

Certificate of Analysis

TMB Stop Buffer

Cat No : T3552

Lot No : 07262415

Exp . Aug. 2027

TEST	Specification	Result
Appearance	Clear liquid	PASS
pH	< 8.0	PASS
Required absorbance at 120 minutes	< 0.05	PASS
Free of precipitate	PASS	PASS

This document has been produced electronically and is valid without signature

Date of Print : 7. 31. 2024