

# Certificate of Analysis

## Bradford (Coomassie) Protein Assay Plus Reagent

Cat No : P7201

Lot No : 09012319

TEST	Specification	Result
Appearance	Clear, Blue Solution	PASS
Protease	None	None
Linear Response OD @ 595nm ( 0.1 mg/ml - 0.5 mg/ml , BSA	0.05—0.7	PASS
Peak A	463 Min - 467 Max	PASS
R Square ( 25 ug/ml - 2000 ug/ml)	> 0.95	0.9779

### QUALITY CONTROL (QC) TESTS & DATA

Tests performed as per standard assay protocol  
Protein (BSA) concentration (0.1-0.5mg/ml)

This document has been produced electronically and is valid without signature

Date of Print : 11. 21. 2023