

Certificate of Analysis

Product Name : 20X PBS Buffer

Cat No : P2200

Lot No : 12082315

TEST	Specification	Result
Appearance	Clear, Colorless Solution	PASS
Sterility		
Aerobic	Negative	NEGATIVE
Anaerobic	Negative	NEGATIVE
pH , 20X Solution <i>pH was taken at room temperature(~ 22°C)</i>	6.1 - 6.3	6.20
Osmolarity	5800 - 6100 mM/kg	6120 mM/kg
Endotoxin <i>Endotoxin testing was performed using LAL detection method.</i>	≤ 1.0 EU/ml	< 0.1 EU/ml
Mycoplasma <i>Mycoplasma test was performed using MycoAlert® Mycoplasma Detection Assay</i>	Negative	NEGATIVE
Animal Cell Culture Toxicity Functionality	Negative ≥ 80%	NEGATIVE PASS

This document has been produced electronically and is valid without signature

Date of Print : 12. 12. 2023