

Certificate of Analysis

DMEM/Ham's F-12 (1:1 Mixture), With L-Glutamine , No Phenol Red

www.gendepot.com

Cat No : CM018	Lot No: 12082319	Exp. Dec. 2025
TEST	Specifications	Results
Appearance	Clear Solution	PASS
Sterility		
Aerobic	Negative	NEGATIVE
Anaerobic	Negative	NEGATIVE
pH pH was taken at room temperature(~ 22°C)	7.10 - 7.40	7.22
Osmolality	320 - 420 mM/kg	302 mM/kg
Endotoxin Endotoxin testing was performed using LAL detection method.	≤ 1.00 EU/ml	< 0.1 EU/ml
Mycoplasma Mycoplasma test was performed using MycoAlert [®] Mycoplasma Detection Assay	Negative	NEGATIVE
Animal Cell Culture Cell promotion Viability Medium was supplemented with 10% FBS and tested using CHO cells grown in a monolayer in a 5% CO2 atmosphere.	3 Passages ≥ 80%	PASS > 80%

This document has been produced electronically and is valid without signature

Date of Print : 12. 15. 2023