

Certificate of Analysis

0.2ml 8-Strip PCR Tubes & Flat Caps, 250 strips of 8 tubes and caps

Cat No : GP203

Lot No : 12132312

TEST	Specification	Result
Appearance	Clear Tube	PASS
DNase	Not detected	PASS
RNase	Not detected	PASS
Endotoxin	Not detected	PASS
Human DNA	Not detected	PASS

All PCR tubes are designed according to industry standards and ensure full contact with cycler blocks. The uniform, thin wall design ensures rapid and consistent heat transfer, reducing variance between samples. Tight sealing caps minimize evaporation, and allow for low reaction volumes. Strip tubes and caps are precision molded to ensure proper sealing and consistent spacing for compatibility with all thermal cycler blocks. All tubes are manufactured in an ISO class 8 cleanroom and undergo QC testing for PCR performance and the absence of contaminating factors. Tubes and caps are made from high-quality, virgin polypropylene, and are certified free from RNase, DNase, endotoxin, and human DNA.

This document has been produced electronically and is valid without signature

Date of Print : 12. 21. 2023