

Certificate of Analysis

Bradford (Coomassie) Protein Assay Plus Reagent

Cat No : P7201

Lot No : 12172118

TEST	Specification	Result
Appearance	Clear, Blue Solution	PASS
Protease	None	None
Linear Response OD @ 595nm (0.1 mg/ml - 0.5 mg/ml , BSA	0.05—0.7	PASS
Peak A	463 Min - 467 Max	PASS

QUALITY CONTROL (QC) TESTS & DATA

Tests performed as per standard assay protocol
Protein (BSA) concentration (0.1-0.5mg/ml)

This document has been produced electronically and is valid without signature

Date of Print : 1. 6. 2022