

# Electrocorder EC-2VA User Instructions



### WARNING!

This product must only be used by suitably qualified personnel; do not attempt to use this product unless you are qualified to do so. High voltages that cause burns and lethal shocks are present during voltage monitoring and recording!

Voltage inputs are not isolated from each other, as one input is energised, other will become live!

T: +44 (0)870 225 1790 F: +44 (0)870 225 1791 E: sales@acksen.com acksen.com

## **Powerfully Measured**

### **GENERAL DESCRIPTION**

Thank you for purchasing the Electrocorder EC-2VA, we hope you enjoy using this product, this package consists of six main components:

- Electrocorder logger (1)
  - The logging unit is housed in a strong ABS case.
- 2. Voltage Input Leads (1)
  - A two-core voltage input lead is provided to allow easy connection to the voltage system.
- 3. Rogowski Coil (1)
  Current input sensor.
- 4. Serial Lead (1)
  - A serial lead is provided to allow connection between the logger and any PC with a 9 pin serial port.
- 5. Software CD (1)
  - Electrosoft software is provided free.
- User Instructions (1)
   These User Instructions are provided to give guidance, to qualified personnel.

### PC HARDWARE REQUIREMENTS

To run Electrosoft you must have certain hardware and software installed on your computer. The system requirements include:-

- Microsoft Windows XP, Vista, 7, 8 or 2003/2008/2012 Server.
- An RS232-compatible. Serial Port and/or USB Port is required for interfacing to an Electrocorder, depending on model.
- Our minimum system specification is a 2Ghz Intel Core2Duo, 2GB RAM, 1GB free disk space.
- 136x768 resolution, 24-bit colour and Windows XP.
- Our recommended system specification is a 3.1Ghz Intel i5-3450, 8GB RAM, 1GB free disk space.
- 1920x1200 resolution, 24-bit colour and Windows 7.

### Installing Electrosoft

When you run the Setup program, it will automatically set a path on When you insert the Electrosoft CD, it will prompt you to run the setup program; follow the on-screen instructions to install Electrosoft. If the setup program does not start automatically, please run "ElectrosoftSetup.exe" on your DVD/CD-ROM Drive.

The setup program will create a shortcut for Electrosoft in the Start menu.

### **GETTING STARTED**

In order to set-up an Electrocorder, you must first run Electrosoft on your PC. Then connect an Electrocorder to the PC serial port using the correct (supplied) serial lead. In Electrosoft, use the 'Setup' dialog box window and input the details of the location to be monitored. The Electrocorder does NOT need to be connected in to the mains voltage to perform this task.

The recording mode is set by default to commence recording when the Electrocorder detects voltage and to stop recording when the memory is full.

Select the recording method - two options are available: 1. Record to EN50160 standard - the Electrocorder will take a sample once every second for 10 minutes. It then averages the samples taken over that 10 minute period and stores the value. In this mode the unit will record for approximately 37 days until the memory is full.

2. Take a sample over a discrete period - the Electrocorder can be set to take an average over a selected period, 1 (one) sec to 15 (fifteen) min and also record the max and min during each period. For example, a unit set to record every 1 (one) second will record for approximately 2 hours. A unit set to record every 12 seconds will record for approximately 1 day. A unit set to record every 15 (fifteen) minutes will record for approximately 75 days.

When the required information has been input, download to the connected Electrocorder by clicking the 'Write Setup' icon. The Electrocorder is now ready to monitor voltage.

When the Electrocorder is recording a flashing red light will show and when it has completed recording, a green light will appear on the unit. The database contained within Electrosoft will also advise that the unit has completed recording and is ready to be collected. To download the recorded data connect the Electrocorder to the PC serial port and click the 'Read Setup' icon. The recorded data is displayed for analysis.

This document is produced in conjunction with the Help file contained in Electrosoft, which contains a detailed explanation of all features and contains information, which should be studied prior to using this product.

### USB to RS232 Serial Converter

If you have purchased a USB to RS232 converter, you must install the drivers. You can use the drivers shipped with the program which may be in the USB sub-folder within the Program Folder, normally C:\Program Files\Electrosoft\USB. You can download them from the website www.electrocorder.com or use the disk, if one came with the converter.

The following describes the XP installation, other operating systems will vary slightly. When you plug the converter into the PC, it will detect it and identify the new hardware as UC232R, Windows will then ask to search for the drivers, choose "Yes, this time only", then on the next screen choose, "Install from a list location" then specify the location of the drivers, possibly the USB sub folder, in the installation folder, or wherever you saved the files to when you downloaded from the internet.

When installed, make a note of the serial or COM port number the converter has been assigned to and when you run Electrosoft, select the appropriate serial port or COM port number.



## **SAFETY TIPS**

The 2-core voltage input is labelled 'L1' and the current input A1. For correct operation this recorder must have a proper Neutral (N) connection. Brown is the live conductor and blue the neutral.

- 1. If possible electrically isolate the conductors to which you wish to connect.
- First, using insulating gloves place the current input sensor around the conductor to be monitored. It must only be used around an insulated conductors, this sensor is rated to 600V CAT II
- 3. Using insulating gloves, connect the each voltage cores in turn to the electrical bus-bars.
- The voltage lead is double insulated, 300V. Always check that the lead is in good condition, if it becomes damaged, then it MUST be replaced.

### Features & benefits of the EC-2VA logger system


Feature	Benefit
Unit is small and lightweight	Easy to install
Easy to use Windows software	Can be used by non-technical staff
Electrosoft contains internal database	Allows effective management of
	distributed Electrocorders
True RMS voltage measurement	Complies to EN50160:1994

### Colour codes around the world

Phase	IEC	Aus/NZ	US	Canadian
	Colour Code	Colour Code	Colour Code	Color Code
L1 (A)	Brown	Red	Black	Red
L2 (B)	Black	White	Red	Black
L3 (C)	Grey	Blue	Blue	Blue
N (Neutral)	Blue	Black	White, grey	White
G	Yell/Green	Yell/Green	Green	Green
(Ground/Earth)	Stripe	Stripe	Yell/Green	bare copper

### Inputs and Connections on Various Systems

Colour and	Single Phase	Single Phase
Input Terminal	(2-Wire)	(3-Wire)
Brown (L1)	X (Live)	X (Live)
Green/Yellow		X (Earth)
(G/E)		
Blue (N)	X (Neutral)	X (Neutral)



# **Single Phase 2 Wire Systems (using inputs 'L1' and 'A1')** Ensure the logger has been set-up with a PC to log for a specific period.

Using insulating gloves place the current sensor around the voltage busbar being monitored, the simply connect the brown wire (of L1) to System Live/Hot and connect the blue (of L1) to System Neutral ('N'). The logger will automatically start to record when voltage on input L1 is detected. If you do not wish to record current (or cannot) then you do not have to connect A1; the logger will then only record voltage.

We would not recommend connecting the Neutral (N) to anything other than System Neutral. We do not recommend connecting this input to an external System Ground (G) or Earth (E) as the floating voltage on the Neutral (internal Star-point) could trip sensitive Earth Fault or Earth Leakage protection equipment.

## Power Measurement (Using inputs 'L1' and 'A1')

Ensure the logger has been set-up with a PC to log for a specific period.

To record the power being taken on a voltage busbar or cable, firstly using insulating gloves place the current sensor around the voltage busbar being monitored, then connect the blue wire of L1 to the system Neutral ('N') then connect brown wire (of L1) to that voltage busbar and . Ensure the current carrying cable is insulated! The logger will automatically start to record when voltage on input L1 is detected.

We do not recommend connecting this input to an external System Ground (G) or Earth (E) as the floating voltage on the Neutral (internal Star-point) could trip sensitive Earth Fault or Earth Leakage protection equipment.

# **Technical specifications**

recrimed specimentions	
Measurement range (Vrms)	0Vac 300Vrms (Ph – N)
Maximum channel input voltage	300Vrms (Ph – N)
Voltage measurement accuracy	±1.5% of reading (10 bit) within 70Vac – 300Vrms (Ph – N); else ±3%. (50/60Hz ±2%)
Current range	0 – 300A (other ranges can be factory set)
Current measurement accuracy	3% of reading 50 to 300A, otherwise 5%
Dimensions of current sensor	Max conductor diameter 120mm (5")
Vmin & Vmax meas time resolution	Always one cycle (50/60 Hz), independent of selected averaging period
Sampling frequency	16 samples per cycle 800Hz @ 50Hz or 960Hz
	@ 60Hz
Data recorded	Average voltage/current, max & min voltage/ currentcycle-value during the averaging period
Memory capacity	128kB able to record 32,000 Voltage levels per channel
Memory type	Non-volatile SEEPROM
Memory - averaging period & duration	1 sec to 60 mins (1sec. avg gives 2 hrs of log ging, 60min. avg gives 300 days of logging)
Real-time clock accuracy	Greater than 0.001%
Input lead length	Metric 1.0 metres Imperial 3'6" (3 feet, 6 inches)
Battery life (while plugged in)	Unlimited - mains powered & battery backup (9,000 hours, 1 year while unpowered)
Battery type	Unit contains one 9V Alkaline batteries (E-Block, PP3, 1604A)
Communications interface type	RS-232 serial, baud of 19,200
Electrosoft software	Windows (9x, 2K, ME, NT, XP, Vista, Windows 7); 1024 x768 min resolution
Environmental (temp & sealing)	-20C to +40C or +14°F to +104°F – Indoor or protected environment only, IP65, NEMA 4X
Dimensions & weight	Metric 145 x 90 x 45mm & 400g Imperial – 5.5" x 3.5" x 2" & 1lb
Standards	Recording - EN50160: 1994 - CAT II,

(subject to change without notice)

### Calibration

Each unit is individually calibrated during testing.

Battery life (while connected)

Unlimited - mains powered and battery back up.

Battery life (while unconnected to mains)

The  $9\dot{V}$  Alkaline batteries should last for at least 9,000 hours (1 year).



# Caution

The battery used in this device may present a risk of fire or chemical burn if mistreated. Do not recharge, disassemble, heat above 100°C or incinerate. Replace with a 9V Lithium or Alkaline battery IEC Type 6-F22 (PP3, MN1604). Use of another battery may present a risk of fire or explosion. Dispose of used batteries promptly. Check for signs of battery (electrolyte) leakage. If leakage has occurred, the PCB must be cleaned in an approved manner by a competent (trained) person. Keep away from children.

### Maintenance

Regularly check the Electrocorder casing for signs of damage (cracks, broken or loose parts) or misuse. If the unit is damaged in any way it must NOT be used and should be returned to the supplier. The unit must not be used for any other purpose than for that recommended by the manufacturer. The unit must not be submerged in any liquid.

### Cleaning

Wipe the outside of the case with a clean cloth damped with IPA (Isopropyl Alcohol).

#### Warranty

All Acksen products carry a minimum 1 year's warranty covering manufacturing defects and component failures. The device contains no user-serviceable parts and as such should only be repaired by skilled and authorised personnel. Failure to comply could result in unsafe operation and should not be attempted under any circumstances. Contact below for a list of approved service agents. Note: Any unauthorised repair or adjustment will automatically render the warranty invalid.

Repair and spare parts

Acksen Ltd, 28 Station Road, Whiteabbey Newtownabbey, Co. Antrim BT37 0AW United Kingdom

Or an approved repair company.

### Returning a product for repair

If returning a product to the manufacturer for repair, it should be sent freight pre-paid to the appropriate address. A copy of the Invoice and of the packing note should be sent simultaneously by airmail to expedite clearance through Customs. A repair estimate showing freight return and other charges will be submitted to the sender, if required, before work on the device commences.

### WEEE

For EU customers Acksen Ltd offer a product take-back service. For customers within the European Union (only) and products manufactured or sold by us; when those products reach the end of their life, simply send them back to us at your expense, we will dispose of them according to the relevant legislation. WEEE Reg. No. WEE/DD2117VU