

Date

# **ANALYTICAL REPORT**

**Client Information** 

Bristol Mask Limited Unit 2B Barton Hill Trading Estate Herapath Street BRISTOL UNITED KINGDOM					
contact@bristolmask.uk					
-					
	Sample Information				
Order Code:	EUAA70-00010345				
<b>Reception Date:</b>	28-Jan-2021				
Analysis Starting Date:	28-Jan-2021				
Analysis Ending Date:	2-Feb-2021				
Sample code Nr.	560-2021-00000901				
Sample described as:	Masks				
	Requirements and decision rule				
Customer requirements: EN 14683:2019+AC:2019 TYPE IIR					
Decision Rule:	Shared risk - Simple acceptance.				
Information provided by the customer*					
Client Reference: BRIDGE30GSM					
Sample Description:					
Purchase Order Number:					
Batch	002				





Page: 2 / 6 AR-21-YL-000946-01 560-2021-00000901 02/02/2021

# SAMPLE PICTURE



Eurofins Textile Testing Spain, S.L.U. Calle Germán Bernácer, 4 03203 Elche SPAIN Phone+34 966 299 638 www.eurofins.com/tex





Page: 3 / 6 AR-21-YL-000946-01 560-2021-00000901 02/02/2021

## CONCLUSION:

TEST PROPERTY	PASS	FAIL	REMARKS
Resistance against penetration by synthetic blood ISO 22609:2004			
Mask	Х		

**Remark:** Test has been performed as per application request





Page: 4 / 6 AR-21-YL-000946-01 560-2021-00000901 02/02/2021

#### COMPONENT LIST:

COMPONENT ID	COMPONENT NAME	MATERIAL DESCRIPTION	COLOR	REMARKS
CUST 01	Mask	Mask	White	





Page: 5 / 6 AR-21-YL-000946-01 560-2021-00000901 02/02/2021

MASKS TESTING	CAS No.	RESULTS	UNC.	LOQ		GUIDELINES
Analyses on:Mask						
Resistance against penet	tration by synthetic blood				Ana	lysis Ending Date: 02/02/2021
ISO 22609:2004						
Number of specimens tested		32		-		
N° of specimens failed		0		-		
N° of specimens passed		32		-	≥29	✓ Pass
At least 29 of 32 specimens mu	ust pass tested at 16KPa					

Complete test data reported at Annex.

AQL information according to ISO 22609:2004:

A single sampling plan providing an AQL of 4,0 % requires 32 test specimens.

An AQL of 4.0 % is met for a single sampling plan when 29 or more specimens show pass results.

AQL= Acceptable Quality Limit.





Page: 6 / 6 AR-21-YL-000946-01 560-2021-00000901 02/02/2021

#### Signed for and on behalf of Eurofins Textile Testing Spain:



Report electronically validated by

Maria Jesus Martinez Puig Chemical Lab manager

#### **EXPLANATORY NOTE**

- Test not covered by ENAC accreditation scope
- Test is subcontracted within Eurofins group and is accredited
- Test is subcontracted within Eurofins group and is not accredited
- Test is subcontracted outside Eurofins group and is accredited
- □ Test is subcontracted outside Eurofins group and is not accredited

N/A = Not Applicable

\*Eurofins Textile Testing Spain S.L.U is not responsible of the information supplied by the costumer and reported as section "Information provided by the costumer\*".

Eurofins General Sales Terms and Conditions Applied.

Results obtained refer only to samples, products or material received in Laboratory, as described in section "Sample information" and tested in conditions shown in present report.

Test uncertainties not reported are at customer disposal, for those tests in which it is possible to evaluate the test uncertainty.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k = 2, which for a normal distribution provides a level of confidence of approximately 95%.

Reproduction of this document is only valid if it is done completely and under the written permission. Any printed version of this document will be considered as a copy.

If you happen to have any comments, please do it by sending email to textile\_spain@eurofins.com and referring to this report number.

End Of Report





#### DETERMINATION RESISTANCE AGAINST PENETRATION BY SYNTETIC BLOOD

Test Method: ISO 22609:2004; Targeting-plate test method

Number of test specimens: 32

Sample size: Circular, diameter 5,58 cm

Sample area tested: 24,5 cm<sup>2</sup>

Pressure: 16 kPa (120,0 mm Hg)

Stream velocity of synthetic blood: 550±10 cm/s

Distance of the face mask target area surface from the tip of the cannula: 30,5 cm

Angle of the pneumatic valve with respect to the face mask target area: 90°

Technique used to enhance visual detection of synthetic blood: Hydrophilic cotton

Conditioning: At least 4 hours. T<sup>a</sup> between 16,7°C and 26°C. RH between 82,8% and 88%

Environmental test conditions 19,3°C; 79,2% Hr

Pre-treatment: None

Results				
Specimen	Pass	Fail		
1	Х			
2	Х			
3	Х			
4	Х			
5	Х			
6	Х			
7	Х			
8	Х			
9	Х			
10	Х			
11	Х			
12	X			
13	Х			
14	Х			
15	Х			
16	Х			
17	Х			
18	Х			
19	Х			
20	Х			
21	Х			
22	Х			
23	Х			
24	Х			
25	Х			
26	Х			
27	Х			
28	Х			
29	Х			
30	X			
31	X			
32	Х			

Conclusion	PASS



## Operating requirements for surgical masks based on EN 14683: 2019+AC: 2019 standard

TEST	ΤΥΡΕ Ι	TYPE II	TYPE IIR
Bacterial filtration efficiency (BFE), (%)	≥ 95	≥ 98	≥ 98
Differential pressure (Pa/cm <sup>2</sup> )	< 40	< 40	< 60
Splash resistance pressure (kPa)	Not required	Not required	≥ 16
Microbial cleanliness (CFU/g)	≤ 30	≤ <b>30</b>	≤ <b>3</b> 0