EVOLT!

Evolt 360 Specifications

Model	Evolt 360
Measuring Method	Tetra-polar electrode method using 8 touch electrodes
Frequency Range	5, 50, 250kHz
Measuring Site	Whole body and segmental measurement (arms, legs and trunk)
Result Sheet Data	Body Composition Results
	Weight, Lean Body Mass, Skeletal Muscle Mass, Protein, Mineral, Total Body Water, Body
	Fat Mass, Subcutaneous Fat Mass, Visceral Fat Mass, Visceral Fat Area, Visceral Fat Level,
	Body Fat Percentage, Intracellular Fluid (ICF), Extracellular Fluid (ECF), BMR, TEE, Bio Age, BWI Score, Segmental Analysis Lean Mass & Fat Mass of Left Arm, Right Arm, Left Leg,
	Right Leg, Torso, Abdominal Circumference, Waist to Hip Ratio, Nutrition – Calories,
	Protein, Carbohydrates, Fat, Supplement Recommendations.
Power Consumption	60VA
Measuring Current	Approx. within 180µA
Power Consumption	Input (AC 100-240V, 50-60Hz), Output (DC 12V, 5A adaptor
Display	8 inch wide Colour LCD Touch Screen
Input Device	Touch Screen, PC Remote Control
Transmission Device	USB Port, RS-232C, Bluetooth (Option)
Dimension	Main Unit 414x636x980mm (WxDxH±10mm)
Weight	Approximately 18kg (main unit)
Measuring Range	100~950Ω
Measurement Time	Within 30 seconds
Applicable Height	50-250cm
Measuring Weight	10-250kg
Applicable Age	7-99 years old
Operation Ambient	Ambient temperature range +5 to +40°C, Relative humidity range 15 to 93% (non- condensing)
Storage Ambient	Ambient temperature range -25 to +70°C, Relative humidity range lower than 93% (non- condensing)



For purpose of improvement, specifications and design are subject to change without notice. This is a medical device. Read precaution and operation method before use.

Evolt IOH Pty Ltd HQ 1/24 Leda Drive, Burleigh, QLD, 4220 TEL +64 7 5659 1958 |Email info@evolt360.com