

TEST REPORT

Report No.: SHT16110079-01E

Date: 2016-12-05

Page 1 of 2

The following information of testing sample

Sample Name	Disposable Surgical Gown	Sample No.	T16110079-01
Sample Type	/	Sample Qty.	4pcs
Applicant	Henan Joinkona Medical Products Stock Co.,Ltd		
Address	Xinxing Road, The South of Industry District, LuShan County, PingDingShan City, Henan Province, China		
Sample Received Date	2016-10-28	Testing period	2016-10-28 to 2016-12-01
Test Items	Water resistance, Hydrostatic pressure test		
Test Method	AATCC 42-2013 AATCC 127-2013		
Test Results	Please refer to next page(s)		

*****For more detailed information, please refer to next page(s)*****

Redacted by:

王玉芳

(Yufang Wang)

Reviewed by:

吴涛

(Tao Wu)

Approved by:

(Authorized signatory: Fei Xia)

TEST REPORT

Report No.: SHT16110079-01E

Date: 2016-12-05

Page 2 of 2

Test Result(s):

1. Water resistance

Test Method: AATCC 42-2013

Determinate Standard: According to applicant's requirement

Test Item	Unit	Test Result	Limited Value	Determination
Water resistance	g	0.4	≤ 1.0	Pass

2. Hydrostatic pressure test

Test Method: AATCC 127-2013

Determinate Standard: According to applicant's requirement

Test Item	Unit	Test Result	Limited Value	Determination
Hydrostatic pressure test	cmH ₂ O	65.2	≥ 50	Pass

Sample Photo



The End

Remark: The samples provided by the customer, the report test results are only responsible for samples.

The test data is only used for reference, not as a social justice data.