

Joie Water PSB1 2/20/20

Lab ID: 2002139-05

Bell Food and Water

METRC Batch ID:

Date Sampled: 02/25/20

Date Printed: 03/2/20

Report cannot be used for OLCC/OHA compliance.

Potency Analysis

Analytical Method: De Backer, Journal of Chromatography b.2009. 11.004 - SOP 102

Cannabinoids	mg/g	LOQ
THCA	< LOQ	0.000375
delta 9-THC	< LOQ	0.000375
delta 8-THC	< LOQ	0.000375
CBGA	< LOQ	0.000375
CBDA	< LOQ	0.000375
CBD	0.0217	0.000375
CBN	< LOQ	0.000375
CBG	< LOQ	0.000375
CBC	< LOQ	0.000375

Total THC
< LOQ mg/g

Total CBD
0.0217 mg/g

<LOQ - Results below the Limit of Quantitation

Acid form of THC/CBD are decarboxylated by heat, lose 12% of original mass as CO2. Result = *bioactive*

"Total" Cannabinoid accounts for decarboxylation and moisture content. Total THC = $[(THCA \times 0.877) + \Delta 9THC] / (100\% - MC)$



Erik Werstler
Lab Director

Joie Water PSB1 2/20/20

Bell Food and Water

Laboratory ID: 2002139-05

Quality Control Potency

Batch: B20B112 - Potency

Blank(B20B112-BLK1)

Analyte	Result	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	< LOQ	0.000374	mg/g		02/27/20 15:49	02/27/20 20:42	
delta 9-THC	< LOQ	0.000374	mg/g		02/27/20 15:49	02/27/20 20:42	
CBGA	< LOQ	0.000374	mg/g		02/27/20 15:49	02/27/20 20:42	
CBDA	< LOQ	0.000374	mg/g		02/27/20 15:49	02/27/20 20:42	
CBD	< LOQ	0.000374	mg/g		02/27/20 15:49	02/27/20 20:42	
CBN	< LOQ	0.000374	mg/g		02/27/20 15:49	02/27/20 20:42	
CBG	< LOQ	0.000374	mg/g		02/27/20 15:49	02/27/20 20:42	
delta 8-THC	< LOQ	0.000374	mg/g		02/27/20 15:49	02/27/20 20:42	
CBC	< LOQ	0.000374	mg/g		02/27/20 15:49	02/27/20 20:42	

LCS(B20B112-BS1)

Analyte	% Recovery	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	116	0.000374	mg/g	85-115	02/27/20 15:49	02/27/20 20:24	Q1
delta 9-THC	104	0.000374	mg/g	85-115	02/27/20 15:49	02/27/20 20:24	
CBDA	103	0.000374	mg/g	85-115	02/27/20 15:49	02/27/20 20:24	
CBD	103	0.000374	mg/g	85-115	02/27/20 15:49	02/27/20 20:24	

Notes and Definitions

- B Analyte detected in method blank, but not associated samples.
 - B2 Analyte detected in sample and associate method blank.
 - C Interference due to co-elution.
 - D Initial result exceeded calibration range, reported data are based on analysis of a dilution.
 - H Non-homogenous sample matrix affecting RPD and/or QC.
 - I Manual Integration was performed.
 - L Duplicate sample relative percent difference (RPD) exceeds QC limits.
 - M Anomalous results due to matrix interference
 - P Peaks manually split.
 - Q1 QC out of limits but still ok
 - Q2 Quality Control outside QC limits. Data considered estimate.
 - Q3 CCV was above the acceptance criteria. Non-detect samples are considered acceptable.
 - Q4 CCV was below the acceptance criteria, however the sample still exceeds the regulatory limit.
 - R Marginal Exceedence.
 - U Reported result is an estimate. The analyte was detected above the calibration range.
 - X Problems with initial analysis, reported data are from reinjection of prepared sample.
- <LOQ - Results below the Limit of Quantitation - Compound not detected



Erik Werstler
Lab Director