

TECHNICAL SPECIFICATIONS

General Specification

Light Source	Xenon Flash Lamp
Detector	CCD (2,048 Pixels)
Wavelength Accuracy	±1 nm
Wavelength Range	200 – 1100 nm
Spectral Resolution	0.3 nm
Dimension	220 x 280 x 220 mm 8.6 x 11.0 x 8.6 in
Weight	5 kg
Operator Voltage	100 – 240 V, 50 – 60 Hz
Power Consumption	80 – 100 W
Warranty	1 Year
LCD Touch Panel	7.0 in

Nano Volume Specificaiton

Absorbance Precision	1% at 100 ng/μl
Absorbance Range	0 – 400 Abs. (10mm equivalent)
Detection Limit	2 ng/μl (dsDNA)
Maximum Concentration	20,000 ng/μl (dsDNA), 200 mg/ml (BSA)
Measurement Time	5 sec
Minimum Sample Size	1 μl
Path Length	0.05 – 1.2 mm (Auto-ranging)

Cuvette Specificaiton

Beam Height	8.5 mm
Absorbance Range	0.002 – 2.0 Abs.
Measurement Time	3 sec



Nabi

UV/Vis NANO SPECTROPHOTOMETER



AFSBio Inc
Toronto, ON, Canada
www.afsbio.com
info@afsbio.com



Nabi

UV/VIS NANO SPECTROPHOTOMETER

Variety & Precision in Detection

- End Point & Kinetic Mode in Detection
- High Speed Spectrum Reading (5 Seconds) Accurate Detection (Less than ± 0.002)
- Wide Range of Wavelength Detection with Cuvette
- Nano Volume Sample Measurement (Nucleic Acid / Protein)
- High Repeatability in Quantitative Measurement for Nucleic Acid

Nabi UV/VIS NANO SPECTROPHOTOMETER



Nabi

UV/VIS NANO SPECTROPHOTOMETER

Simple & Light in Design

- Compact Size and Light Weight
- Easy Operation with Minimized Steps for Measurements and Analysis
- 7.0 inch Touch LCD
- Quick Boot – No Lamp Warm-up Time
- Stand-Alone – Perform All Functions without PC

Nabi

UV/VIS NANO SPECTROPHOTOMETER

Result Analysis

- Single Wavelength / Spectrum Selection
- End Point / Kinetic Mode Selection
- Quantitative Measurements for Nucleic Acid, Protein, and etc.
- Cell Concentration
- Simple Data Backup and Software Upgrade via USB Flash Drive
- Easy to Confirm the Result via Graphical Presentation

