Caligreen Laboratory 13340 Saticoy Street Unit H North Hollywood, CA 91605 (818) 979-9474 www.caligreenlaboratory.com Lic# C8-0000104-LIC

1 of 1

Breathe

Sample ID: 2009CGL1145.3062

Strain: Face Mask

Matrix: Topical

Type: Lotion Sample Size: 1 units; Batch: Produced:

Collected:

Received: 09/11/2020 Completed: 09/15/2020

Batch#:

Client

Halsa Botanicals

Lic.#

1055 W 7th St #3100

Los Angeles, CA 90017



Summary

Test

Batch

Cannabinoids Moisture

Date Tested

09/14/2020

09/15/2020

Complete

Result

Complete NR - Complete

Cannabinoids Complete

0.029%

0.029%

Total THC

0.694%

0.694%

Total CBD

0.758%

0.758%

Total Cannabinoids

Analyte	LOD	LOQ	Result	Result		
	mg/g	mg/g	%	mg/g		
THCa	0.01	0.02	ND	ND		
Δ9-THC	0.02	0.02	0.029	0.29		
Δ8-THC	0.02	0.02	ND	ND		
THCV	0.02	0.02	ND	ND		
CBDa	0.01	0.02	ND	ND		
CBD	0.02	0.02	0.694	6.94		
CBDV	0.02	0.02	0.004	0.04		
CBN	0.01	0.02	<loq< th=""><th><loq< th=""><th></th><th></th></loq<></th></loq<>	<loq< th=""><th></th><th></th></loq<>		
CBGa	0.01	0.02	0.003	0.03		
CBG	0.02	0.02	0.005	0.05	t E B	
CBC	0.01	0.02	0.022	0.22		
Total THC			0.029	0.29		
Total CBD			0.694	6.94		
Total			0.758	7.58		

Date Tested: 09/14/2020

 $\begin{array}{l} \text{Enter Constant} \\ \text{Im L} = 1.2601g. \\ \text{Method: CGL-POT-UHPLC-v1.0} \\ \text{Total THC} = \text{THCa*0.877} + \Delta 9\text{-THC}; \\ \text{Total CBD} = \text{CBDa*0.877} + \text{CBD} \\ \end{array}$

LOQ = Limit of Quantitation; The reported result is based on a sample weight with the applicable moisture content for that sample; Unless otherwise stated all quality control samples performed within specifications established by the Laboratory.

> C. S. Le Vergue Cynthia LeVesque - Laboratory Director

Confident Cannabis All Rights Reserved support@confidentcannabis.com (866) 506-5866

Cyntina Levesque - Laboratory Director

9/15/2020

Www.confidentcannabic.com

WAMC-ROHAUS. The instruments used are the Rotronic HC2-AW-USB Water Activity and Temperature meter and OHAUS MB27 Moisture Analyzer. The SOP for foreign material is CGL-FM-DINO performed with the Dino-Lite A904126 digital microscope. All LQC samples were performed and met the prescribed acceptance criteria in 16 CCR section 5730, pursuant to 16 CCR section 5720, pursuant to 16 CCR section 5730, pursuant to 16 CCR