

EC Declaration of Conformity

Manufacturer: PHASE Scientific International Limited
Name: Kent Cheng, Ph.D.
Address: 32F, Gravity, 29 Hing Yip St., Kwun Tong, KLN, HK
Tel: +852 3892 7200
E-Mail: info@phasesci.com

Authorized Representative:
Name: Lotus NL B.V.
Address: Koningin Julianaplein 10,1e
Verd, 2595AA, The Hague, Netherlands.
E-mail: peter@lotusnl.com

We, the manufacturer, here with declare that the product(s)

Product Name	INDICAID[®] COVID-19 Rapid Antigen Test	Specification	IVD
Intended Use	The INDICAD [®] COVID-19 Rapid Antigen Test is a lateral flow immunoassay designed for the qualitative detection of SARS-CoV-2 antigens in direct nasal swab samples.		
Classification	Others		

Conformity Assessment Route : IVDD98/79/EC Annex III.

Applicable Standards:

ISO 13485:2016
ISO 14971:2019
EN ISO 18113-1:2011

EN ISO 18113-2:2011
EN 13641:2002
ISO 15223-1:2016

EN 13612:2002
ISO 23640:2015
EN 62366-1:2015



We, the manufacturer, here declare with sole responsibility that our product/s mentioned above meet/s the provisions of the Directive 98/79/EC of the European Parliament and of the Council on In-Vitro Diagnostic Medical Devices.

We agree to develop, implement and maintain a documented post-production monitoring process.

Name of Product Director	Kent Cheng, Ph.D.
Signature	
Date	20 October 2020
Place	Hong Kong, China.
Seal (Manufacturer)	

