

is a 3Rs refinement technique that reduces stress in CD-1 mice


Kendra Steele¹, Barbara Stone¹, James Hester¹, Brett Spear², and Angelika Fath-Goodin¹

¹ParaTechs Corporation, Lexington, KY, ²The Markey Cancer Center, University of Kentucky, Lexington, KY, USA

Introduction

Non-surgical embryo transfer (NSET) can be used to transfer blastocysts to recipient mice during the generation of transgenic mice, after cryopreservation of embryos, after in-vitro fertilization, and during rederivation of mouse strains. Non-surgical embryo transfer is a refinement (one of the 3Rs) of the surgical embryo transfer procedure. The NSET™ device is a small tapered catheter that attaches to a P2 pipette and deposits embryos directly into one of the mouse's uterine horns. Use of an NSET™ device in mice should substantially reduce pain, suffering, and recovery time of mice as the surgical procedure is replaced with an alternative method. No anesthesia is required for non-surgical embryo transfer and the mice require no post-procedure recovery. Because the NSET procedure is non-surgical and does not necessitate anesthesia, we hypothesized that use of the NSET™ device is less stressful than surgery for the mouse. We therefore compared the stress responses of pseudopregnant CD-1 mice (2.5 days post coitum) following uterine embryo transfer surgery and non-surgical embryo transfer procedures using the NSET™ device. ELISA analysis was used to quantify levels of fecal corticosterone, a biomarker of stress. Electrocardiograms were performed using a non-invasive device, the ECGenie by Mouse Specifics, Inc., to measure cardiac indicators of stress.

Electrocardiograms with the ECGenie™



Electrocardiograms are recordings of the electrical currents that pass through the heart and signify the activity associated with each heart beat. The non-invasive ECGenie instrument was used to record electrocardiograms from mice without restraint or a surgical implant. Readings were made by placing the mouse on a sensory pad outfitted with three monitoring electrodes. The cardiac activity was measured passively through the paws. Time necessary to record an acceptable cardiac signal depended on the mobility of the mouse and averaged between 5 - 8 minutes.

<http://www.bcm.edu/phenotyping/ecgenie.htm>

Embryo Transfer

NSET™

vs


Surgery



Photo courtesy of Marcelo F. G. Nogueira (UNESP)

<http://surgery.wustl.edu>

Heart rate decreases dramatically in response to anesthesia.




Heart rate in beats per minute (BPM) are graphed at various hours (h) post procedure for 2.5 dpc pseudopregnant CD-1 mice. Data was analyzed by Mouse Specifics using EzCG analysis software. Data presented are means and standard deviations (N=11). Statistical significance ($P \leq 0.05$) for the comparison of the control group versus the experimental groups is represented by an asterisk (*).

The NSET procedure is effective for embryo transfer of blastocysts.


NSET™ Advantages

- Simple procedure
- Does not require anesthesia or an analgesia
- Requires less time to
 - Prepare the mouse
 - Perform the procedure
 - Monitor the mouse for post-surgical complications
- Requires less equipment
- Uses less space
- Reduces costs up to 75%




Surgical and NSET procedures are both effective for blastocyst transfer to the uterine horn. Blastocysts (e3.5) were transferred to the uterine horn of 2.5 dpc pseudopregnant female CD-1 mice. Pregnancy rate (% of females becoming pregnant after embryo transfer) and birth rate (% of pups per embryos transferred to females that became pregnant) are compared.

Variability in cardiac rhythm increases in response to surgery.



Variability in beat-to-beat rhythm of the heart is measured as the coefficient of variance (CV%) and is graphed at various hours (h) post procedure for 2.5 dpc pseudopregnant CD-1 mice. The data was analyzed by Mouse Specifics using EzCG analysis software. Data presented are means and standard deviations (N=11). Statistical significance ($P \leq 0.05$) for the comparison of the control group versus the experimental groups is represented by an asterisk (*).

Fecal corticosterone levels are higher in response to surgery than to insertion of the NSET™ device.



Fecal samples were collected from pseudopregnant mice (2.5 dpc) assigned to five experimental treatment groups (N=15). Corticosterone levels were measured by ELISA assay (Cayman Chemical Co., Ann Arbor, MI) according to the manufacturer's directions. (*) represents a significant difference between the control group and other groups. (•) represents a significant difference between the surgery group and other groups.

Conclusions

- The NSET procedure is effective for uterine transfer of blastocyst stage embryos.
- Surgery causes a significant increase in fecal corticosterone levels, whereas the NSET procedure does not.
- Anesthesia causes a 38-50% decrease in the number of heart beats per minute, whereas the NSET procedure does not alter heart beats per minute.
- Surgery increases cardiac variability by 8%, whereas anesthesia alone and the NSET procedure do not significantly increase cardiac variability.
- Use of the NSET™ device is less stressful to mice than surgical embryo transfer.

Acknowledgements. We thank Kim Carrico for her contributions towards developing the ELISA protocol, and Mouse Specifics, Inc. for their generous support regarding the ECGenie device and analysis of the data. Research reported in this publication was supported in part by the Office of the Director, National Institutes of Health under Award Numbers 2R44RR025737-02 & 8R44OD010958 and by the Kentucky Cabinet for Economic Development, Office of Commercialization and Innovation under the grant agreement KSTC-184-512-11-115 with the Kentucky Science and Technology Corporation.