
Datex-Ohmeda Cardiocap™/5 for Anesthesia

Software S-XANE01

User's Reference Manual





Conformity according to the Council Directive 93/42/EEC concerning Medical devices

CAUTION: U.S. Federal law restricts this device to sale by or on the order of a licensed medical practitioner. Outside the USA, check local laws for any restriction that may apply. All specifications subject to change without notice. Document Number M1031868-02

November 2004

Datex-Ohmeda Inc. P.O. Box 7550 Madison, WI 53707-7550, USA Tel: +1-608-221 1551. fax: +1-608-222 9147 www.us.datex-ohmeda.com/

Datex-Ohmeda Division, Instrumentarium Corporation P.O. Box 900, FIN-00031 DATEX-OHMEDA, FINLAND Tel: +358 10 39411, fax: +358 9 1433310 www.datex-ohmeda.com/

NOTICE

Intended use

The Datex-Ohmeda Cardiocap TM/5 and accessories are indicated for indoor monitoring of hemodynamic (ECG, impedance respiration, NIBP, temperature, SpO₂, and invasive pressure), respiratory (CO₂, O₂, N₂O, respiration rate, anesthetic agent, and agent identification), ventilatory (airway pressure, volume, and flow), and relaxation status (NMT) of all hospital patients.

With the N-XOSAT option, monitoring of arterial oxygen saturation includes monitoring hospital patients during conditions of clinical patient motion.

The Cardiocap/5 is indicated for patients weighing 5 kg (11 lb.) or more.

Impedance respiration measurement is indicated for patients ages 3 years and older.

The monitor is indicated for use by qualified medical personnel only.

CAUTION: US Federal law restricts this device to sale by or on the order of a licensed medical practitioner. Outside the USA, check local laws for any restrictions that may apply.

Classifications

IEC 60601-1:

- Type of protection against electric shock: Class I equipment.
- Degree of protection against electric shock (indicated by a symbol on the monitor beside each connector): Type BF applied part or Type CF applied part.
- The equipment is not suitable for use in the presence of a flammable anesthetic mixture with air or with oxygen or nitrous oxide.
- Mode of operation: Continuous.

IEC 60529 (degree of protection against harmful ingress of water): IPX1

EU Medical Device Directive: IIb

CISPR 11: Group 1, class A

Responsibility of the manufacturer

Datex-Ohmeda Division, Instrumentarium Corp. is responsible for the safety, reliability and performance of the equipment only if:

- Assembly, operation, extensions, readjustments, modifications, service, and repairs are carried out by personnel authorized by Datex-Ohmeda.
- Electrical installation complies with appropriate requirements.
- The equipment is used in accordance with the Cardiocap/5 User's Guide and serviced and maintained in accordance with the Cardiocap/5 Technical Reference Manual.

Trademarks

Datex®, Ohmeda®, and other trademarks (Cardiocap/5, AS/3, CS/3, S/5, S/5 Light, D-lite, Pedi-lite, D-fend, D-fend+, MemCard, ComWheel, EarSat, FlexSat, OxyTip, PatientO₂, and Patient Spirometry) are the property of Instrumentarium Corp. or its subsidiaries.

Nellcor® is a registered trademark of Mallinckrodt Inc. Durasensor®, Dura-Y® Oxiband®, OxiCliq®, and Oxisensor® are registered trademarks of Nellcor Puritan Bennett.

All other product and company names are the property of their respective owners.

© 2004 General Electric Company. All rights reserved.







Cardiocap/5 User's Reference Manual—Anesthesia

Part I - Setup and General Use

- **Monitor Description** 1
 - Monitoring Basics 2
 - Monitor Setup 3
 - Alarms 4
- Trends and Snapshots 5
- Patient Data Management 6
 - Recording and Printing 7
 - Troubleshooting 8
- Maintenance and Cleaning 9

Part II - Measurement Parameters

E	CG	10
	vu	

- Impedance Respiration 11
 - Pulse Oximetry 12
- Non-Invasive Blood Pressure (NIBP) 13
 - Invasive Blood Pressure 14
 - Temperature 15
 - Airway Gases 16
 - Patient Spirometry 17
- NeuroMuscular Transmission (NMT) 18



Introduct	ion	1-1
	Additional information	1-1
Clinic	cal application	1-1
	iguration	
	Configuration passwords	
Cardioca	p/5 model and options for Anesthesia software	1-2
	odynamic model with airway gas measurement (F-MXG)	
	collection and management options	
	he monitor	
Patient co	onnectors panel	1-5
Rear pane	el	1-6
Rear	panel connections	1-6
Symbols .		1-7
Delayed s	tandby function	1-8
Back-up I	battery	1-8
-	ry time	
Safety pro	ecautions	1-10
Warr	nings	1-10
******	Failure of operation	
	Explosion hazard	
	Electrical shock hazard	
	Patient safety	1-11
Cauti	ions	1- 12
	Airway gas measurement	1-12
	Invasive blood pressure	1-13
	Cleaning	1-13
	Disposal	1-13
Specifica	tions	1-15
Gene	eral	1-15
	Power supply	1-15
	Back-up battery	1-15
	Environmental conditions	
ECG	1-15	
	ST segment analysis	
	Heart rate	
	Impedance respiration	1-17
NIBF)	1-17

Invasive blood pressure	1-17
Pulse rate	
Temperature	1-17
Pulse oximetry, standard	1-17
SpO ₂	
Pulse rate	
Default alarm limits	
Sensor emitter wavelength ranges	
Pulse oximetry, Datex-Ohmeda enhanced (N-XOSAT option)	
<u>SpO2</u>	
Pulse rate	
Default alarm limits	
Sensor emitter wavelength ranges	
Pulse oximetry, Nellcor compatible (N-XNSAT option)	
SpO ₂ Pulse rate	
Default alarm limits	
Sensor emitter wavelength ranges	
Airway gases	
Respiration rate (RR)	
Agent identification	
Carbon Dioxide (CO ₂), Oxygen (O ₂), and Nitrous Oxide (N ₂ O)	
Anesthetic Agent (AA)	1-20
Patient Spirometry	1-20
NMT 1-21	
Stimulator	1-21
Regional block mode	1-21
Recorder	1-21
Table of Figures	
Figure 1-1. Cardiocap/5 monitor (F-MXG)	1-4
Figure 1-2. Patient connectors (F-MXG)	1-5
Figure 1-3. Cardiocap/5 rear panel (F-MXG)	1-6









Introduction

This *User's Reference Manual* contains detailed information about the features of the Datex-Ohmeda Cardiocap/5 monitor with Anesthesia software (S-XANE) installed.

Additional information

The Datex-Ohmeda Cardiocap/5 *User's Guide for Anesthesia* ships with the monitor. It contains basic information for using the Cardiocap/5.

The Datex-Ohmeda Cardiocap/5 *Technical Reference Manual* describes installation, maintenance, service, and repair procedures to be performed by authorized service personnel only.

For information about other devices closely related to the Cardiocap/5, see the Datex-Ohmeda S/5 Network and Central, User's Reference Manual

Licensed accessories are listed in the Datex-Ohmeda Supplies and Accessories Catalog.

Clinical application

This manual contains specific information about clinical and technical aspects of the Cardiocap/5. In addition, Datex-Ohmeda produces application guides that contain detailed information about clinical applications, such as patient spirometry, CO_2 monitoring, ST segment analysis, impedance respiration, etc.

Configuration

This manual describes most configurable features of the Datex-Ohmeda Cardiocap/5. Due to the possibility of different factory configurations, some menus, displays, and functions described in this manual may not be available in the monitor you are using.

When your monitor was delivered, it was configured with the default settings and ready to use. Instructions for changing settings to make your own permanent configuration are located within the appropriate chapters in this manual.

Configuration passwords

Some features can be configured only from the Install/Service menu. A password is required to access the Install/Service options. The default password is 16–4–34.

To enter the Install/Service menu:

- 1. Press the ComWheel and select Monitor Setup from the Main Menu.
- 2. Select Install/Service.
- 3. Turn the ComWheel until the first number (16) appears in the adjustment window, then press the ComWheel to select the number. Select the other password numbers (4, then 34) in the same way. When you finish, the Install/Service menu opens.

Most modifications are temporary unless you save them in the Save Modes menu, a password-protected submenu of the Install/Service menu. The default password for entering the Save Modes menu is 13–20–31.

Cardiocap/5 model and options for Anesthesia software

The Cardiocap/5 F-MXG model is used for Anesthesia software. The F-MXG is a hemodynamic model with airway gas measurement that can be equipped with built-in options.

- All measurement parameter options (and the Recorder option, N-XREC) are factory-configured and cannot be added after purchase.
- Data collection and management options (N-XNET and N-XDNET) can be added later.

Hemodynamic model with airway gas measurement (F-MXG)

The F-MXG model with Anesthesia software installed measures NIBP, ECG (3-lead and 5-lead), pulse oximetry (SpO₂), temperature (T1), impedance respiration, and airway gases. Gas measurement depends on which airway gas option is installed (N-XC, N-XCO, or N-XCAiO):

N-XC Carbon Dioxide (CO₂)

N-XCO CO_2 , N_2O , and Patient Oxygen (O_2)

N-XCAiO CO₂, anesthetic agents, agent identification, N₂O, and O₂

The F-MXG can also be equipped with each of these built-in options:

N-XP Two invasive pressure channels and second temperature (T2)

N-XV Patient Spirometry (N-XCO or N-XCAiO option required)

N-XREC Recorder

In addition, the F-MXG can be equipped with **one** of the following built-in options:

N-XOSAT Datex-Ohmeda enhanced pulse oximetry (SpO₂)
N-XNSAT Nellcor® compatible pulse oximetry (SpO₂)

N-XNMT NeuroMuscular Transmission (NMT) for relaxation measurement and

plexus stimulation (N-XCAiO option required)

Data collection and management options

These options can be factory-built or added later as upgrades:

N-XNET Network

N-XDNET Data card and Network



Order codes

Items and corresponding order codes are given in the table below:

Abbreviation	Description	Order code
F-MX	Cardiocap/5 Hemodynamic Frame	6050-0005-614
F-MXG	Cardiocap/5 Hemodynamic Frame with gas measurement	6050-0005-617
S-XANE01	Cardiocap/5 Anesthesia Software	6050-0005-615
N-XC	Airway Gas Option (CO ₂)	6050-0005-611
N-XCO	Airway Gas Option (CO ₂ , O ₂ and N ₂ O)	6050-0005-612
N-XCAIO	Airway Gas Option (CO ₂ , O ₂ , N ₂ O and anesthesia agents with automatic identification)	6050-0005-613
N-XV	Patient Spirometry Option	6050-0005-620
N-XP	Invasive Pressure Option with second temperature channel	6050-0005-940 or 6050- 0005-939
N-XREC	Recorder Option	6050-0005-941
N-XNET	Network Option	6050-0005-622
N-XDNET	DataCard and Network Option	6050-0005-700, 6050- 0005-735 or 6050-0005- 736
N-XNMT	NeuroMuscular Transmission (NMT) Option	6050-0005-914
N-NSAT	Nellcor Compatible SpO ₂ Option	6050-0005-916
N-XOSAT	Enhanced Datex-Ohmeda SpO ₂	6050-0005-917

Parts of the monitor




Figure 1-1. Cardiocap/5 monitor (F-MXG)

- (1) Power On/Standby key
- (2) External power indicator / Battery charge status LED
- (3) Alarm indicators
- (4) Insertion slots for two memory cards NOTE: A cover for the slots is available.
- (5) Direct access keys
- (6) Adjustable rear support
- (7) ComWheel
- (8) Recorder (N-XREC option)
 NOTE: The two-button recorder (shown) is for Cardiocap/5 monitors using software version 3.0 or higher. A one-button recorder was available previously.
- (9) Patient connectors
- (10) Spirometry connectors
- (11) NIBP connector
- (12) D-fend housing





Patient connectors panel





Figure 1-2. Patient connectors (F-MXG)

- (1) Spirometry (N-XV option)
- (2) NMT (N-XNMT option)
- (3) Invasive pressure, P2 (N-XP option)
- (4) Temperature, T2 (N-XP option)
- (5) Temperature, T1
- (6) Invasive pressure, P1 (N-XP option)
- (7) SpO₂ (connector differs depending on pulse oximetry configuration)



- (8) ECG
- (9) NIBP
- (10) D-fend housing

Rear panel




Figure 1-3. Cardiocap/5 rear panel (F-MXG)

- (1) Built-in handle
- (2) Gas outlet, X6
- (3) Remote Control connector, X5
- (4) Ethernet connector, X4
- (5) Network connection LEDs
- (6) Network identification plug connector, X3
- (7) Serial communication interface/local printer connector, X2
- (8) Analog/digital output connector, X1 (includes nurse call and defibrillator synchronization signals)
- (9) Mounting attachment
- (10) Dust filter
- (11) Potential equalization
- (12) Fuse and voltage information
- (13) Receptacle for mains power cord

Rear panel connections

WARNING: Electrical shock hazard. Use only hospital-grade, grounded electrical outlets.

CAUTION: Turn off the power before making any rear panel connections.

- Place the monitor on a flat surface that can hold at least 10 kg (22 lb.).
- Use only the power cord provided to connect the monitor to the wall outlet.



Symbols



Attention! Read accompanying instructions, including all warnings and cautions, before using this device.

This symbol has the following meanings when it appears on the screen:

- On the front panel indicates that protection against cardiac defibrillator discharge is due in part to the accessories for pulse oximetry (SpO₂), temperature (T) and invasive pressure (P) measurement.
- When displayed beside the O₂ value, indicates that the FiO₂ lowalarm limit is set below 21%.
- When displayed next to the HR value, indicates that there is a risk that the monitor counts pacemaker spikes (pacer is set ON R) or the monitor counts T-waves (a wide QRS is selected).



Alarm silence indicator. When displayed at the upper left corner of the screen, indicates that all alarms are silenced. When in a menu or digit field, indicates that the alarm source has been turned off.



Main Menu. Located beside the ComWheel to indicate you can open the Main Menu by pressing the ComWheel when no other menu is displayed.



Submenu. Appears in a menu to indicate the selection will open a new menu. A submenu contains functions that are used less frequently.



The monitor is connected to the Datex-Ohmeda Network.



Data card (green) and/or the Menu card (white) is inserted.



Flashes next to the heart rate value or pulse rate value to indicate the beats detected.



Appears next to the respiration rate value to indicate that the respiration rate is calculated from the impedance respiration measurement.



Battery operation and remaining capacity.



Battery charging.



Type BF applied part (IEC 60601-1). Defibrillator-proof protection against electric shock.



Type CF applied part (IEC 60601-1). Defibrillator-proof protection against electric shock.



Gas outlet (in airway gas models only).



Ethernet connectors.



Equipotentiality. Monitor can be connected to potential equalization conductor.



Alternating current.



Fuse.



Power On/Standby.





ESD warning symbol for electrostatic sensitive devices. Pins of connectors identified with the ESD warning symbol should not be touched. Connections should not be made to these connectors unless ESD precautionary procedures are used. See "Safety precautions: ESD precautionary procedures" for details.



Symbol for non-ionizing electromagnetic radiation. Interference may occur in the vicinity of equipment marked with the symbol.

Delayed standby function

A delayed switch-off mechanism makes it possible to return to monitoring if the monitor is accidentally switched to standby.

When the **Power On/Standby** key is pressed, a message box is displayed in the center of the waveform display and the waveform fields are frozen. After eight seconds, the monitor display is switched off and the monitor enters standby mode.

Pressing the **Power On/Standby** key during this 8-second delay will allow you to continue monitoring. The message box will be removed and all data and configurations will be retained.

Back-up battery

If mains power is lost during monitoring, the Cardiocap/5 can run on power from the internal battery. The capacity of a fully charged internal battery is 15 minutes minimum.

NOTE: The monitor can be started only when connected to mains power.

Battery time

A symbol on the display indicates the remaining operating time of the battery. It is displayed only when the monitor is not connected to mains power.



The operating time that remains is updated every minute.

- When about five minutes of battery operating time is left, a "Batt low" alarm is activated and the monitor makes one beep.
- When about one minute of operating time is left, a "Batt empty" alarm is activated.

The battery is charged automatically when the monitor is connected to mains power. Battery charging time is normally five hours.

A charging symbol is displayed during charging:



The symbol is removed from the display when the battery is fully charged.

During battery charging, the standby LED is flashing in standby mode.



Installation and service

A separate "Technical Reference Manual" describes installation, interfacing, connectors, service, maintenance and reparation procedures of the monitor. Medical electrical equipment needs special precautions regarding EMC and needs to be installed and put into service according to the EMC information provided in the "Technical Reference Manual" by qualified personnel. Service and reparations are allowed for authorized service personnel only.

WARNING: Before starting to use the system, ensure that the whole combination complies with the international standard IEC 60601-1-1 and with the requirements of the local authorities.

Do not connect any external devices to the system other than those specified.

Safety precautions

This section lists the warnings and cautions you should know before using the monitor.

Warnings

A WARNING indicates a situation in which the user or the patient may be in danger of injury or death.

Failure of operation

It is possible for any device to malfunction; therefore, always verify unusual data by performing a formal patient assessment.

The monitor or its components should not be used adjacent to or stacked with other equipment. If adjacent or stacked use is necessary, the monitor and its components should be observed to verify normal operation in the configuration in which it will be used.

Explosion hazard

To avoid an explosion hazard, do not use the monitor in the presence of flammable anesthetics. The monitor measures only non-flammable anesthetics.

Electrical shock hazard

When you connect equipment to the Cardiocap/5 input and output connectors, you are configuring a medical system and are responsible for ensuring that the system complies with IEC/EN 60601-1-1 and with local requirements.

Do not connect external equipment to the system, except equipment specified by Datex-Ohmeda.

Make sure external equipment is hospital-grade and grounded before connecting. Use only hospital-grade, grounded electrical outlets and power cords.

Pins of connectors identified with the ESD warning symbol should not be touched. Connections should not be made to these connectors unless ESD precautionary procedures are used. See "ESD precautionary procedures" later in this section (Users Guide: See "Safety precautions: ESD precautionary procedures" in the "User's Reference Manual) for details.

NMT

Always stop NMT measurement before handling stimulation electrodes.

Do not place the NMT stimulation electrodes on the patient's chest.

Cleaning

Before cleaning, disconnect the monitor from the electrical outlet.

After cleaning, make sure that every part of the monitor is dry before reconnecting it to the power supply.

If liquid has accidentally entered the equipment, disconnect the power cord from the power supply and have the equipment serviced by authorized service personnel.



Patient safety

All invasive procedures involve risks to the patient. Use aseptic technique. Follow the instructions provided by the catheter manufacturer.

The output signals are not floating and they must not be connected directly to a patient.

Connect only one patient to one monitor at a time.

Constant attention by a qualified professional is needed whenever a patient is under anesthesia or connected to a ventilator. Some equipment malfunctions may pass unnoticed in spite of the monitor alarm.

Do not use antistatic or electrically-conductive breathing tubes. They may increase the risk of burns when an electrosurgery unit is used.

Do not use the Cardiocap/5 during magnetic resonance imaging (MRI).

Use only approved accessories, mounts and defibrillator-proof cables and invasive pressure transducers. For a list of approved supplies and accessories, see the "Supplies and Accessories" catalog. Other cables, transducers and accessories may cause a safety hazard, damage the equipment or system, result in increased emissions or decreased immunity of the equipment or system or interfere with the measurement. Protection against cardiac defibrillator discharge is due in part to the accessories for pulse oximetry (SpO₂), temperature (T) and invasive pressure (P) measurement. Single-use accessories are not designed to be re-used. Re-use may cause a risk of contamination and affect the measurement accuracy.

To prevent erroneous readings, do not use physically damaged sensors or sensor cables. Discard a damaged sensor or sensor cable immediately. Never repair a damaged sensor or cable; never use a sensor or cable repaired by others. A damaged sensor or a sensor soaked in liquid may cause burns during electrosurgery.

ECG, Impedance respiration, Invasive blood pressure

Ensure proper contact of the return electrode of the electrosurgery unit to your patient to avoid possible burns at sensor sites.

ECG, Impedance respiration, NMT

Make sure that the lead set clips or snaps do not touch any electrically conductive material, including earth.

Impedance respiration

In obstructive apnea, respiration movements and impedance variations may continue.

NIBP

The monitor sets the inflation pressure automatically according to the first measurement. Reset the case to reset the inflation limit before measuring a new patient.

Invasive blood pressure

Make sure that no part of the patient connections touches any electrically conductive material, including earth.

Patients with pacemakers or arrhythmia

The monitor may count pacemaker pulses as heart beats during cardiac arrest, some arrhythmias, and with certain types of pacemakers particularly in ONR mode. Do not rely entirely upon rate meter alarms. Keep patients with pacemakers and arrhythmias under close surveillance.

The impedance respiration measurement may cause rate changes in Minute Ventilation Rate Responsive Pacemakers. Set the pacemaker rate responsive mode to off or turn off the impedance respiration measurement on the monitor.

Gases

MAC values are empirical, not absolute. Cardiocap/5 MAC values correspond to those of healthy adults and cannot be applied to children. Age and other individual factors influencing the effect of volatile agents are not taken into account.

Pulse oximetry

Use clean and dry sensors and cables only. Moisture and debris on connectors may affect measurement accuracy.

Patient conditions (such as reddening, blistering, skin discoloration, ischemic skin necrosis, and skin erosion) may warrant changing the sensor site frequently or using a different style of sensor. For details, refer to the instructions supplied with the sensor.

Conditions that may cause inaccurate readings and impact alarms include interfering substances, excessive ambient light, electrical interference, ventricular septal defects (VSD), excessive motion, low perfusion, low signal strength, incorrect sensor placement, poor sensor fit, and/or movement of the sensor on the patient.

To prevent erroneous readings, do not use an inflated blood pressure cuff or arterial blood pressure measurement device on the same limb as the oximeter sensor.

Temperature

To prevent patient injury, use Datex-Ohmeda temperature probes only.

Alarms

Make sure that necessary alarm limits are set and alarms are on when you start monitoring a patient.

When alarms are suppressed, observe the patient frequently.

Cautions

A CAUTION indicates a situation that may cause damage to the unit or devices connected to it..

US Federal law restricts this device to sale by or on the order of a licensed medical practitioner.

The system is intended for use by qualified medical personnel only.

Do not apply pressurized air to any outlet or tubing connected to the monitor. Pressure may destroy sensitive elements.

Do not store the monitor outside the specified temperature range (-10 to + 50 $^{\circ}$ C / 14 to 122 $^{\circ}$ F).

Do not subject memory cards to excessive heat, bending, or magnetic fields.

Leave space for circulation of air to prevent the monitor from overheating.

Turn off the power before making any rear panel connections.

Use only cables and accessories approved by Datex-Ohmeda. Other cables and accessories may damage the system or interfere with measurement. Single-use accessories are not designed to be reused.

Vibrations during transport may disturb SpO_2 , ECG, impedance respiration, and NIBP measurements.

Airway gas measurement

Strong scavenging suction may change the operating pressure of the monitor and cause inaccurate readings or internal damage.

Invasive blood pressure

Mechanical shock to the invasive blood pressure transducer may cause severe shifts in zero balance and calibration and may produce erroneous readings.

Cleaning

Do not disinfect or open the water trap cartridge. To avoid damage, do not touch or try to clean the water trap hydrophobic membrane (other than to rinse it with water). Do not sterilize any part of the monitor with steam autoclave or ethylene oxide. Do not immerse any part of the device in liquids or allow liquid to enter the interior. Do not use hypochlorite, acetone-based, phenol-based, or ammonia-based cleaners.

Disposal

Dispose of the whole device and its parts in accordance with local environmental and waste disposal regulations.

Points to note

Medical electrical equipment needs special precautions regarding electromagnetic compatibility and needs to be installed and put into service according to the electromagnetic compatibility information provided in the "Technical Reference Manual" by qualified personnel.

Portable and mobile RF communications equipment can affect the medical electrical equipment.

The allowed cables, transducers and accessories for the system are listed in the "Supplies and Accessories" catalog delivered with the monitor.

The equipment is suitable for use in the presence of electrosurgery. Please notice the possible limitations in the parameter sections and in the "Safety precautions" section.

Service and reparations are allowed for authorized service personnel only.

CISPR 11 classifications:

- Group 1 contains all ISM (Industrial, scientific and medical) equipment in which there is intentionally generated and/or used conductively coupled radio-frequency energy which is necessary for the internal functioning of the equipment itself.
- Class A equipment is suitable for use in all establishments other than domestic and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.

ESD precautionary procedures

To avoid electrostatic charges to build up, it is recommended to store, maintain and use the components at a relative humidity of 30% or greater. Floors should be covered by ESD dissipative carpets or similar. Non-synthetic clothing should be used when working with the component.

To prevent applying a possible electrostatic discharge to the ESD sensitive parts of the component, one should touch the metallic frame of the component or to a large metal object located close to the component. When working with the component and specifically when the ESD sensitive parts of the component may be touched, a grounded wrist strap intended for use with ESD sensitive equipment should be worn. Refer documentation provided with the wrist straps for details of proper use

ESD precautionary procedure training

It is recommended that all potential users receive an explanation of the ESD warning symbol and training in ESD precautionary procedures.

The minimum content of an ESD precautionary procedure training should include an introduction to the physics of electrostatic charge, the voltage levels that can occur in normal practice and the damage that can be done to electronic components if they are touched by and operator who is electrostatically charged. Further, an explanation should be given of methods to prevent build-up of electrostatic charge and how and why to discharge one's body to earth or to the frame of the equipment or bond oneself by means of a wrist strap to the equipment or the earth prior to making a connection.



Specifications

All specifications are subject to change without notice.

WARNING: Operation of the monitor outside the specified values may cause inaccurate results.

General

Power supply

Rated voltages and frequencies: 100-240 V 60/50 Hz

Allowed voltage fluctuations: $\pm 10\%$ Maximum power consumption: 80 VA

Fuses (2): T2AH/250V

Back-up battery

Type: 12V 2.6AH lead acid

Back-up battery time: at least 15 minutes when fully charged

Charging time (typical): 5 hours

Charging indicator: Green LED On: full charge, battery on the holding voltage

Green LED flashing: charging

Environmental conditions

Operating temperature: +10 to +40°C (50 to 104°F)

Storage and transport temperature: -10 to +50°C (14 to 122°F)

Relative humidity: 0 to 85% noncondensing; in airway 0 to 100% condensing

Atmospheric pressure: 660 to 1060 hPa (500 to 800 mmHg)

ECG

Waveform display (with 50 Hz power supply frequency):

Monitoring filter 0.5 to 30 Hz

ST filter 0.05 to 30 Hz

Diagnostic filter 0.05 to 100 Hz

Waveform display (with 60 Hz power supply frequency):

Monitoring filter 0.5 to 40 Hz

ST filter 0.05 to 40 Hz

Diagnostic filter 0.05 to 100 Hz

Minimizing the effects of the line isolation monitor transients: Crystal controlled oscillator is used as the operating frequency source of the patient isolation power supply.

Offset voltage range: ±0.4 V

WARNING. The ± 0.4 V offset voltage range of the ECG measurement may be insufficient to handle the offset potentials when using ECG electrodes of dissimilar metals

The isolation barrier capacitance has been minimized to reduce the hazard of burns in the event of a defect in the ESU return electrode connection.

Direct current for leads-off detection through any patient electrode : ≤50 nA

The normalized respiration sensing current : \leq 3.0 μ A

Frequency of respiration sensing current: 31.25 kHz

Maximum Tall T wave amplitude that does not disturb the heart rate calculation time (according to ANSI/AAMI EC13 4.1.2.1): 10 mV

Pacemaker pulse detection:

detection level: 2 to 700 mV pulse duration: 0.5 to 2 ms

The monitor is specified for both of the methods A and B required in ANSI/AAMI EC13 4.1.4.2.

Pacer pulse rejection of fast ECG signals:

0.2 V/s with Sensit pacemaker selection and 1.6 V/s with other selections according to the test defined in ANSI/AAMI EC13 section 4.1.4.3.

ST segment analysis

Measured and displayed simultaneously for up to three ECG leads

ST level range: -6 to +6 mm (-0.6 to +0.6 mV)

Display resolution: 0.1 mm (0.01 mV)

Averaging: Calculated from 16 QRS complexes

Display update interval: 5 seconds

Test result of ST segment measurement algorithm testing.

The algorithm testing has been performed by using The European Society of

Cardiology ST-T Database

Average results from ischemic ST detection

Episode sensitivity: 77

Episode positive predictive accuracy: 67

Duration sensitivity: 74

Duration positive predictive accuracy: 61

Heart rate

Measurement range: 30 to 250 bpm Measurement accuracy: \pm 5% or \pm 5 bpm Pacemaker pulse detection level: 2 to 500 mV Pacemaker pulse duration: 0.5 to 2 ms

Display averaging time: 10 seconds
Display update time: 5 seconds

The heart rate calculation operates with irregular rhythms of ANSI/AAMI EC13 4.1.2.1

(e) as follows: a): 40 bpm b): 87 bpm c): 60 bpm d): 117 bpm

Average heart rate response time and time range of response time (according to

ANSI/AAMI EC13 4.1.2.1 (f)):

Response time 80 to 120 bpm: 5.0s (3.7 to 6.2 s)

Response time 80 to 40 bpm: 6.5s (4.1 to 9.2 s)

The average time and time range () to alarm for tachycardia are as follows ANSI/AAMI

EC13 4.1.2.1 (g)): Figure 4a halved amplitude: 5.5 s (4.8 to 5.9 s)

Figure 4a normal amplitude: 6.6 s (4.8 to 7.2 s)

Figure 4a doubled amplitude: 7.2 s (5.3 to 9.4 s)

Figure 4b halved amplitude: $7.4 ext{ s}$ ($7.0 ext{ to } 7.7 ext{ s}$)

Figure 4b normal amplitude: 6.2 s (4.7 to 8.1 s)

Figure 4b doubled amplitude: 6.4 s (4.4 to 8.2 s)

Auxiliary ECG output:

Bandwidth of auxiliary output: 0.5 to 40Hz

Gain: 1 mV ECG signal is 1 V at the auxiliary output.

Propagation delay: < 15 ms

The pacemaker pulses are absent but may cause interference at the auxiliary ECG output.

An auxiliary device that fulfils the requirements of the IEC 60601-1 standard can be connected to the auxiliary output. There are no other limitations, because the auxiliary output of the monitor is galvanically isolated from patient applied part of the ECG measurement.

Impedance respiration

Respiration range: 4 to 120 resp/minute Accuracy: \pm 5% or \pm 5 resp/minute

NIBP

Measurement range:

Adult 25 to 260 mmHg; Child 25 to 195 mmHg; Infant 15 to 145 mmHg

Pulse rate range accepted: 30 to 250 bpm

Typical measuring time: Adult 23 seconds; Infant 20 seconds

Invasive blood pressure

Measurement range: -40 to 320 mmHg Measurement accuracy: $\pm 5\%$ or ± 2 mmHg

Transducer sensitivity: 5 µV/V/mmHg, 5 VDC, max 20 mA

Pulse rate

Measurement range: 30 to 250 bpm Accuracy: $\pm 5\%$ or ± 5 bpm

Temperature

Measurement range: 10 to 45°C (50 to 113°F)

Measurement accuracy:

25 to 45.0 °C \pm 0.1 °C (77 to 113 °F \pm 0.2 °F) 10 to 24.9 °C \pm 0.2 °C (50 to 76.8 °F \pm 0.4 °F)

Probe type: Datex-Ohmeda only

Pulse oximetry, standard

Display update time: 5 seconds Averaging time: adjustable

Plethysmographic waveform scaling: adjustable

Sp₀2

Calibration range: 50 to 100%

Calibrated against functional saturation

Measurement range: 40 to 100%

Measurement accuracy (% SpO₂ ± 1SD):

80 to 100% ± 2 digits; 50 to 80% ± 3 digits; Below 50% unspecified NOTE: SpO₂ measurement accuracy is based on deep hypoxia studies using Datex-Ohmeda FingerSat sensors on volunteers. Arterial blood samples were analyzed by a Radiometer OSM CO-oximeter. Refer to the sensor instructions for specific accuracy data.

Pulse rate

Measurement range: 30 to 250 bpm Measurement accuracy: \pm 5% or \pm 5 bpm

Default alarm limits

SpO₂: high Off, low 90% Pulse rate: high 160, low 40 NOTE: Limits are adjustable.

Sensor emitter wavelength ranges

Red LED: 660 nm Infrared LED: 900 nm

Pulse oximetry, Datex-Ohmeda enhanced (N-XOSAT option)

Display update time: 5 seconds Averaging time: 12 seconds

Plethysmographic waveform scaling: automatic

Sp02

Calibration range: 70 to 100%

Calibrated against functional saturation

Measurement range: 1 to 100% Measurement accuracy (± 1SD):

70 to $100\% \pm 2$ digits

70 to $100\% \pm 3$ digits during conditions of clinical patient motion

Below 70% unspecified

NOTE: SpO_2 measurement accuracy is statistically derived and correlated to simultaneous arterial blood gases measured on a Radiometer OSM3 CO-oximeter. Refer to the sensor instructions for specific accuracy data.

Pulse rate

Measurement range: 30 to 250 bpm

Measurement accuracy: $\pm 2\%$ or ± 2 bpm (whichever is greater)

Default alarm limits

SpO₂: high Off, low 90% Pulse rate: high 160, low 40 NOTE: Limits are adjustable.

Sensor emitter wavelength ranges

Red LED: 650 to 665 nm Infrared LED: 930 to 950 nm Average power: ≤ 1 mW

Pulse oximetry, Nellcor compatible (N-XNSAT option)

Display update time: 5 seconds

Averaging time: 5 to 7 seconds

Plethysmographic waveform scaling: automatic

Sp02

Calibrated against functional saturation

Measurement range: 1 to 100%

Measurement accuracy (% SpO₂ ±1SD):

70 to 100% (\pm 2 digits to \pm 3.5 digits, depending on the sensor)

Below 70% unspecified

See the Pulse Oximetry chapter for a list of approved sensors and accuracy details.

NOTE: SpO₂ measurement accuracy is based on testing healthy adult volunteers in induced hypoxia studies.

Pulse rate

Measurement range: 30 to 250 bpm Measurement accuracy: ±3 digits

Default alarm limits

SpO₂: high Off, low 90% Pulse rate: high 160, low 40 NOTE: Limits are adjustable.

Sensor emitter wavelength ranges

Red LED: 660 nm Infrared LED: 920 nm

Airway gases

Sampling rate (typical value): 200 ml/minute

Sampling delay: 2.5 seconds typical with a 3 meter sampling line

Total system response time: 2.9 seconds typical with a 3 meter sampling line, including

sampling delay and rise time

Warm-up time: 2 to 5 minutes, 30 minutes for full spec.

Default alarm limits:

NOTE: Limits are adjustable within the measurement range. Alarm limits and

their adjustment range may vary depending on the mode used.

EtCO₂ high 8%, low 3% FiCO₂ high 3%, low Off EtO₂ high Off, low 10% FiO₂ high Off, low 18% FiN₂O high 82% FiHal high 2.2%, low Off

EtEnf high 3.4%, low Off Filso high 3.4%, low Off EtIso high 2.3%, low Off FiDes high 18%, low Off EtDes high 12%,low Off FiSev high 5.1%, low Off

EtSev high 3.4%, low Off

FiEnf high 5.1%, low Off

EtHal high 1.5%, low Off

Ethanol C₂H₅OH (< 0.3%) Methane CH₄ (< 0.2%) Acetone (< 0.1%) Nitrogen N₂

Carbon monoxide CO

Nitric oxide NO (< 200 ppm)

Water vapor

Non-disturbing gases:

Maximum effect on readings

 CO_2 : < 0.2 vol%

 $O_2, N_2O: < 2 \text{ vol}\%$

Anesthetic agents: < 0.15 vol%

Effect of helium: Decreases CO2 readings < 0.6 vol% typically

Respiration rate (RR)

Measurement range: 4 to 60 breaths/minute Detection criteria: 1% variation in CO₂

Agent identification

Identification threshold (typical value): 0.15 vol%

Carbon Dioxide (CO_2), Oxygen (O_2), and Nitrous Oxide (N_2O)

Measurement	Carbon Dioxide (CO ₂)	Oxygen (O ₂)	Nitrous Oxide (N ₂ O)
Range	0 to 15%, (0 to 15 kPa) (0 to 113 mmHg)	0 to 100%	0 to 100%
Accuracy (typical)	≤ 0.3 vol%	≤ 2 vol%	≤ 3 vol%
Rise time	< 400 ms typical	< 400 ms typical	< 450 ms typical
Gas cross effects	< 0.2 vol% (O ₂ , N ₂ O, and anesthetic agents)	<pre>< 1 vol% (anesthetic agents); < 2 vol% (N2O)</pre>	< 2 vol% (anesthetic agents)

Anesthetic Agent (AA)

Measurement	Halothane, isoflurane, Enflurane	Sevoflurane	Desflurane
Range	0 to 6%	0 to 8%	0 to 20%
Accuracy (typical)	≤ 0.2 vol%	≤ 0.2 vol%	0 to 5% ≤ 0.2 vol% 5 to 10% ≤ 0.5 vol% 10 to 20% ≤ 1.0 vol%
Rise time	< 400 ms typical		
Gas cross effects	< 0.15 vol% N ₂ O		

Patient Spirometry

Detection through D-lite[™] or Pedi-lite[™] flow sensor and gas sampler:

Measurement	D-lite flow sensor	Pedi-lite flow sensor
Tidal volume Measurement range Accuracy (typical value)	150 to 2000 ml ± 6% or 30 ml	15 to 300 ml ± 6% or 4 ml
Minute volume Measurement range Accuracy (typical value)	2 to 20 l/minute ± 6%	0.5 to 5 l/minute ± 6%
Airway pressure Measurement range Accuracy (typical value)	-20 to +100 cmH ₂ O ± 1 cmH ₂ O	–20 to +100 cmH₂O not applicable
Flow Measurement range	1.5 to 100 l/minute	0.25 to 25 l/minute
Compliance		

Measurement range	4 to 100 ml/cmH ₂ O	4 to 100 ml/cmH ₂ O
Airway Resistance Measurement range	0 to 40 cm H ₂ O/l/second	0 to 40 cm H ₂ O/l/second
Sensor specifications		
Dead space	9.5 ml	2.5 ml
Resistance at 30 l/minute	0.5 cmH ₂ O	not applicable
Resistance at 10 l/minute	not applicable	1.0 cmH ₂ O

NMT

Stimulation modes:

Train of four (TOF)

Double burst (3.3) (DBS)

Single twitch (ST)

50 Hz tetanic + post-tetanic count (PTC)

Measurement intervals:

TOF and DBS: Manual; 10 seconds, 12 seconds, 15 seconds, 20 seconds, 1 minute,

5 minutes, 15 minutes

ST: Manual; 1 second, 10 seconds, 20 seconds

Stimulator

Stimulus pulse: Square wave, constant current

Pulse width: 100, 200 or 300 μs

Stimulus current range (supramax and manual): 10 to 70 mA with 5 mA steps

Stimulus current accuracy: 10% or ± 3mA (whichever is greater)

Max load: $3 \text{ k}\Omega$ Max voltage: 300 V

Regional block mode

Stimulation mode: Single twitch (ST) Intervals: 1 second, 2 seconds, 3 seconds Stimulus pulse: Square wave, constant current

Pulse width: 40 µs

Stimulus current range: 0 to 5.0 mA with 0.1 mA steps

Stimulus current accuracy: 20% or 0.3 mA (whichever is greater)

Recorder

Principle: thermal array

Print resolution:

Vertical: 8 dots/mm (200 dots/inch)

Horizontal: 32 dots/mm (800 dots/inch) at speed of 25 mm/second and slower

Paper width: 50 mm; printing width 48 mm

Traces: selectable; 1, 2, or 3 traces

Print speed: 1, 6.25, 12.5, 25 mm/second


For your notes





Chapter 2. Monitoring Basics

Introduction	2-1
Keys and controls	2-1
Key functions	2-2
Remote control	2-3
Remote control keys	
Using menus	2-4
Opening the Main Menu	
Using the direct access keys	2-5
Choosing menu options	2-5
Returning to the normal display	2-5
Monitoring a patient	2-6
Preparations	2-6
Starting monitoring	2-6
During monitoring	2-7
Ending monitoring	2-7
Main Menu map	2-8
Direct access key maps	2-11
Abbreviations	2-12
able of Figures	
Figure 2-1. Cardiocap/5 keys and controls	
Figure 2-2. Remote Control	
Figure 2-3. Parts of a menu	
Figure 2-4. Main Menu	2-4





2. MONITORING BASICS

Introduction

You can control Cardiocap/5 monitoring with the ComWheel, the direct access keys, or the optional Remote Control.

- · The ComWheel provides access to all monitor functions.
- The direct access keys provide direct access to frequently-used functions that are also available through the Main Menu.
- The Remote Control Menu key lets you enter all monitoring functions. Each Remote Control direct access key lets you enter a common function.

See Main Menu map and Direct access key maps later in this chapter for the menus you can access from the Main Menu and by pressing the direct access keys.

Keys and controls




Figure 2-1. Cardiocap/5 keys and controls

- (1) Power On/Standby key
- (2) Silence Alarms key
- (3) Trends direct access key
- (4) Pulse Oximetry, Invasive Pressures, or NMT direct access key (depending on configuration)
- (5) ECG direct access key
- (6) NIBP direct access key
- (7) NIBP Start/Cancel key
- (8) Normal Screen key
- (9) ComWheel

Key functions



Power On/Standby key.

NOTE: The monitor will start only when connected to mains power.



Silences an active alarm or pre-silences all alarms for two minutes (press the key once) or five minutes (press the key for three seconds). Pressing the key once again clears the alarm message field of all alarm messages and enables new alarms. A lighted yellow LED indicates a yellow alarm; a lighted red LED indicates a red alarm.



Displays trends (numerical and graphical) and snapshots. Press the **Normal Screen** key to return to monitoring mode.

The next oval-shaped key can be **Pulse Oximetry, Invasive Pressures,** or **NMT**, depending on which built-in options your monitor contains.



Displays the menu for adjusting Oxygen saturation measurement settings (all models with no Invasive Pressure or NMT option)



Displays the menu for adjusting Invasive Pressure measurement settings (all models with the Invasive Pressure option and no NMT option)



Displays the menu for adjusting NMT measurement settings (all models with the NMT option).



Displays the menu for adjusting ECG measurement settings.



Displays the menu for adjusting NIBP measurement settings.



Starts or cancels a single non-invasive blood pressure measurement, STAT, and manual measurements; stops venous stasis.



Returns to the normal monitoring display.



The **ComWheet** is the main tool for all menu functions. You turn and/or press it to access, navigate, and choose all menu options.



Pushing the ComWheel while no menu is displayed opens the Main Menu from which you have access to all monitor functions.

Recorder keys

See the *Recording and Printing* chapter for descriptions of the keys on the optional recorder.





Remote control

The optional Remote Control (K-CREMCO) contains keys that help you in most common situations and tasks. You can use these direct access keys to start or end a function.




Figure 2-2. Remote Control

Remote control keys

Key	Function
Take Snapshot	Creates a freeze frame, which is saved to the monitor memory. Pressing the key also marks a number beside the numerical trend.
Menu	Opens the Main Menu from which you can access all menus and functions of the monitor.
Silence Alarms	Silences an active alarm or pre-silences all alarms for two or five minutes.
Zero ALL Pressures	Zeroes all invasive pressure channels.
Freeze	Freezes the waveform sweep for 60 seconds. To unfreeze the waveforms, press the Freeze key again (or press the Normal Screen key).
Record Wave	Records the selected real-time waveforms.
Other Patients	If the monitor is connected to the network, it is possible to view alarms and general views for another site.
Start NIBP	Starts a single NIBP measurement and cancels any measurement.
Normal Screen	Closes open menus and returns to normal monitoring display.
Start Wedge	Displays the customized wedge view and starts wedge measurement.
Special	Not in use with Cardiocap/5.
Start C.O.	Not in use with Cardiocap/5.

Using menus

A menu is a list of functions or commands displayed on the screen.




Figure 2-3. Parts of a menu

- (1) Header of the menu
- (2) Indicates the present selection
- (3) Adjustment window with other options
- (4) List of menu selections
- (5) Entry indicator to submenus
- (6) Short instructions

Opening the Main Menu

If you press the ComWheel when no menu is displayed, you will open the Main Menu.




Figure 2-4. Main Menu

The Main Menu provides access to all other menus. To display a different menu, such as the Parameters menu, select it from the Main Menu.


You can enter the Main Menu from other menus by using the ComWheel to select the previous menu until you reach the Main Menu.

To exit the Main Menu, select **Normal Screen** or press the **Normal Screen** key.






The direct access keys allow you to bypass the Main Menu and go directly to the menu activated by the direct access key. For example:



Choosing menu options

Once a menu is displayed, you use the ComWheel to make selections and adjustments. For example, to make adjustments in the ECG menu:



Returning to the normal display

Press the Normal Screen key to exit a menu and return to the normal monitoring display.

Monitoring a patient


Use only supplies and accessories approved by Datex-Ohmeda for use with the Cardiocap/5. See the Datex-Ohmeda Supplies and Accessories Catalog.

Preparations

- 1. Check that the monitor and patient accessories are clean and intact.
- 2. Press the Power On/Standby key to turn on the monitor.

NOTE: The monitor will start only when connected to mains power.

- 3. Tilt the monitor to the optimal viewing angle:
 - Press the center of the foot and adjust.
 - · Make sure both feet are at the same angle.



NOTE: If wall mounting is used, make sure the monitor's front and back attachment bars fit tightly to the edges of the mounting plate and the locking bolt in the back of the monitor locks into place.

- 4. If necessary, change the operating mode:
 - Press the ComWheel and select Monitor Setup from the Main Menu.
 - Select Select Mode.

The operating mode defines what is displayed on the screen and in the trends. Note that changing the mode also changes, for example, the alarm limits.

Starting monitoring

- 1. Prepare the patient connections. Refer to the applicable measurement parameter chapter(s) later in this manual.
- 2. Alarms are operative and the parameter default settings are active when the patient is connected to the monitor. To review alarm limits, press the ComWheel and select **Alarms Setup** from the Main Menu.

You can also select Alarms Setup through the corresponding parameter menu.

- 3. Check that you have the desired waveforms and digits in the fields. To adjust them:
 - Press the ComWheel and select Monitor Setup from the Main Menu.
 - Select Screen Setup.
 - Select Waveform Fields or Digit Fields.



- 4. Carry out measurement-specific start-up. For example, start NIBP measurement to get the reference values.
- Enter patient data (press the ComWheel and select Patient Data from the Main Menu).

During monitoring

If you need to suppress alarms, press the Silence Alarms key.


- To silence the alarms for two minutes, press the key once.
- To silence the alarms for five minutes, press the key for more than three seconds.

If you accidentally press the **Power On/Standby** key during monitoring, the "Switching to standby mode in X seconds" message is displayed. If the **Power On/Standby** key is not pressed again within eight seconds, the monitor will switch to standby mode.

To return to monitoring, press the Power On/Standby key again within eight seconds.

Ending monitoring


6. Press the ComWheel to open the Main Menu.



- 7. To print necessary information, select Record/Print. Wait until printing is finished.
- 8. To clear patient data and return settings to their defaults:
 - Select Reset Case in the Main Menu.
 - Select Reset ALL and YES.
- 9. Turn the power switch to standby if the monitor will not be used.
- 10. Clean the monitor according to the instructions.


Main Menu map






* Menu is displayed when the NMT Regional Block Adapter is connected to monitor.





Reset Case 🕨











Direct access key maps



^{*} Menu is displayed when the NMT Regional Block Adapter is connected to monitor.

Abbreviations



AA anesthetic agent
ABP arterial blood pressure
AirW airway temperature

APN apnea Art arterial Asy asystole

ATMP atmospheric pressure

ATPD ambient temperature and pressure, dry gas AVF augmented voltage foot—usually left (ECG)

AVL augmented voltage left (ECG)
AVR augmented voltage right (ECG)

Axil axillary temperature

B-To-B beat to beat

Blad bladder temperature
BP blood pressure
BSA body surface area

BTPS body temperature and pressure, saturated gas

°C degrees Celsius
Casc. cascaded (ECG)
CO₂ carbon dioxide
Compl compliance

Cont. continuous recording
Contrl controlled ventilation
Core central temperature
CVP central venous pressure

DBS double burst stimulation (NMT)

Des desflurane
Dia diastolic pressure
Diagn diagnostic (ECG filter)

DIFF difference Dyn dynamic

E expiratory

ECG electrocardiography
ECG1 first ECG waveform (top)

ECG1/r real time ECG

ECG2 second ECG waveform ECG3 third ECG waveform

EMC electromagnetic compatibility
EMG electromyography (NMT)

Enf enflurane

ESD electrostatic discharge
Eso esophageal temperature
ET, Et end tidal concentration





Exp expired

OF degrees Fahrenheit

FI, Fi fraction of inspired gas

FiO₂ fraction of inspired oxygen

Flow airway gas flow

FVloop flow volume loop

Graph. graphical

Hemo hemodynamic
Hal halothane
Hgb hemoglobin
HR heart rate

I inspiratory

ICP intra cranial pressure

ID identification
Imped. impedance
indep. independent
Inv. invasive
Iso isoflurane

LAP left atrial pressure LED light emitting diode

MAC minimum alveolar concentration

Mean mean of diastolic and systolic pressures

Monit monitoring (ECG filter)

MV minute volume

MYO myocardial temperature

N₂O nitrous oxide

Naso nasopharyngeal temperature

Net network

NIBP non-invasive blood pressure NMT neuromuscular transmission

NTPD normal temperature and pressure, dry gas

Num. numerical trends

 O_2 oxygen Oxy oxygen

P1, P2 invasive pressure channels
PA pulmonary arterial pressure
PaCO₂ arterial carbon dioxide level

Pedi pediatric
Paw airway pressure

PCWP pulmonary capillary wedge pressure
PEEP_{tot} positive end expiratory pressure
Pleth plethysmograph waveform
Ppeak airway peak pressure

Pplat airway plateau (pause) pressure

PR pulse rate Prev. previous

PTC post tetanic count (NMT)
PVloop pressure volume loop

PVC premature ventricular contraction

RAP right atrial pressure
Raw airway resistance
Rect rectal temperature
Resp respiration rate
Room room temperature
RR respiration rate

RVP right ventricular pressure

S/D systolic/diastolic

SaO₂ arterial oxygen saturation

Sev sevoflurane Skin skin temperature

Snaps. snapshot Spiro spirometry

SpO₂ oxygen saturation measured by pulse oximeter

Spont spontaneous breathing

ST single twitch stimulation (NMT)

ST ST segment of ECG

ST inf. ST segment on inferior leads (II, III, aVF)
ST lat. ST segment on lateral leads (I, aVL, V5)
STAT continuous NIBP cuff inflation for 5 minute

Stfilt ST filter (ECG)

STPD standard temperature and pressure, dry gas

Surf skin temperature

Sw software

Sys systolic pressure

T1, T2 temperature

T1% first stimulus as a percentage of the reference value (NMT)

Tab. tabular trends Temp temperature

TOF train of four stimulation(NMT)

TOF% ratio of the fourth response to the first response (NMT)

TV tidal volume

Tymp tympanic temperature

V precordial (chest) lead (ECG)

Vol volume

Chapter 3. Monitor Setup

Overview	3-1
Selecting a user mode	3-1
Changing and saving configuration settings	3-2
Parts of the display	3-2
Display setup	
Modifying waveform fields	3-4
Changing sweep speeds	3-5
Modifying digit fields	3-5
Splitting the screen display	3-7
Minitrend view	3-8
Setting time and date	3-8
Display brightness	3-9
Displaying pulse rate	3-9
Other adjustable features	3-10
Changing parameter colors	
Table of Figures	
Figure 3-1. Parts of the display	
Figure 3-2. Display fields	
Figure 3-3. Waveform field	
Figure 3-4. Digit fields Figure 3-5. Split screen view	





3. MONITOR SETUP

Overview

The Cardiocap/5 has setup options for display, parameters, alarms, etc. You can change the settings to suit your specific needs. General monitor settings can be changed in the Monitor Setup menu. Other settings are changed in setup menus for each parameter.

Modifications remain in effect until the case is ended, a user mode is changed, or for 15 minutes after the monitor is turned off. Replacing the CPU Board also returns settings to their defaults.

NOTE: You can load or reload the settings using the Datex-Ohmeda Network.

Selecting a user mode

User modes are predefined groups of settings that include general monitor settings and some measurement-specific settings. A user mode defines, for example, what is displayed on the screen and in trends. The preconfigured, factory-default user modes are General, Regional, and Pediatric.

GENERAL Mode for basic general anesthesia and for low- and medium-risk adult

patients.


REGIONAL Mode for all types of regional anesthesia and anesthesia care.

PEDIATRIC Mode for basic general anesthesia and for low- and medium-risk

pediatric patients.

The monitor always starts in startup mode. Startup mode is one of your user modes, chosen during configuration. You can change mode settings during monitoring.

To change the startup mode to another mode:



- Press the ComWheel and select Monitor Setup from the Main Menu.
- 2. Select Select Mode.
- 3. Select GENERAL, REGIONAL, or PEDIATRIC.



The mode you select is marked with an arrow.

Changes made separately are maintained when returning to the previous mode.

Changing and saving configuration settings

Permanent changes to the configuration are made in the Install/Service menu, a submenu of the Monitor Setup menu. You need a password to enter this menu and change the configuration. Almost all modifications are only temporary unless you save them in the Save Modes menu, which is a submenu of the Install/Service menu. A password is also required to enter this menu and save changes.

See *Configuration* at the beginning of chapter 1 for the default passwords needed to enter the Install/Service menu and the Save Modes menu.

If you wish to make permanent changes, we recommend that you contact the person responsible for the entire configuration. When settings are changed and saved, they should be recorded.

Parts of the display




Figure 3-1. Parts of the display

- (1) Alarm silence indicator
- (2) Alarm message field (for five messages)
- (3) Actual time
- (4) Waveform fields with corresponding numeric information on the right
- (5) Digit fields

Waveform fields and digit fields are configurable.



Display setup

At startup, the display is arranged according to the startup mode definitions. Unused parameters are not displayed and no space is reserved for them.





Figure 3-2. Display fields

Changing setup temporarily

You can choose which waveforms and numeric information are displayed, and how they are arranged on the screen.

To change the setup:



- Press the ComWheel and select Monitor Setup from the Main Menu.
- Select Screen Setup.
- 3. Select Waveform Fields, Digit Fields, Split Screen, or Minitrend Length.



NOTE: Choosing to display the same parameter in the waveform field and the digit field makes the previously chosen field disappear so that the same information is not displayed in two fields at the same time.

Modifying waveform fields

You can choose which parameters are displayed on the waveform fields, and in which order they are arranged.





Figure 3-3. Waveform field

Up to six waveforms can be displayed at a time. When fewer than six waveforms are displayed, the remaining waveforms are enlarged to fill the entire waveform area.

Changing the waveform also changes the numerical field on the right of the waveform. It may also change the digit fields. If you choose the same measurement in the waveform field that is currently in the digit field, the digit field will disappear so that the same information will not be displayed in two fields at the same time.

To set the parameters for the waveform fields:



- Press the ComWheel and select Monitor Setup from the Main Menu.
- 5. Select Screen Setup.
- Select Waveform Fields.
- 7. Choose the parameter to be displayed in each field.



The invasive pressure waveforms are displayed only when the transducer is connected to the monitor.

Selecting **Combine Pressures** in the Waveform Fields menu displays invasive pressures in the same waveform field with individual scales.

When you use a 5-lead set in the ECG measurement, up to three different ECG leads can be displayed simultaneously in different fields.





You can change the waveform sweep speed. Alternatives are **Fast** (6.25 mm/second) and **Slow** (0.62 mm/second). Slow waveforms have a sweep speed one tenth of normal. Slow waveforms show changes better than fast waveforms.

To change the sweep speeds:

- 8. Press the ComWheel and select Monitor Setup from the Main Menu.
- 9. Select Sweep Speeds.
- 10. Select the parameter and choose the the sweep speed.

Modifying digit fields

Measured patient data may be displayed in up to four digit fields, which are located on the lower part of the screen. You may change the contents or turn each field OFF. If a digit field is turned OFF, the digit field to its left is usually enlarged to fill the space.





Figure 3-4. Digit fields

The digit fields are numbered from left to right in the Digit Fields menu (Lower Field 1 is the left-most field; Lower Field 4 is the right-most field).

Changing the digit field may also change the waveform field setup. If you choose the same measurement in the digit field that is currently in the waveform field, the measurement waveform field will disappear so that the same information will not be displayed in two fields at the same time.

To modify digit fields:



- 11. Press the ComWheel and select **Monitor Setup** from the Main Menu.
- 12. Select Screen Setup.
- 13. Select Digit Fields
- 14. Choose the parameter to be displayed in each digit field.







Splitting the screen display

You can split a display so that one part always displays spirometry or trend data.





Figure 3-5. Split screen view

To select a split screen view:



- 15. Press the ComWheel and select **Monitor Setup** from the Main Menu.
- 16. Select Screen Setup.
- 17. Select **Split Screen** and choose the option you wish.
 - Spiro 1: Basic spirometry view information.
 - Spiro2: Basic spirometry view with additional values.
 - Trend: Minitrend of the parameters selected to the waveform field.



Minitrend view

You can choose to view the Minitrend data from the last five minutes or the last 30-minute period next to the waveform field for the parameter. 5-minute minitrend is updated every 10 seconds, 30-minute minitrend is updated once every minute.

To modify the trend split screen view:

- 18. Press the ComWheel and select Monitor Setup from the Main Menu.
- 19. Select Screen Setup.
- 20. Select Minitrend Length and choose 5 min or 30 min.


Setting time and date

The time is shown in the upper right corner of the screen. The clock format is 24 hours. Turning the monitor off does not affect the clock.

If the monitor is connected to the Datex-Ohmeda S/5 Network and Central, the Time and Date menu is not available. The monitor follows the time and date settings for the center.

NOTE: To prevent the loss of trend data, you cannot change the time settings after starting a new case.

To set the time and date:




- 21. Press the ComWheel and select **Monitor Setup** from the Main Menu.
- 22. Select Time and Date.
- 23. Select the option you want to adjust.
 - Hour, minutes, and seconds.
 - Day, month, and year.





Display brightness

To adjust display brightness according to environmental conditions:



- 24. Press the ComWheel and select **Monitor Setup** from the Main Menu.
- 25. Select Display Brightness.
- 26. Choose the desired adjustment (30% to 100%).



Displaying pulse rate

Combined Heart Rate and Pulse Rate can be displayed next to the ECG waveform. The current HR source is displayed with bigger font and the QRS symbol flashes next to it.



To display a combined Heart Rate and Pulse Rate:

- 27. Press the ECG key.
- 28. Select ECG Setup.
- 29. Select Display with HR.

Other adjustable features


This section describes other adjustable features associated with setting up the monitor. A password is required for entering the Install/Service menu where the adjustments are made. If you wish to make changes, we recommend that you contact the person responsible for the entire configuration.

Changing parameter colors

You can select the color of each parameter to be yellow, white, green, red or blue.

To change the color:

- 30. Press the ComWheel and select Monitor Setup from the Main Menu.
- 31. Select Install/Service and enter the password.
- 32. Select Colors.
- 33. Select the desired parameter and the color you wish for it in the opened adjustment window.











Changing units of measurement

You can choose the units for height, weight, and for individual parameters.

To change the units:

- 34. Press the ComWheel and select Monitor Setup from the Main Menu.
- 35. Select Install/Service and enter the password.
- 36. Select Installation.
- 37. Select Units and choose the units for Height and Weight.
- 38. To review and/or change the units for a specific parameter, select Parameters.










For your notes:





Chapter 4. Alarms

Overview	4-1
Alarm categories	4-1
ISO pattern alarms	4-2
ISO2 pattern alarms	4-2
General pattern alarms	4-2
Alarm activation	4-3
Displaying alarms	4-3
Alarms Setup menu options	
Adjusting alarm limits	4-5
Changing limits	4-5
Cancelling changes	
Returning to default limits	4-5
Choosing automatic limits	
Calculation formulas for automatic limits	4-6
Changing alarm sources	4-7
Suppressing alarms	4-8
Suppressing audible alarms permanently	4-8
Reactivating alarms	
Suppressing audible alarms for two or five minutes	
Reactivating alarms.	
Suppressing alarms for individual parameters	
Transferring alarms	4-10
Recording alarms automatically	4-10
Other adjustable features	4-11
Displaying limits	
Changing the tone pattern	
Latching alarms	
Automatic snapshots and printouts of them	4-12
ble of Figures	
Figure 4-1. View of alarms	4-1
Figure 4-2. Alarms Setup menu and limits screen	4-3





4. ALARMS

Overview

When an alarm for the monitored parameter becomes active, an audible tone sounds and the screen changes as shown and described below:




Figure 4-1. View of alarms

- (1) Alarm messages appear in order of priority in the message field.
- (2) The measurement value flashes. The background color signifies the alarm category.
- (3) In some cases, a message in the digit or waveform fields gives more detailed information, using the color of the corresponding parameter.

WARNING: Patient safety. Make sure that necessary alarm limits are set and alarms are on when you start monitoring a patient.

WARNING: Patient safety. When alarms are suppressed, observe the patient frequently.

Alarm categories

Alarms are classified into three categories according to their priority: RED ALARM (highest priority), YELLOW ALARM, and NOTE (white color).

The alarm type depends primarily on the physiological significance and the duration (generally a minimum of 20 seconds) of the alarm. Thus, for example, asystole advances rapidly to a red priority, whereas apnea is allowed a slightly longer duration.

The monitor has three choices of alarm tones and patterns: ISO, ISO2 (the default), and General.

NOTE: There is also a sound for catastrophic situations. A continuous beep sounds if the FiO_2 is less than 18% and EtO_2 less than 10%, FiN_2O is high, or Ppeak is high.

To change the alarm pattern, contact the person responsible for the configuration.

ISO pattern alarms

VISUAL	TONE PATTERN	MEANING
Red LED and measurement	Triple + double beep every 5 seconds.	Life-threatening situation.
value flash	5	
	Exception: The second tone pattern is produced 1 second after the first pattern.	
Yellow LED and measurement value flash	Triple beep every 19 seconds 19 19	Serious, but not a life- threatening situation.
White	Single beep -	Advisory note that may be equipment-related, such as "Leads off."

ISO2 pattern alarms

The difference between the ISO pattern and the ISO2 pattern is the rising sound of the tone pattern in the ISO2.

VISUAL	TONE PATTERN	MEANING
Red LED and measurement value flash	Rising triple + double beep every 5 seconds. Exception to this is the second tone pattern which is produced 1 second after the first pattern.	Life-threatening situation.
Yellow LED and measurement value flash	Triple beep every 19.5 seconds 19.5 19.5	Serious, but not a life- threatening situation.
White	Single beep. —	Advisory note that may be equipment-related, such as "Leads off."

General pattern alarms

VISUAL	TONE PATTERN	MEANING
Red LED and measurement value flash	Continuous beeping. —	Life-threatening situation.
Yellow LED and measurement value flash	Double beep every 5 seconds555	Serious, but not a life- threatening situation.
White	Single beep.	Advisory note that may be equipment-related, such as "Leads off."



To make the alarms operative, connect the patient cables. Alarms are operative even when the measurement is not selected on the display (except for the respiration measurement), unless the alarm source is selected Off.

Individual alarms have their own specific requirements before they become active:

- Apnea requires five breaths to be activated.
- Invasive pressures are required to be within alarm limits for 20 seconds after zeroing.

Displaying alarms

Patient alarm limits can be viewed and adjusted in the Alarms Setup menu.

To display the Alarms Setup menu, press the ComWheel and select Alarms Setup from the Main Menu.




Figure 4-2. Alarms Setup menu and limits screen

- (1) List of selections.
- (2) Exit the alarm limit adjustment area and return to the Alarms Setup menu.
- (3) Parameter box with high- and low-limit values with a 10-minute trend displayed to show the current status.

Alarms Setup menu options

Adjust Limits Adjusts individual measurement alarm limits. The adjustment menu

can also be accessed through each parameter menu.

Auto limits are calculated from the displayed patient reading at the

time when auto limits are selected. They are calculated from the last

two minutes.

Default Limits Sets the alarms to the default alarm limits. Default limits can be

changed and saved to modes.

NOTE: Resetting the case also returns alarm limits to their defaults.

Cancel Changes Returns all limits to the ones set before entering the Alarms Setup

menu, if you have not yet exited the menu.

Alarm Volume Adjusts the volume of the audio alarms. The range is from 1 (silent) to 10 (loud). Alarm Volume does not totally silence audible alarms.

Alarms Setup				-1	6		
Adjust Limits 📋	∢ Ex	n j	Alarını Vo	lume 6			
Auto Limits	HR	NIBP Sys	Àrl Sys	CVP Sys	Sp02	CO2 F1	02 E1
Default limits	72	170	172	22	160	3.0	100
Cancel Changes							
Alamı Volunic 🔠							
Antho ON/OF⊢ 🔯							
Alami History				l		herya e construir a	
Remove Menu	43	86	B?	5	92	,0:0	10
Main Menn	12.0	20	35	14	39.8		
		_					
Adjust volume	0.0	9	15	10	37.7		
of audible	Des	Resp	p	MV	Тепи		
alarm time	ĽĬ.	Rate	peak	esp	11_		

Audio ON/OFF

Displays the following options:

- Activate Alarms activates the silenced alarms.
- Silence Apnea silences Apnea and disconnection alarms as well as respiration rate, CO₂, low Ppeak, PEEP, and minute volume limit alarms.
- Silence ECG silences asystole alarm and heart rate limit alarms.
- Silence apnea & asy is the same as Silence Apnea and Silence ECG combined (see above).
- Silence ALL silences all alarms permanently, except FiO_2 <18% (or $EtCO_2$ < 10%), FiN_2O >82%, and Ppeak > 70 cmH₂O alarms.

Alarm History

Displays a list of the last 20 alarms which have reached the yellow or the red alarm level. The alarm list items have the occurrence time and the type beside them.

Remove Menu


Displays only the alarm limits and 10-minute trends.






Changing limits

To change alarm limits for parameters:




- I. Press the ComWheel and select Alarms Setup from the Main Menu.
- 2. Select Adjust Limits.
- 3. Turn the ComWheel to highlight the measurement you wish.
- Press the ComWheel to display an adjustment window.



Turn the ComWheel to change the High alarm limit and/or press the ComWheel to accept the limit.

The cursor moves to the Low setting.



- Turn the ComWheel to change the Low alarm limit and/or press the ComWheel to accept the limit.
- 7. To adjust other measurements, select **Previous Menu.**

OI

To return to the Alarms Setup menu, select **Alarms Setup**.

or

To return to the normal monitoring display, press the **Normal Screen** key.

You can also enter the alarm limit adjustment window through the measurement menus. When you are in a parameter menu, select the alarm for that parameter or enter the setup menu for that parameter, then select the alarm.

NOTE: The ST alarm limits, located in the HR alarm limits menu, can be set after the ST learning sequence. For ST, you receive advisory messages only.

Canceling changes

Before you exit the Alarms Setup menu, select Cancel Changes to return all limits to those that were in effect before you entered the Alarms Setup menu.

Returning to default limits

Select **Default Limits** in the Alarms Setup menu to set the alarms to the factory default alarm limits.

Choosing automatic limits

To get automatic patient-related alarm limits that enable close patient control, select **Auto Limits** in the Alarms Setup menu. These limits are calculated using the displayed patient readings during the two minutes immediately prior to the time when **Auto Limits** is selected.

Calculation formulas for automatic limits

The basis for calculating automatic limits is shown below.

Parameter	Calculation formula
Pulse Oximetry	
SpO ₂ high	(100) = OFF
SpO ₂ low	SpO ₂ – 5%
Temperature	
T1 high	T1 + 1 °C
T1 low	T1 – 1 °C
T2 high	T2 + 1 °C
T2 low	T2 – 1 °C
Hemodynamic Parameters	
HR high	HR * 1.25
HR low	HR * 0.75
NIBP sys, dia, mea high	NIBP sys/dia/mea * 1.25 + 10
NIBP sys, dia, mea low	NIBP sys/dia/mea * 0.75 – 10
P1, P2 high	P*1.25 + 5/10 depending on the label
P1, P2 low	P * 0.75 - 5/10 depending on the label
Gases	
EtCO ₂ high	EtCO ₂ + 1%
EtCO ₂ low	EtCO ₂ – 1%
Resp high	RR * 1.25 + 2
Resp low	RR * 0.75 – 2
Spirometry	
Ppeak high	Ppeak + 10
Ppeak low	Ppeak – 10
PEEP high	PEEP + 5
PEEP low	PEEP - 5
Mvexp high	Mvexp + 2
Mvexp low	Mvexp - 2



Changing alarm sources

For NIBP, invasive pressure, O_2 , anesthetic agents, and temperature, you can select which measured values trigger the alarm.


- For the pressures, the choices are **Systolic**, **Diastolic**, **Mean**, or **OFF**. If **OFF** is selected, the parameter box with the 10-minute trend is not displayed.
- For O₂ and anesthetic agents, the alternative alarm sources are ETO₂ or FiO₂.
- Temperature choices are T1 and T2. Only the last-modified source is active.

To change the alarm source:

- 8. Press the ComWheel and select Alarms Setup from the Main Menu.
- 9. Select Adjust Limits.
- 10. Turn the ComWheel to highlight the measurement you wish, then press the ComWheel.

An adjustment window is displayed.

- 11. Wheel as many times as required to get from the limit setting to the menu selections.
- Select the desired alarm and press the ComWheel to change the selection to OFF or ON.



Suppressing alarms



APN.

When alarms are suppressed, this warning symbol is displayed.


NOTE: If alarms are suppressed and a power interruption occurs, or if the monitor is turned off for up to 15 minutes, check the alarm status before you resume monitoring or begin monitoring a new patient.

Warning: Patient Safety. When alarms are suppressed, observe the patient frequently.

Suppressing audible alarms permanently

For special cases, such as lung operations, it may be desirable to suppress certain audible alarms entirely.

- 13. Press the ComWheel and select Alarms Setup from the Main Menu.
- 14. Select Audio ON/OFF.
- 15. Select Silence Apnea, Silence ECG, Silence Apn&ECG or Silence ALL.



If an active alarm is suppressed, a reminder beep sounds every two minutes. You can adjust the volume of the reminder beep in the Audio ON/OFF menu.

Reactivating alarms

Select Activate Alarms to turn audible alarms on.



- To suppress audible alarms for two minutes, press the Silence Alarms key once.
- To suppress audible alarms for five minutes, press the Silence Alarms key for more than three seconds.

If the alarms are not active when you press the **Silence Alarms** key, they are pre-silenced for two minutes or five minutes. Exception: $FiO_2 < 18\%$, $EtO_2 < 10\%$, $FiN_2O > 82\%$, and high Ppeak alarms are silenced for 20 seconds. New alarms are displayed.

Reactivating suppressed alarms

Press the **Silence Alarms** key during the silent period to activate new alarms immediately, activate silenced alarms (after one minute), and the apnea alarm (after five breaths).

Suppressing alarms for individual parameters

You can turn off the alarms without changing the alarm limits for the following parameters:

- FiCO₂
- EtO₂ and FiO₂
- NIBP (Sys, Dia, and Mean)
- · ST (Lat. and Inf.)
- Temperature (T1 and T2)

To turn off or turn on alarms for a parameter:

- 16. Press the ComWheel and select Alarms Setup from the Main Menu.
- 17. Select Adjust Limits.
- 18. Turn the ComWheel to highlight the measurement and press the ComWheel.
- 19. In the adjustment window that is displayed, press the ComWheel as needed to move to the menu selections.
- 20. Select the parameter alarm to be suppressed from the menu and press the ComWheel to turn the selection **0FF**.



This symbol is displayed in the digit field for each alarm you suppress.

or

Select the parameter alarm to be reactivated from the menu and press the ComWheel to turn the selection **ON**.

Transferring alarms


All alarms can be transferred to the Datex-Ohmeda S/5 Network and Central or to another monitor that is connected to the Datex-Ohmeda monitor network.

- 21. Press the ComWheel and select Patient Data from the Main Menu.
- 22. Select Other Patients.
- 23. Select Receive Alarms.
- 24. Select one of the sites on the list that is displayed.

Recording alarms automatically

You can choose to generate an automatic strip chart recording when a systole, heart rate, and pressure alarms reach the red alarm level. The recorder will print the ECG \pm Art waveforms.

To record on alarms automatically:



- Press the ComWheel and select Record/Print from the Main Menu.
- 26. Select Record Waveforms.
- 27. Select Start on Alarms and choose YES.



Recording will start from the following alarms:

Alarm	Recorded parameters		
Asystole	ECG + P1 waveforms, 25 mm/sec		
HR High/Low	ECG + P1 waveforms, 25 mm/sec		
P1 High/Low	ECG + P1 waveforms, 25 mm/sec		

Recording time is 30 seconds: 12 seconds recording from the recorder memory, 18 seconds real-time recording. Alarm recordings are annotated with the alarm source.







This section describes the other adjustable features related to alarms. A password is required for entering the Install/Service menu where the adjustments are made. If you wish to make changes, we recommend that you contact the person responsible for the entire configuration.

Displaying limits

Alarm limits can be displayed next to the numerical parameter value as shown below.



To display alarm limits:



- 28. Press the ComWheel and select **Monitor Setup** from the Main Menu.
- 29. Select Install/Service and enter the password.
- 30. Select Installation.
- 31. Select Alarms Options.
- 32. Select Show Limits and choose YES or NO.



Make sure the alarm source is selected. The alarm limits become visible when patient monitoring begins and the monitor receives patient data.

Changing the tone pattern

The monitor has three choices of alarming tone patterns: ISO, ISO2, and General.

To change the tone pattern:


- 33. Press the ComWheel and select Monitor Setup from the Main Menu.
- 34. Select Install/Service and enter the password.
- 35. Select Installation.
- 36. Select Alarms Options.
- 37. Select Alarm Tones and choose ISO, ISO2, or General.

Latching alarms

Latching alarms enables unattended monitoring. If **Latching Alarms** is active, alarm messages continue to be displayed even if the initial alarm condition has ceased. A reminder beep sounds every 10 seconds.

- To clear all inactive alarm messages and the beep, press the Silence Alarms key once.
- To clear all inactive and active alarm messages, enabling only new upcoming alarm messages, press the **Silence Alarms** key twice.

To select latching alarms:



- 38. Press the ComWheel and select Monitor Setup from the Main Menu.
- 39. Select install/Service and enter the password.
- 40. Select installation.
- 41. Select Alarms Options.
- 42. Select Latching Alarms and choose YES.




Alarm behavior

If the alarm mode is latched, the technical alarms are latched as well. This does not comply with the NIBP (IEC 60601-2-30) and invasive pressure (IEC 60601-2-34) standard requirements.

Automatic snapshots and printouts of them

When an Asystole, HR high or low, or P1 high or low alarm occur, the monitor can automatically create and print a snapshot that you can view in the Trends menu. A snapshot is created when the Asystole, heart rate, or P1 red alarm is activated. The snapshot contains data from the last 15 seconds.

To create a snapshot on alarms and print it automatically:



- 43. Press the ComWheel and select Monitor Setup from the Main Menu.
- 44. Select Install/Service and enter the password.
- 45. Select Trends&Snapshot.
- 46. Select Snapshot.
- 47. Select Create on Alarms and choose YES.
- 48. Select Automatic Print and choose Alarms.





For your notes:





Chapter 5. Trends and Snapshots

Overview	5-1
Trended parameters	5-1
Factory default parameters	5-1
Most common tasks	5-2
Recording trend data	5-3
Changing graphical trend parameters	
Printing trend data	
Printing selected pages of graphical trend data	
Erasing trend data	
Graphical trend view	
Graphical trend view symbols	
Changing the time scale (trend length and resolution)	
Moving on graphical trend pages	5-6
Minitrend view	5-6
Adding a minitrend	5-7
Choosing the length of minitrend data	5-7
Removing a minitrend	5-7
Numerical trend view	5-8
Scrolling numerical trend pages	5-8
Snapshots	5-9
Viewing snapshots	5-10
Creating snapshots	5-11
Marking an event	
Erasing snapshots	5-11
Other adjustable features	5-12
Creating snapshots on alarms	5-12
Printing snapshots automatically	5-12
Printing loops on snapshots	5-13
Configuring snapshot and graphical trend pages	5-13

Table of Figures

Figure 5-1. Graphical trend view	5-4
Figure 5-2. Minitrend view	
Figure 5-3. Numerical trend page	
Figure 5-4. Waveform snapshot page	
Figure 5-5. Waveform snapshot	
Figure 5-6. Snapshot fields	
Figure 5-7. Marking numerical trends	









5. TRENDS AND SNAPSHOTS

Overview

A trend is a collection of the measured patient data. The monitor displays two types of trend information: graphical and numerical. You can create snapshots (frozen frames saved to memory) of this information.

The monitor collects graphical and numerical trend data automatically from the trended parameters. You can select between 20 minutes and 24 hours of trend time.

You can view the trends in the Trends menu, or you can continuously display graphical "minitrends" next to the waveform fields. A minitrend is a view of collected trend data from the last 5-minute period or 30-minute period, depending on which period of time you chose.

Trended parameters

The trend parameters are listed below

- Electrocardiograph (HR)
- ST analysis
- · Invasive pressure
- Noninvasive blood pressure
- Oxygen saturation (Pleth, SpO₂)
- Gases (CO₂, O₂, N₂O, AA, MAC, Resp)
- Respiration rate (RR)
- Temperature
- Airway pressure
- Airway flow
- Derived values from airway pressure and flow
- NMT

Trend data is stored in memory for 15 minutes after the power is turned to Standby.

Factory default parameters

For factory default parameters, contact the person responsible for the entire configuration.

The configuration and the mode determine which parameters are displayed.

The unit of the parameter matches the real-time waveform settings for that parameter.

To change the snapshot or graphical trend settings, see *Configuring snapshot and graphical trend pages* at the end of this chapter.

NOTE: You cannot change the contents of numerical trend fields. Numerical trend fields cannot be configured.

Most common tasks

To display trends	Press the Trends key to open the Trends menu.
	The trend (graphical or numerical) that was displayed last appears beside the Trends menu.
To scroll time with the	1. Select Cursor.
trend cursor	2. Turn the ComWheel to move the cursor.
	For the graphical trend, the numeric measurement values for the displayed time are next to the cursor.
	3. Press the ComWheel to return to the Trends menu.
To see more parameters on other pages	Select Next Page .
To change time scale	Select Time Scale and the trend time (20 min or 1 , 2 , 4 , 6 , 8 , 10 , 12 , 24 hours).
To change trend type or a snapshot to be viewed	Select Graphical , Numerical or Snapshot .
To change scaling	Select Trend Scales.
To select minitrend split screen	1. Select Monitor Setup in the Main Menu.
	2. Select Screen Setup.
	3. Select Split Screen and Trend .
To create snapshots	Select Take Snapshot in the Trends menu.
To record trends	Press the Record Trend/Stop key.
	or
	1. Select Record/Print in the Main Menu.
	2. Select Record Trends.
	3. Select Record Numerical or Record Graphical.
To print trends	To print the currently viewed trend data, press the Trends key and select Print Page .
	To print graphical trend data, select Print Graphical in the Record/Print menu.
	To print numerical trend data, select Numerical and then Print Page .
To erase trend history	1. Select Reset Case in the Main Menu.
	Select Reset Trends and Yes in the adjustment window.

Recording trend data

Press the Record Trend/Stop key.

or

- Press the ComWheel and select Record/Print from the Main Menu.
- 2. Select Record Trends.
- 3. Select Record Numerical or Record Graphical.

Graphical trends are recorded from the time period that corresponds to the time scale (20min - 24h) chosen for them in the Trends menu.

Changing graphical trend parameters

To change the recorded parameters for graphical trends:

- 4. Press the ComWheel and select Record/Print from the Main Menu.
- 5. Select Record Trends.
- 6. Select Graphic Trend 1 and choose the parameter to be recorded.
- Select Graphic Trend 2 and choose the parameter to be recorded.

Printing trend data

- 8. Press the Trends key.
- 9. Select Graphical or Numerical.
- 10. Select Print Page.

If you wish to print trend data located on other pages, select **Next Page** and **Print Page**.

Printing selected pages of graphical trend data

- 11. Press the ComWheel and select Record/Print from the Main Menu.
- 12. Select Print Graphical.
- 13. Select the page (Page 1, Page 2, etc.).

Erasing trend data

- 14. Press the ComWheel and select Reset Case from the Main Menu.
- 15. Select Reset Trends.
- 16. In the opened window, select YES to erase the trends.

or

Select NO to cancel the function.

Graphical trend view

Graphical trend pages contain up to four pages with five fields on each page. All five fields are usually visible and can be printed.

Parameter scale, label, unit, and color follow the real-time waveform settings for each parameter. Exceptions: you can select the HR and temperature scales in the Trend Scales menu.




Figure 5-1. Graphical trend view

- (1) Trends menu
- (2) Trend measurement field
- (3) Real-time ECG
- (4) Numeric value of a measurement at the trend cursor point
- (5) Trend page number
- (6) Indication of the amount of data gathered and viewed
- (7) Time and marker field



Graphical trend view symbols

Trend bar, parameter scale to the left. The gap shows the mean value (except in the Paw field where it indicates Pplat).

Trend bar, parameter scale to the right. The gap shows the mean value (except in the Paw field where it indicates Pplat).

NIBP trend bar.

A dotted vertical line across the trend field indicates a change, such as a change of ECG lead, anesthetic agent, or zeroing the invasive blood pressure channel.

A blue, white, or red line above the marker field indicates the following:

- **Blue line**: the left end of the line is the point at which gathering of trend data started. The right end shows the last moment data was gathered.
- White line: the proportion of data shown on the screen. If the line is on the left, leaving space on the right, there is more data to see after the current view. If the line is on the right, there is more data before the current view. If the line is in the middle, there is more data at the beginning and end of the case.
- Red line: the time period during which a 20-minute trend was gathered.

Changing the time scale (trend length and resolution)

Adjust the time scale (and/or parameter scale) as needed to view the desired level of trend detail.

To change trend length and resolution:

- 17. Press the Trends key and select Time Scale in the Trends menu.
- 18. Select the trend length.

Trend length on the display	Resolution	Trended Time Period
20 min	10 seconds	last 30 minutes
1 h	1 minute	last 24 hours
2 h	1 minute	last 24 hours
4 h	2 minutes	last 24 hours
6 h	3 minutes	last 24 hours
8 h	4 minutes	last 24 hours
10 h	5 minutes	last 24 hours
12 h	6 minutes	last 24 hours
24 h	12 minutes	last 24 hours

Moving on graphical trend pages

Numerical measurement values for trended parameters appear next to the cursor on the graphical trend. The displayed values were measured at the time indicated by the cursor.

To view trend values for specific times:

- 19. Press the Trends key and select Cursor.
- 20. Turn the ComWheel to move the cursor and scroll the values for specific times.
- 21. To return to the Trends menu, press the ComWheel.
- 22. To see more parameters on other pages, select Next Page in the Trends menu.

Minitrend view

17:00




Figure 5-2. Minitrend view

You can split the **Normal Screen** display so that one fourth of the display (on the left) shows continuous graphical minitrends. Minitrends show the measurements selected for the waveform fields.


NOTE: The split screen option is possible only with displays that have waveforms on the **Normal Screen**.





Adding a minitrend

To split the screen and show minitrends:



- 23. Press the ComWheel and select **Monitor Setup** from the Main Menu.
- 24. Select Screen Setup.
- 25. Select Split Screen and choose Trend.



Choosing the length of minitrend data

You can view a minitrend consisting of the last 5 minutes or the last 30 minutes of trend data. A 5-minute minitrend is updated every 10 seconds; a 30-minute minitrend is updated once every minute.

To choose the length of the minitrend data:

- 26. Press the ComWheel and select Monitor Setup from the Main Menu.
- 27. Select Screen Setup.
- 28. Select Minitrend Length and choose 5 min or 30 min.

Removing a minitrend

To remove the minitrend from the display:

- 29. Press the ComWheel and select Monitor Setup from the Main Menu.
- 30. Select Screen Setup.
- 31. Select Split Screen and choose None.

Numerical trend view

Numerical trends contain:

- One to four pages of trend information (24 hours maximum).
- · Real-time ECG at the top of each page.

The resolution is five minutes.




Figure 5-3. Numerical trend page

- (1) Trend menu
- (2) Page number
- (3) Real-time ECG
- (4) Real-time digit fields, if the waveform display is Normal Screen

Scrolling numerical trend pages

Use the ComWheel to scroll the trend vertically.

When the highlight reaches the top or the bottom of the view, the next five minutes of information appears.



Snapshots

A snapshot is a frozen frame that is saved to the monitor memory. A snapshot captures the contents of the frozen frame chosen during configuration, not the image you are currently viewing when you take the snapshot. A snapshot can contain waveforms, numerical trends, and graphical trends.




Figure 5-4. Waveform snapshot page

- (1) Trends menu
- (2) Real-time ECG
- (3) Waveform snapshot
- (4) Time of the snapshot (If created on an alarm, the cause of the alarm appears.)
- (5) Real-time digit fields, if the waveform display is Normal Screen
- (6) Page number
- (7) Waveform cursor
- (8) Graphical trend snapshot
- (9) Time and marker field

To change the content of the snapshot page, see *Configuring snapshot and graphical trend pages* at the end of this chapter.

Viewing snapshots

A snapshot includes 15 seconds of a waveform and/or graphical trend from the time period that corresponds to the time scale selected for the graphical trend and/or the numerical trend.




Figure 5-5. Waveform snapshot

- (1) Numerical label
- (2) Time when saved

The rest of the settings depend on the current settings for each real-time waveform.

There is one configurable waveform snapshot page. The page contains up to six fields that can include snapshots of waveforms and graphical and numerical trends.

You can see five fields on the display but six on the printout.




Figure 5-6. Snapshot fields

To view snapshots:

- 32. Press the Trends key and select Snapshot from the Trends menu.
- 33. Select Next Snapshot.
- 34. Turn the ComWheel to move the cursor in the time and marker field.
- 35. When the cursor reaches the saving time of a snapshot (marked with an "S" in the time axis), the snapshots that were saved at that time appear.
- 36. Press the ComWheel to return to the menu.



Creating snapshots

To create a snapshot, press the Trends key and select Take Snapshot.

The waveforms and trends that are currently selected to the Snapshot page will be saved. You can take up to 16 snapshots, depending on the data load and the amount of fields chosen to contain information on the snapshot page.

Marking an event

At the time **Take Snapshot** is selected, a number is inserted in the numerical trends beside the trend information. This number marks the event at which a snapshot was created.

Also, NIBP measurement marks a trend line. If there are several measurements during one minute, a one-minute average appears instead of individual measurements.

Page	4 Of 4	T1	T2	Sp02
Mark	Time	T1	T2	<u> </u>
	12:25			
	12:30			
	12:35			
1	12:39	38.8	20.0	
	12:40	38.8	20.0	
2	12:43	38.8	20.0	
	12:45	38.8	20.0	1

Figure 5-7. Marking numerical trends

The event number also appears on numerical trend printouts.

Erasing snapshots

The snapshots will be erased when you reset the case or reset the trends. They are also erased after 24 hours automatically.

To erase snapshots:

- 37. Press the ComWheel and select **Reset Case** from the Main Menu.
- 38. Select Reset Trends or Reset ALL, and choose YES.


As memory fills up, older snapshots are overwritten automatically.

Other adjustable features

This section describes the other adjustable features related to trends and snapshots. A password is required for entering the Install/Service menu where the adjustments are made. If you wish to change the adjustments, contact the person responsible for the entire configuration.

Creating snapshots on alarms

The monitor can automatically create snapshots when the following alarms occur: Asystole, HR high, HR low, P1 high, P1 low.




- 39. Press the ComWheel and select Monitor Setup from the Main Menu.
- 40. Select Install/Service and enter the password.
- 41. Select Trends & Snapshot.
- 42. Select Snapshot.
- Select Create on Alarms and choose YES.



Printing snapshots automatically

Snapshots can be printed on request, on alarms, or when they are created.



- 44. Press the ComWheel and select **Monitor Setup** from the Main Menu.
- 45. Select Install/Service and enter the password.
- 46. Select Trends & Snapshot.
- 47. Select Snapshot.
- 48. Select Automatic Print.

Choose N0, if you want printing to occur only on request.

or

Choose **ALARMS**, if you want to print only snapshots that are created on alarms.

or

Choose **ALL**, if you want to print all snapshots immediately after they are created.



Printing loops on snapshots

You can have Patient Spirometry loops printed whenever snapshots are taken.

- 49. Press the ComWheel and select Monitor Setup from the Main Menu.
- 50. Select Install/Service and enter the password.
- 51. Select Trends & Snapshot.
- 52. Select Snapshot.
- 53. Select Print Loops.
- 54. Select YES.

Configuring snapshot and graphical trend pages

You can change the parameters on the snapshot page and on graphical trend pages.

- For the snapshot page, you can also select graphical trend boxes.
- For graphical trend pages, you can also select events. An event list will appear if there is no menu to hide it.

NOTE: Numerical trend pages cannot be configured.

To configure the snapshot page and/or graphical trend pages:

- 55. Press the ComWheel and select Monitor Setup from the Main Menu.
- 56. Select Install/Service and enter the password.
- 57. Select Trends & Snapshot.
- 58. Select Snapshot.

or

Select Graphical Trends and select the trend page you want to configure.

59. Select the parameter for each field.

The field numbers start on the top of the display. Select one parameter for each field on the trend page, or turn the field **OFF**.

When all fields are **0FF**, the page is displayed with empty fields. The time scale and page number appear at the bottom of the page.

If several equal fields are selected on top of each other, they will form one higher field. Equal fields can not be defined separate from each other.

For your notes:



Chapter 6. Patient Data Management

Overview.		6-1
Network o	ptions	6-1
Using	the network	6-2
	Connecting to the network	6-2
Using	memory cards	
	Inserting a memory card	6-3
Resetting	a case	6-4
Adding de	mographics	6-4
Loading a	previous case	6-5
	Load Prev. Case	6-5
	Patient from Net. (with N-XNET or N-XDNET)	
	Patient from Card (with N-XDNET only)	6-5
Marking e	events	6-6
Viewing o	ther sites (with N-XNET or N-XDNET option)	6-7
View	ing other site alarms (with N-XNET or N-XDNET option)	6-8
Table of Fi	gures	
	Inserting and removing a memory card	
Figure 6-2.	Marked events in trends	6-6
Figure 6-3	Data from another site	6-7

		**efage.gr	



6. PATIENT DATA MANAGEMENT

Overview

The Cardiocap/5 monitor continuously collects and saves patient data, such as trends. Saving is activated when the monitor receives vital signs. With a single, stand-alone monitor, you can add patient demographics and place markers on the trends.

The data management capabilities can be expanded if your monitor is connected to the Datex-Ohmeda S/5 Network and iCentral or if you use memory cards to store and transfer data.


Network options

The monitor can be equipped with the following options to enhance network and data continuity capabilities:

N-XNET Network

N-XDNET Data card and Network

- The N-XNET option allows real-time information sharing between your Cardiocap/5 and monitors that are connected to the same monitor network (the Datex-Ohmeda S/5 Network and iCentral). When the monitor is connected to the network, patient information is saved to the network computer. The network offers the possibility of viewing real-time patient information, such as alarms, from another monitoring site.
- The N-XDNET option offers the same capabilities as the N-XNET option and, in
 addition, it allows you to load trend data from a Data card. With this option, you
 can assure the continuity of patient data by gathering all previously collected
 information for a patient, then adding new information to the same patient file.



Using the network

With the N-XNET or N-XDNET option, you can connect the monitor to the Datex-Ohmeda S/5 Network and iCentral for central monitoring and printing services.



This symbol is displayed when the monitor is connected to the network.

- Time and date are set by the network and cannot be adjusted in the monitor.
- The patient's ID entered in the Patient Data menu can be displayed on the Datex-Ohmeda S/5 Network and iCentral screen and printed on the laser printouts.
- Measured and entered patient data can be viewed on the Datex-Ohmeda S/5 Network and iCentral.
- Measured patient data and alarms can be viewed on other Datex-Ohmeda S/5, CS/3, and AS/3 monitoring systems, and vice versa.

Connecting to the network

To connect your monitor to the network:

- Connect the ethernet and location ID plug to the corresponding connectors on the rear panel of the monitor.
- Configure the location IDs in the Datex-Ohmeda S/5 Network and iCentral.

Using memory cards

With the N-XDNET option, you can load collected patient data from a monitor that is connected to the same monitor network or from a Data card.

The Data card and the Menu card are memory cards. Each is housed internally in a built-in memory module.

- The Data card is for storing and transferring trend data.
- The Menu card is for storing menus.

When the Data card is inserted, continuous patient trend data is stored on it automatically. The capacity of the card is approximately 48 hours, depending on the complexity of the saved data. When the card is full, the oldest data is erased.

The trend data stored on a Data card can be transferred to and viewed at all S/5, AS/3, and CS/3 systems as well as other Cardiocap/5 monitors.

See Loading a previous case later in this chapter for information about loading trend data.

CAUTION: Do not subject memory cards to excessive heat, bending, or magnetic fields.

Inserting a memory card

You can insert a memory card into either slot. The monitor recognizes whether the inserted card is a Data card or a Menu card.




Figure 6-1. Inserting and removing a memory card

Insert the end of the card with the small connectors first. Push the card firmly into place. The release button will come up. To remove the card, press the release button, then pull the card out of its slot.




This symbol appears on the display to indicate that a memory card is inserted. The "Data card inserted" or "Menu card inserted" message appears in the message field of the display.

Resetting a case

At the beginning of a new procedure, you should erase the screen layout, trend data, and all alarm and parameter settings of the previous case.

To reset a case:




- 1. Press the ComWheel and select **Reset Case** from the Main Menu.
- Select Reset ALL and choose YES.



Selecting **Reset ALL** returns you to the starting mode and its settings (screen layout, trend parameters, alarm limits, etc.).

Adding demographics

To add demographics:



- 1. Press the ComWheel and select **Patient Data** from the Main Menu.
- 2. Select **Demographics**.
- 3. Enter the patient's height and weight.



Height Adjustable from 15 cm to 250 cm.

Weight Adjustable from 1 kg to 250 kg.


BSA

The body surface area is calculated automatically using the du Bois formula. The body surface area is used when calculating index values of certain parameters.



Loading a previous case

To load data from a previous case:



- 4. Press the ComWheel and select **Patient Data** from the Main Menu.
 - Select Load Prev. Case.



5.

Load Prev. Case

When the network is not in use, you can retrieve the trend data of the previous case from the monitor memory. You can retrieve trend data that was stored during the previous 24 hours under these circumstances:

- The monitor was turned off for less than 15 minutes.
- The monitor remained on (even if the case was reset).

Reloading trend data after the case has been closed for several hours is not recommended. The period that contains no information results in empty record pages.

Patient from Net. (with N-XNET or N-XDNET)

You can retrieve the last 24 hours of trend data from the network if the monitor is connected to it.

Patient from Card (with N-XDNET only)

When the data memory card is inserted, you can load trend data from a case on that card. The Data card saves previous patient data from the time it is inserted in the monitor. You can retrieve approximately 24 hours of information. The memory capacity depends on the complexity of the saved data.



Marking events

During monitoring, you can place numbers beside the numerical trend to mark the event.

To mark events:

- 6. Press the Trends key.
- 7. Select Take Snapshot.

The events are numbered from 1 to 16, depending on the data load. A number is displayed beside the numerical trend information that was recorded at the time **Take Snapshot** was selected.

Page	4 Of 4	T1	T2	Sp02
Mar	kTime	T1	T2	
	12:25			
	12:30			
	12:35			
1	12:39	38.8	20.0	
	12:40	38.8	20.0	
2	12:43	38.8	20.0	
	12:45	38.8	20.0	1

Figure 6-2. Marked events in trends

A trend printout is also marked with this number.

A snapshot is created at the same time. For more information about Snapshots, refer to the *Trends and Snapshots* chapter.








Viewing other sites (with N-XNET or N-XDNET option)

With your Cardiocap/5 monitor you can see the numerics, waveforms, and alarms of another Cardiocap/5, S/5 Light, S/5, AS/3 or CS/3 monitor if both monitors are connected to the same monitor network.

To view data for a patient at another site:



- 1. Press the ComWheel and select **Patient Data** from the Main Menu.
- 2. Select Other Patients.
- 3. Select Show Vital Signs.
- 4. Select the site you wish to view.







Figure 6-3. Data from another site

Viewing other site alarms (with N-XNET or N-XDNET option)

With a networked Cardiocap/5 monitor, you can view the alarms of another monitor on the same network, including the Cardiocap/5, S/5 Light, S/5, AS/3, or CS/3, or the iCentral station.

To view the alarms of a monitor at another site:



- 8. Press the ComWheel and select **Patient Data** from the Main Menu.
- 9. Select Other Patients.
- 10. Select Receive Alarms.
- 11. Select the site you wish to view.



The sites from where the alarms are forwarded to your monitor are marked by a bell next to the site name.





7-1
7-1
7-2
7-2
7-2
7-3
7-4
7-4 7-5
7-6
7-6
7-6 7-7
7-7
7-8
7-9
7-9
7 - 9
7-10
.7-10
7-10
7-1
7-2 7-6



Overview

You can record waveforms, values, and loops and print them to a laser printer that is either connected to the monitor or accessible over the Datex-Ohmeda monitor network.

Recording and printing options are located in the Record/Print menu and its submenus.

- You can record or print just one loop or one view of trends from within the corresponding measurement menu.
- You can record or print several pages, several parameters, or all gathered information from the Record/Print menu.

Recorder (N-XREC option)

You can use the optional built-in recorder (N-XREC option) to do the following:

- Record the graphical or numerical trend and print up to 24 hours of trend data.
- Record up to three real-time waveforms simultaneously.
- Display recordings of numerical information (in horizontal and vertical plane).

NOTE: Thermal paper printouts may be destroyed when exposed to light, heat, alcohol, etc. Make a photocopy for your archives.




Figure 7-1. Built-in recorder

NOTE: The two-button recorder (shown) is for Cardiocap/5 monitors using software version 3.0 or higher. A one-button recorder was available previously.

- Record Waveform/Stop key. Press to start or stop recording selected real-time waveforms. This key functions like the Record Start/Stop key on a one-button recorder.
- (2) **Record Trend/Stop** key. Press to start or stop recording selected numerical or graphical trend data. Since this key is not available on a one-button recorder, use the Record/Print menu options to print trend data.
- (3) Key to release and open the recorder paper compartment.
- (4) Recorder paper.
- (5) Recorder paper compartment.

Recording waveforms




Figure 7-2. Sample recording

NOTE: Waveform scaling follows the displayed parameter scaling, when applicable.

Start/Stop recording waveforms

To record waveforms:

- 1. Make sure paper is installed in the recorder.
- 2. Press the Record Waveform/Stop key.

OI.

Press the ComWheel and select Record/Print from the Main Menu.

Select Record Waveforms.

Select Record Wave.


The waveforms are recorded according to the current waveform selections.

3. To stop recording, press the Record Waveform/Stop key again. or

Select Stop Wave in the Record Waveforms menu.

Controlling paper speed

To see the waveforms more clearly or more generally, you can adjust the paper speed:



- 1. Press the ComWheel and select **Record/Print** from the Main Menu.
- 2. Select Record Waveforms.
- 3. Select **Paper Speed** and choose the speed you prefer (1, 6.25, 12.5, or 25 mm/second).









To select the waveforms you want to record:

4. Press the ComWheel and select Record/Print from the Main Menu.




- Select Record Waveforms.
- 6. Select Waveform 1, Waveform 2, or Waveform 3 and choose the parameter to be recorded. You can record up to three waveforms simultaneously. If you want to record only one waveform, select OFF for the other waveform fields.
- Choose the parameter for the waveform.
 Scroll toward -More- to see additional options.



Controlling the recording time

You can select 30 seconds (the default) or continuous as the recording time.

To select the length of the recording:



- 8. Press the ComWheel and select **Record/Print** from the Main Menu.
- 9. Select Record Waveforms.
- 10. Select Length and choose 30 s or Cont.



NOTE: When recording is activated by alarms, the recording time is always 30 seconds.

Selecting the recording time initiation

The recorder can record only real-time data (delay of 0 seconds) or the first 12 seconds from memory (delay of 12 seconds, the recording delay default).

You can set the recording delay to 0 seconds and record only real-time data.

To change the delay:

- 11. Press the ComWheel and select Record/Print from the Main Menu.
- 12. Select Record Waveforms.
- 13. Select Delay and choose 0 sec.


NOTE: When recording is activated by alarms, the delay is always 12 seconds.



Recording waveforms on alarms

You can generate a strip chart recording automatically when asystole, heart rate, and pressure alarms reach the red-alarm level, by setting up the recorder to start recording on alarms.

To start recording on an alarm:



- 14. Press the ComWheel and select **Record/Print** from the Main Menu.
- 15. Select Record Waveforms.
- 16. Select Start on Alarms and choose YES.



The alarms listed below will start the recording:

Alarm	Recorded parameters
Asystole	ECG + P1 waveforms, 25 mm/second
HR High/Low	ECG + P1 waveforms, 25 mm/second
P1 High/Low	ECG + P1 waveforms, 25 mm/second

Alarm recordings are annotated with the alarm source.

Recording trends

Adjust the parameter and time scales to record the desired level of trend detail.

Start/Stop recording trends

To record numerical or graphical trends:

- 17. Make sure paper is installed in the recorder.
- 18. Press the Record Trend/Stop key.

or

Press the ComWheel and select Record/Print from the Main Menu.

Select Record Trends.

Select Record Numerical or Record Graphical.

A numerical or graphical trend is recorded according to the current selections.

19. To stop recording, press the Record Trend/Stop key again.

Or

Select Stop Numerical or Stop Graphical in the Record Trends menu.

Numerical trend printout

You do not need to choose parameters for a numerical trend recording.




Figure 7-3. Numerical trend printout

Numerical trend recordings contain the following parameters:

Parameter	Printed values and units
HR and SpO ₂	bpm/SpO_2
NIBP	Sys/dia or mean mmHg
P1 "Art"	Sys/dia or mean mmHg
P2 "CVP"	Sys/dia or mean mmHg
T1	Celsius or Fahrenheit
T2	Celsius or Fahrenheit
CO_2	Et/Fi %, kPa or mmHg
O_2	Et/Fi %
AA	Et/Fi %
Resp. Rate	Breaths per minute

Either Sys/Dia or mean pressures are recorded, depending on the digit format selected in the pressure setups.





To select the resolution for a numerical trend record:


- 20. Press the ComWheel and select Record/Print from the Main Menu.
- 21. Select Record Trends.
- 22. Select Trend Resolution and choose from the options.

Selecting graphical trend parameters

You can record graphical trends for two parameters.

To choose the parameters for a graphical trend recording:

- 23. Press the ComWheel and select Record/Print from the Main Menu.
- 24. Select Record Trends.
- 25. Select Graphic Trend 1 or Graphic Trend 2.
- Choose the parameter for the graphical trend.
 Scroll toward -More- to see additional options.



Recorder paper replacement




Figure 7-4. Recorder paper replacement

To change the paper in the recorder:

- 27. Press the key that releases the paper compartment door.
- 28. Remove the roll and insert a new roll.
- 29. Close the paper compartment door.



Printing to a laser printer


You can print the following to a laser printer:

- Graphical and numerical trends.
- Patient Spirometry values and loops.

You can print several pages, all saved loops, or trend data from the Record/Print menu. You can print one page or one loop from the corresponding parameter menu.

Changing the printer

To check or change the printer connection:



- Press the ComWheel and select Record/Print from the Main Menu.
- 31. Select Printer Connection.
- 32. Select Serial, Net, or None.
 - Choose Serial if the printer is connected to your monitor.
 - Choose Net if your monitor is connected to the Datex-Ohmeda S/5 Network and Central.

Printing trend data

Adjust the parameter and time scales to print the desired level of trend detail.

To print a specific trend view:

- 33. Press the ComWheel and select Trends from the Main Menu.
- 34. Select the trend type you wish to print: Snapshot, Numerical, or Graphical.
- 35. Select Next Page to choose a specific trend page.
- 36. Select Print Page.

To print all numerical trend data:

- 37. Press the ComWheel and select Trends from the Main Menu.
- 38. Select Numerical.
- 39. Select Print Page.

To print several pages of graphical trend data:

- 40. Press the ComWheel and select Record/Print from the Main Menu.
- 41. Select Print Graphical.

Printing spirometry loops

To print a pair of loops:

- 42. Press the ComWheel and select Parameters from the Main Menu.
- 43. Select Airway Gas.
- 44. Select Spirometry Loops.
- 45. Select Print Loop.

To print all saved loops:

- 46. Press the ComWheel and select Record/Print from the Main Menu.
- 47. Select Print Loops.

Other adjustable features

This section describes the other adjustable features related to recording and printing. A password is required for entering the Install/Service menu where the adjustments are made. If you wish to make changes, we recommend that you contact the person responsible for the entire configuration.

Changing the format for recorded numerical trends

You can select the kind of numerical trend that will be recorded when you select **Record Numerical** from the Record/Print menu

- 48. Press the ComWheel and select Monitor Setup from the Main Menu.
- 49. Select Install/Service and enter the password.
- 50. Select Installation.
- 51. Select Printer & Recorder.
- 52. Select Num Trend Type.
- 53. Select Num. or Tab.








Checklist	8-1
General	8-1
ECG	8-1
Impedance respiration	8-1
Pulse oximetry	8-2
Non-invasive blood pressure (NIBP)	8-2
Invasive blood pressure	8-2
Temperature	8-3
Airway gases	8-3
Patient spirometry	8-3
NeuroMuscular Transmission (NMT)	8-3
Messages	8-4
Other situations	8-8







8. TROUBLESHOOTING

This chapter consists of three parts that should help you solve common monitoring problems: *Checklist*, *Messages*, and *Other situations*.

Checklist

Check the following items to ensure you remember to make all essential preparations before monitoring. They also help if problems occur during monitoring.

General

Check that

- The monitor has no visual defects, such as cracks or loose parts.
- The power cord is connected to an electrical wall outlet and to the monitor.
- The D-fend water trap is empty and properly pushed into its place (with F-MXG).
- The sampling line is connected to the monitor (with F-MXG).
- Patient connection cables are attached to the connectors so that the color coding on the plug matches the color coding on the connector.
- Trends of the previous patient are erased.
- · Alarm limits are suitable for the patient.

ECG

Check that

- Electrodes are positioned correctly, electrode gel is moist, and skin contact is good.
- Correct lead-wire set is selected (3-lead or 5-lead) and correct lead-wire type is selected in the ECG Setup menu.
- Lead-wire set is properly connected to the ECG extension cable.
- ECG extension cable is plugged into the blue connector.
- Pacemaker selection in the ECG Setup menu is Show when a pacemaker is used.
- ECG is selected for display (Main Menu Monitor Setup Screen Setup).

Impedance respiration

Check that

- Electrodes are positioned correctly, electrode gel is moist, and skin contact is good.
- Correct lead-wire set is selected (either 3 or 5 lead) and correct lead-wire type is selected in the ECG Setup menu.
- Lead-wire set is properly connected to the extension cable.
- ECG extension cable is plugged into the blue connector.
- Resp is selected for display (Main Menu Monitor Setup Screen Setup).

Pulse oximetry

Check that

- Correct sensor is selected for the size of the patient and sensor is positioned correctly on the patient.
- The sensor is completely dry after cleaning.
- Sensor cable is plugged into the gray connector and is properly connected to an
 extension cable or sensor, as applicable.
- SpO₂ is selected for display (Main Menu Monitor Setup Screen Setup).

Non-invasive blood pressure (NIBP)

Check that

- Correct NIBP hose is used (black hose for adults and children; white hose for infants).
- NIBP cuff is not loose; correct cuff for patient size is used.
- O-rings on hose connectors and cuff connector are intact; cuff connector is firmly pushed inside the cuff tube.
- NIBP hose is properly connected to the monitor connector and will not detach by pulling.
- There are no holes or cracks in the cuff bladder or cuff tube.
- Symbol indicating the center of the bladder is over the artery.
- All residual air is squeezed out of the cuff before wrapping it around the arm.
- Cuff is at heart level.
- · Cuff tubes and NIBP tube are not kinked or squeezed together.
- Non-invasive blood pressure is selected for display (Main Menu – Monitor Setup – Screen Setup).

If a leak is suspected, start venous stasis and check that the pressure is stable during stasis.

Invasive blood pressure

Check that

- Invasive blood pressure transducer cable is plugged into a red connector.
- Pressure transducer is connected to the cable.
- Patient catheter is connected to the pressure line.
- · There is no air in the transducer dome or catheter line.
- Transducer is at mid-heart level.
- Pressure transducer is zeroed.
- Invasive blood pressure is selected for display (Main Menu – Monitor Setup – Screen Setup).





Temperature

Check that

- You are using a Datex-Ohmeda temperature probe.
- Temperature probe is positioned correctly.
- Temperature probe is inserted properly into the appropriate temperature connector.
- Temperature is selected for display (Main Menu Monitor Setup Screen Setup).

Airway gases

Check that

- D-fend water trap is properly placed in the monitor.
- Water trap container is empty.
- A new sampling line is used after each patient.
- Sampling line is connected to the water trap and to the airway adapter.
- Monitor is turned on and self-check is performed with the sampling line attached.
- Humidification and/or bacteria filter is placed correctly.
- Breathing circuit or accessories are free of residuals from alcohol-based disinfectants.
- Desired gas parameter is selected for display (Main Menu Monitor Setup Screen Setup).

Patient spirometry

Check that

- Correct flow sensor is used (D-lite for adults and Pedi-lite for pediatrics).
- Correct flow sensor type is selected
 (Main Menu Parameters Airway Gas Spirometry Setup).
- Straight luer connectors of the spirometry tube are attached to the connector on the monitor and the angled luer connectors are attached to the sensor.
- Spirometry tube is not kinked or squeezed.
- D-lite and spirometry tube are free of water drops.
- Humidification and/or bacteria filter is placed correctly.
- Desired spirometry parameter is selected for display (Main Menu - Monitor Setup - Screen Setup).

NeuroMuscular Transmission (NMT)

Check that

- · The application site of the electrodes is cleaned of oil and dirt.
- Electrodes are positioned correctly, electrode gel is moist, and skin contact is good.
 Avoid placement over lesions or excessive body hair.
- NMT is selected for display (Main Menu Monitor Setup Screen Setup).

Messages

Message	Explanation and/or Corrective Action	
Acknowl. alarms silenced	Acknowledged alarms are silenced. (Silence Alarms key pressed during silencing period).	
Air leakage	Air leakage in NIBP cuff or hose. Check all connections and test the cuff tightness using venous stasis.	
Apnea	No change detected in CO ₂ during the last 20 seconds.	
	(Impedance respiration) No change detected in impedance respiration waveform for 20 seconds.	
Apnea deactivated	Apnea alarm is silenced until reactivation after five breaths.	
Artifact(s)	Unsuccessful NIBP measurement. Patient is moving, shivering, or breathing deeply; also marked arrhythmia or irregular beats. Calm the patient and start a new measurement.	
Asystole	Asystole, no QRS detected in ECG.	
	NOTE: This does not depend on the heart rate source.	
Back-up batt. failure	Backup battery is discharged or faulty. Use mains power for 4 hours, then switch to battery power. If the message reappears, contact authorized service personnel.	
Cable off	NMT or regional block cable is not connected.	
Calibrate Agent ID	Agent identification error. Perform gas calibration.	
Calibration not protected	(NIBP) Contact authorized service personnel.	
Calibration switch ON	(NIBP) Calibration protection switch at the bottom of the monitor is turned to the right. Contact authorized service personnel.	
Call service: Error X	NIBP hardware error. Note the error number (X) and contact authorized service personnel.	
Check D-fend	Check that the water trap is properly attached.	
Check Sp02 probe	Pulse search active for 20 seconds but no acceptable pulse is found. Either there is no detectable SpO_2 signal, the sensor is faulty, or it is detached from the patient.	
Check stim. electrodes	NMT stimulus current could not be delivered due to poor stimulus electrode connection or damaged cable.	
Control measurement	NIBP alarm limit was exceeded and a new measurement was started automatically.	
Cuff loose	NIBP cuff is loose or not attached to the patient. Hose is not connected to the monitor.	
Cuff occlusion	Check NIBP cuff hose and tubes; restart measurement. If the problem persists, contact qualified service personnel.	
Cuff over-pressure	NIBP cuff is squeezed during measurement and pressure safety limits are exceeded.	
EEPROM Error	Faulty EEPROM circuit on the CPU board. Contact authorized service personnel.	



Message	Explanation and/or Corrective Action	
EMG electrodes off	(NMT) EMG recording electrodes are off. Electrodes or stimulus clip is loose.	
EtAA high / EtAA low	Measured EtAA is equal to or above/below the alarm limit.	
EtCO2 high / EtCO2 low	Measured EtCO ₂ is equal to or above/below the alarm limit.	
EtO2 high / EtO2 low	Measured EtO ₂ is equal to or above/below the alarm limit.	
FIAA high / FIAA low	Measured FiAA is equal to or above/below the alarm limit.	
FiN20 high	Measured FiN_2O is equal to or above 82%.	
Fi02 high / Fi02 low	Measured FiO_2 is equal to or above/below the alarm limit.	
HR high / HR low	Heart rate (ECG or invasive pressure) or pulse rate (pulse oximetry) is equal to or exceeds the upper/lower alarm limit.	
Infl. limits! Check setup	Adult or child NIBP cuff is being used but the selected infant mode keeps the inflation pressure too low.	
Leads off	One or more cables are disconnected or off the patient. Check ECG cable, all leads, and the neutral electrode (RL/N). Offset voltage between two electrodes is too high.	
	NOTE: This message may appear during defibrillation.	
Long measurement time	NIBP measurement is prolonged over the maximum measurement time (two minutes with adult or child inflation limits; one minute with infant inflation limits). Reapply the cuff, calm the patient, and start a new measurement.	
Measurement off	(NMT) Cable is connected but measurement has not started.	
	(Impedance respiration) Measurement has been turned off.	
Monitor is overheating	Temperature inside the monitor is above the maximum limit. Check the dust filter at the back of the monitor.	
MVexp high / MVexp low	Measured expiratory minute volume value is above/below the set volume limit.	
MVexp< <mvinsp< th=""><th>Exhaled volume is noticeably smaller than inhaled. Check the whole system for leakage.</th></mvinsp<>	Exhaled volume is noticeably smaller than inhaled. Check the whole system for leakage.	
MVexp<0.5 I/min (MVexp<0.2 I/min with Pedi-lite)	Measured volumes are too small for reliable calculation and, for example, waveforms and loops may be unreliable.	
Network connection down	Too many monitors may be active on the network.	
NIBP high / NIBP low	NIBP value is equal to or above/below the alarm limit.	
NIBP manual	Autocycling is interrupted because of air leakage or loose cuff.	
No P1/P2 transducer	Invasive pressure or transducer cable not connected.	
No Sp02 probe	SpO ₂ sensor is not connected to the monitor. Faulty or wrong type of SpO ₂ sensor. Use only sensors approved for use with the type of pulse oximetry installed in the monitor.	
No SpO2 pulse	Pulse signal is found but then lost for 10 seconds or longer. Try a different site.	

Message	Explanation and/or Corrective Action	
Noise	(ECG) Unreliable heart rate (HR) calculation or distorted waveform due to:	
	Too much high frequency or electrical mains noise in ECG.	
	The DC offset voltage between two electrodes is too high.	
	NOTE: This message may appear during defibrillation.	
P1/P2 high P1/P2 low	Invasive pressure of marked channel is equal to or above/below the alarm limit.	
PEEP high / PEEP low	Measured PEEP value is above/below the set pressure limit.	
Performing temp test	Monitor makes a two point calibration test for temperature immediately after the warm-up time and after that once in 10 minutes. Test lasts 10 seconds.	
Ppeak high / Ppeak low	Measured peak pressure is above/below the pressure limit setting.	
Printer error	Printer is out of paper or paper is jammed. Printer is not online or is not turned on. Printer cable is loose or broken.	
Printer failure	Printer is not responding, see Printer error. Select another printer.	
Printing	Laser printer is printing.	
Pulse search	Monitor is searching for pulse oximetry signal.	
RAM Error	Faulty RAM circuit on the CPU board. Contact authorized service personnel.	
Rebreathing	Measured $FiCO_2$ is equal to or above the alarm limit.	
Recorder: cover open	The lid of the recorder is open.	
Recorder: input voltage high Recorder: input voltage low	Call service to check the monitor.	
Recorder: out of paper	Insert a new roll of paper.	
Recorder: system error n	(n can be 1, 2 or 3.) Contact authorized service personnel.	
Recorder: thermal array overheat	Wait for the recorder to cool down.	
Reference not stable	(NMT) Deviation between the four reference stimulation responses is too big, causing reference setting to fail. Can be caused by movement artifact; may occur when patient is relaxed.	
Replace D-fend	D-fend water trap is partially blocked. This decreases air flow.	
Resp high / Resp low	Respiration rate (RR) is equal to or above/below the alarm limit.	
Response too weak	(NMT) The maximum gain is insufficient to increase the response signal amplitude to a measurable level. Can occur if:	
	Stimulation current is too weak. Stimulation electrodes are not connected or they are improperly placed on the nerve. Recording electrodes are disconnected.	
	One or more electrodes are dry and should be replaced.	
	The skin at the electrode site is not properly prepared. Reposition the stimulating or recording electrodes. Change the site of the measuring electrodes.	



Sample line blocked	The gas sampling line inside or outside the monitor is blocked or the water trap is occluded.	
Message	Explanation and/or Corrective Action	
Sensor inop.	Gas measuring sensor is inoperative or the temperature in the monitor has increased. Contact authorized service personnel.	
Setting reference	NMT reference search in progress.	
Small resp. curve	(Impedance Respiration) Signal amplitude is less than 0.4 ohms.	
Snapshot memory full	Creating a new snapshot will erase older ones.	
Sp02 high / Sp02 low	${ m SpO}_2$ is equal to or above/below the alarm limit.	
Sp02 probe off	SpO ₂ sensor is not attached to the patient, site is unsuitable, or sensor is faulty. Check connections; try a new site or sensor.	
SRAM Error	Faulty SRAM circuit on the CPU board. Contact authorized service personnel.	
Supramax not found	(NMT) Supramaximal stimulus current (70 mA) was not found. Stop measurement, reposition the stimulating or recording electrodes, and restart the measurement.	
Supramax search	(NMT) Supramaximal stimulus current search in progress.	
Temperature error	A repeated temperature test has failed. Contact authorized service personnel.	
Tetanic	(NMT) Tetanic stimulation is on.	
Unable to measure Dia	(NIBP) Accurate diastolic pressure not achieved because of artifacts, weak pulsation, etc.	
Unable to measure Sys	(NIBP) Initial inflation pressure is not high enough to enable proper measurement. Systolic pressure is higher than maximum inflation pressure; artifacts interfere in systolic area.	
Unstable zero pressure	Pressure is unstable when starting the NIBP measurement. Calm the patient and retry.	
Voltage error	Erroneous voltage level detected. Contact service personnel.	
Wait for the measurement to be done	Venous stasis is attempted while a regular measurement is under way.	
Weak pulsation	(NIBP) Weak or unstable oscillation signal due to:	
	Improper cuff position or attachment; small air leakage.	
	Weak or abnormal blood circulation.	
	Slow heart rate associated with artifacts.	
	Patient is moving or disturbed during measurement.	
lead off	A lead wire () is disconnected.	

Other situations

Other problem situations which may occur during monitoring, with possible explanations and instructions, are listed below.

Situation	Cause and/or Recommended Action
Extra Asystole alarms	Bad electrode quality or positioning.
	Remove the ECG cable from the monitor and insert again.
	Enlarge the size from 1.0 mV to 2.0 mV .
ECG baseline is thick	ECG lead cable is looped.
	Other electrical power cables are near the ECG wires.
	Incorrect ECG filter.
	Incorrect power frequency of the monitor.
ECG is noisy	Patient is shivering.
	Incorrect ECG filter.
	Bad electrode quality or positioning.
NIBP readings are high	Cuff too small or loose.
	Cuff is under mid-heart level.
Pacer markers are not	Check that:
visible	The pacer markers have been selected 0N .
	The pacemaker was adjusted correctly and is not above R.
	The pacemaker functions correctly; check that ECG cables, electrodes, and setup are correct.
Patient spirometry accuracy is poor—TV less than 50 ml reported as 0 (zero) ml	An adult sensor is used for a pediatric patient and the monitor's spirometry sensor selection is Pediatric . Use a pediatric sensor.
Patient spirometry values are 1/4 of real values	An adult sensor is used for an adult patient, but the monitor's spirometry sensor selection is Pediatric . Change the selection to Adult .
Patient spirometry values are 4 times larger than real values	A pediatric sensor is used for a pediatric patient, but the monitor's spirometry sensor selection is Adult . Change the selection to Pediatric .
Patient spirometry Paw- Vol loop is vertical	A pediatric sensor is used for an adult patient and the monitor's spirometry sensor selection is Pediatric . Use an adult sensor and change the selection to Adult .




Situation	Cause and/or Recommended Action
Pressure readings are high	Air bubbles or a swinging catheter are causing a resonating waveform.
	Wrong zero because the transducer position in relation to the patient has changed (too low).
	Transducer failure.
	Incorrect calibration.
Pressure readings are low	Dampened waveform because of air bubbles, clotting or kinked tubes.
	Transducer is positioned above patient's mid-heart level.
	Transducer failure.
	Incorrect calibration.
Temperature is not	Datex-Ohmeda temperature probe is not being used.
displayed on monitor screen	The temperature may be outside the measurable range (10 to 45 °C).
	If problem persists, contact authorized service personnel.
Cannot see some trend details	Adjust the parameter and time scales to see the desired trend detail.

For your notes:





	9-1
eventive maintenance	9-1
Daily tasks	
ECG and respiration	9-1
Pulse oximetry	
NIBPInvBP	
Temperature	
Airway gases and spirometry	
NMT	
Every month	9-2
Cooling fan dust filter	
Every two months	9-2
Black D-fend	9-:
Every six months	9-2
Gas calibration	
Backup battery check	
Every 12 months	
Planned Maintenance check	
Calibration check of temperature, NIBP and invasive b	
eaning	
	9
Monitor	
ECG cables	9-
	9-
ECG cables	9-
ECG cables Pulse oximetry sensors	9- 9-
ECG cables	9-1 9-1 9-1
ECG cables Pulse oximetry sensors NIBP cuff hose NIBP cuff Cables for invasive pressure	9- 9- 9- 9-
ECG cables Pulse oximetry sensors NIBP cuff hose NIBP cuff Cables for invasive pressure Temperature probes	9-1 9-1 9-1 9-1 9-1
ECG cables Pulse oximetry sensors NIBP cuff hose NIBP cuff Cables for invasive pressure Temperature probes NMT sensor cable, sensors, and regional block adapter.	9- 9- 9- 9- 9- 9-
ECG cables Pulse oximetry sensors NIBP cuff hose NIBP cuff Cables for invasive pressure Temperature probes NMT sensor cable, sensors, and regional block adapter Reusable D-lite sensor	9- 9- 9- 9- 9- 9-
ECG cables Pulse oximetry sensors NIBP cuff hose NIBP cuff Cables for invasive pressure Temperature probes NMT sensor cable, sensors, and regional block adapter	9-1 9-1 9-1 9-1 9-1 9-1
ECG cables Pulse oximetry sensors NIBP cuff hose NIBP cuff Cables for invasive pressure Temperature probes NMT sensor cable, sensors, and regional block adapter Reusable D-lite sensor Airway adapter Sampling line	9-1 9-1 9-1 9-1 9-1 9-1 9-1
ECG cables Pulse oximetry sensors NIBP cuff hose NIBP cuff Cables for invasive pressure Temperature probes NMT sensor cable, sensors, and regional block adapter Reusable D-lite sensor Airway adapter Sampling line D-fend water trap	9-1 9-1 9-1 9-1 9-1 9-1 9-1 9-1
ECG cables Pulse oximetry sensors NIBP cuff hose NIBP cuff Cables for invasive pressure Temperature probes NMT sensor cable, sensors, and regional block adapter. Reusable D-lite sensor Airway adapter Sampling line D-fend water trap. Removing the water trap.	9-1 9-1 9-1 9-1 9-1 9-1 9-1 9-1 9-1 9-1
ECG cables Pulse oximetry sensors NIBP cuff hose NIBP cuff Cables for invasive pressure Temperature probes NMT sensor cable, sensors, and regional block adapter Reusable D-lite sensor Airway adapter Sampling line D-fend water trap	9-1 9-1 9-1 9-1 9-1 9-1 9-1 9-1 9-1 9-1
ECG cables Pulse oximetry sensors NIBP cuff hose NIBP cuff Cables for invasive pressure Temperature probes NMT sensor cable, sensors, and regional block adapter. Reusable D-lite sensor Airway adapter Sampling line D-fend water trap. Removing the water trap.	





9. MAINTENANCE AND CLEANING

Introduction

For safe, reliable function and operation of the monitor, regular care has to be carried out according to the instructions in this manual and the Planned Maintenance procedures described in the Cardiocap/5 *Technical Reference Manual*.

If the monitor does not function as it should, and information in the *Troubleshooting* chapter does not provide help, contact your sales representative. Do not perform cleaning or maintenance procedures other than those described in the Cardiocap/5 manuals.

Preventive maintenance

Daily tasks

Check that all accessories, cables and monitor parts are clean and intact.

Clean the device as described in Cleaning later in this chapter.

Check the parameter items as directed below:

ECG and respiration

The "Leads off" message is displayed when the cable is connected to the monitor and the leads are not connected. Shorting all the leads should make the message disappear (ASY alarm appears).

Pulse oximetry

Sensor functions when properly connected to the patient.

NIBP

Cuff hose detection (Adult/Infant) works properly.

Pump is not restarting in Venous Stasis mode. If it starts, there may be a leak in the cuff.

InvBP

The monitor recognizes the cable connection (activates the display) for all the pressure channels used.

Zeroing of all transducers is working correctly.

Temperature

Measurement starts when a temperature probe is connected.

Airway gases and spirometry

The "Sample line blocked" message appears within 30 seconds after you have occluded the airway adapter; gas waveforms show zero at the same time.

The D-fend water trap is empty. If the green D-fend+ is used (patient with extensive mucus secretion), replace it daily with a new one.

NMT

Check that the electrodes are correctly positioned on the ulnar nerve and the message 'Supramax search' is displayed. Ensure that you get a stimulus response. If the supramaximal stimulus current is not found, the message 'Supramax not found' is displayed. If the current is set manually the message "Setting reference" is displayed directly. Always check the electrode quality.

Every month

Cooling fan dust filter

Check the fan filter on the rear panel of the monitor. Replace the filter if it is damaged.

Clean the fan filter every month or whenever needed:

Pull the filter out.

Wash in detergent solution.

Allow to dry before reinserting. Do not use pressurized air.

Every two months

Black D-fend

Replace the D-fend water trap at least every two months, or when a "Sample line blocked" or "Replace D-fend" alarm message persists.

Every six months

Gas calibration

Perform gas calibration according to the instructions in the Airway Gases chapter.

If gas measurement is in extensive use, calibration is recommended **every two months** to ensure that measurement accuracy stays within specifications.

Backup battery check

The power supply includes a backup battery that saves up to 15 minutes of information.

To check the battery:

Switch on the monitor.

When the battery is fully charged, the charging symbol is not displayed on the screen.

Disconnect the monitor from mains power.

The monitor should function normally for 15 minutes.

Every 12 months


Planned Maintenance check

The Planned Maintenance check detailed in the *Technical Reference Manual* requires trained service person and appropriate testing tools and equipment.

Calibration check of temperature, NIBP and invasive blood pressures

Calibration check of temperature, NIBP and invasive blood pressures should be performed at least once a year by qualified service personnel as a part of the Planned Maintenance, see "Technical Reference Manual."

Cleaning



Method	Why?	Which items?	How?
Cleaning	To remove dirt and most microbes.	All parts daily. Items to be disinfected or sterilized.	Wipe away visible organic material. Rinse as recommended by the detergent manufacturer. Allow to dry completely.
Disinfecting	To kill and remove pathogenic microbes.	Accessories and supplies (other than SpO ₂ sensors) in contact with the patient's skin or airways. NOTE: Some SpO ₂ sensors cannot be disinfected. Read the sensor instructions before attempting to disinfect a sensor.	With heat: Wash in a washing machine at 85 °C (185 °F) for at least one minute. With chemicals: Immerse in 2% glutaraldehyde for 10 minutes (pH=6.5). After tuberculosis, HIV, or hepatitis patient, immerse for 20 minutes. Rinse carefully or wipe with alcohol and let dry.
Sterilizing	To kill microbes and bacterial spores.	Accessories and supplies penetrating the patient's tissue.	Steam autoclave for 15 minutes at 121° C (250 °F) maximum. Use ethylene oxide mixture at 50 to 60°C (120 to 140°F). Follow the sterilizer manufacturer's recommendations for required aeration times. Immerse in 2% glutaraldehyde for three hours, rinse carefully, and let dry.

WARNING: Electrical Shock Hazard. Before cleaning, disconnect the monitor from the electrical outlet.

WARNING: Electrical Shock Hazard. After cleaning, or if liquid has accidentally entered the interior of the monitor, make sure that every part of the monitor is dry before reconnecting it to the power supply.

CAUTION: Do not sterilize any part of the monitor with steam autoclave or ethylene oxide.

CAUTION: Do not immerse any part of the monitor in liquids or allow liquid to enter the interior.

CAUTION: Do not use hypochlorite, acetone-based, phenol-based, or ammonia-based cleaners.



Monitor

To clean the monitor:

3. When the display surface becomes dusty, wipe it gently with absorbent cotton, chamois or other soft material. The display may also be cleaned using a small amount of normal hexane. Do not use acetone, toluene or alcohol because they cause chemical damage to the polarizer.

Wipe the monitor with a mild detergent solution such as the Datex-Ohmeda Cleaning Fluid. Do not leave liquid spills on any metal part.

Let dry completely before connecting to power source.

The internal sampling system does not need to be cleaned nor sterilized. The D-fend water trap functions as a bacteria filter and there is no reverse flow to the patient.

ECG cables

Wipe the cables with a mild detergent solution. Disinfect when necessary.

Pulse oximetry sensors

WARNING: Patient Safety. A damaged sensor or a sensor soaked in liquid may cause burns during electrosurgery.

Disposable sensors are for single-patient use only. Do not attempt to clean them.

For cleaning instructions and recommended cleaning agents for a reusable sensor, consult the instructions for using that sensor.

Consider possible patient allergies when selecting the cleaning agent.

In general, to clean any reusable sensor or sensor cable, detach it from the patient and the monitor. Wipe it with a pad moistened with a mild cleaning solution, such as 70% isopropyl alcohol. Allow the sensor and/or cable to dry completely before use.

NIBP cuff hose

Wipe the cables and hose with a mild detergent solution. Disinfect when necessary.

NIBP cuff

Clean only when necessary. Remove the bladder from the cuff. Wash the bladder and the cuff in mild detergent solution. Do NOT use alcohol.

Cables for invasive pressure

Wipe the cables with a sterile alcohol-based detergent. After cleaning, rinse surfaces by wiping them with a cloth damped with sterile water. Dry with a dry cloth.

Temperature probes

Clean with a mild detergent solution and rinse with water. Disinfect or sterilize when necessary.

NMT sensor cable, sensors, and regional block adapter

Wipe the cable, sensor, and adapter with mild detergent solution.

If disinfection is required, use ethanol or isopropyl alcohol. Do not use hypochlorite. Do not immerse in liquid.

Reusable D-lite sensor

Reusable D-lite sensors can be washed in a dishwasher. After cleaning, allow the sensor to dry thoroughly.

Make sure that the sensor is intact. Connections must be tight for exact measurements.

Airway adapter

Replace the single-use adapter after each patient.

A reusable adapter can be disinfected with glutaraldehyde or alcohol. A reusable steel adapter may be autoclaved.

If you want to clean the adapter before use, submerge it in 70% alcohol solution for 30 seconds and rinse carefully with water.

Rinse away all traces of alcohol or detergent. Dry the adapter before connecting to the patient.

Sampling line

Do not reuse the sampling line. Attempting to clean and reuse a sampling line may affect measurement results.

D-fend water trap

The D-fend water trap contains a hydrophobic membrane that prevents water and secretions from entering the measuring chamber. Condensed water and saliva are collected into a washable container.

The green D-fend+ is for patients with extensive mucus secretion and for single-patient use only. Replace it every 24 hours or when a "Sample line blocked" or "Replace D-fend" message persists.

The water trap container can be cleaned with disinfecting solutions or sterilized using cold chemicals or ethylene oxide.

CAUTION: Do not disinfect or open the water trap cartridge. To avoid damage, do not touch or try to clean the water trap hydrophobic membrane (other than to rinse it with water).



To lengthen the lifetime of your monitor and minimize downtime:

- Empty the water trap container whenever it is more than half full.
- Do not open, wash, or sterilize the D-fend water trap cartridge.
- After washing or disinfecting the airway adapter or water trap container, make sure no alcohol or detergent remains. Traces of alcohol or other organic cleaning solutions may affect measurement.
- Do not force air or oxygen through the D-fend; do not allow smoke or dust to enter the D-fend.
- While administering nebulized medication, disconnect the gas sampling line from the patient circuit for 30 minutes.
- If a "Sample line blocked" alarm occurs, replace the sampling line and empty the water trap container.

Removing the water trap




Figure 9-1. Emptying the water trap container

To remove the water trap:

4. Push the water trap latch to the right. The water trap is spring-loaded. The "Check D-fend" message appears.

Detach the container from the water trap cartridge by pulling it carefully downward.

Empty and clean the container.

Attach the container back into the cartridge tightly.

Push the whole unit into its housing on the front panel until the latch is set.

Press the **Normal Screen** key to restart monitoring. Check that the "Check D-fend" message is no longer displayed.

Other accessories

See the accessory packages for checkout and cleaning instructions. Do not reuse disposable accessories.

For your notes:



Chapter 10. ECG

Overview	10-1
ECG and heart rate display	10-1
Patient connections	10-2
Patient preparation	10-2
Placing electrodes	10-2
Cable set	10-3
Color and letter coding	10-4
Connecting the lead wires and the trunk cable	10-4
ECG setup menu	10-5
Selecting a lead	10-7
Viewing a cascaded ECG	10-7
Adjusting ECG size	10-8
Adjusting beat sound volume	10-8
ST segment analysis	10-9
Monitoring the ST segment	
Accessing ST analysis view	
ST display in a digit field	
Setting ST point manually	
Setting ISO point manually	10-11
Setting ECG alarms	10-11
Pacemaker markers	10-12
Other adjustable features	10-13
Analog signal output	

Table of Figures

Figure 10-1. ECG and HR display	10-1
Figure 10-2. Standard electrode positioning with 3-lead set	10-2
Figure 10-3. Modified electrode positioning, CB ₅ , with 3-lead set	10-2
Figure 10-4. Standard electrode positioning with 5-lead set	10-3
Figure 10-5. ECG setup	10-3
Figure 10-6. ST analysis view	10-10
Figure 10-7 ST number field	10-10







10. ECG

Overview

The electrocardiograph, ECG, reflects the electrical activity generated by the heart muscle. ECG monitoring is used for heart rate measurement and to detect arrhythmias, pacemaker function, and myocardial ischemia.

When you use a 5-lead set, you may monitor the waveforms of up to three different ECG leads. When you use a 3-lead set, the monitor displays one ECG lead.

When monitoring ECG, the monitor simultaneously analyzes ST segment changes. Changes of up to three different ECG leads are analyzed depending on your lead set.

ECG and heart rate display




Figure 10-1. ECG and HR display

- (1) ECG1 displayed first
- (2) ECG2 displayed below ECG1
- (3) ECG3 displayed below ECG2
- (4) Selected lead label
- (5) ECG gain bar (1 mV reference)
- (6) Heart rate (HR) label
- (7) Heart rate calculation source and message field for HR messages
- (8) HR value is always displayed with ECG1
- (9) Heart beat detector is flashing with every detected heart beat
- (10) ST values appear next to ECG1; the lead with the largest absolute ST value is also displayed
- (11) Selected ST leads
- (12) ST label
- (13) Message field for parameter messages

NOTE: You can choose the number of ECG waveforms on the display in the Monitor Setup menu (select **Screen Setup** and **Waveform Fields**). You choose the ECG leads in the ECG menu.

Patient connections

Patient preparation

Good signal quality requires good skin contact with the electrodes. Excessive body hair or skin oil can affect the contact. Pre-gelled electrodes are recommended. Check that the electrodes are moist and have not dried out during storage.

WARNING: Make sure the lead set clips or snaps do not touch any electrically conductive material including earth.

Placing electrodes

When placing the electrodes, avoid bones close to the skin, obvious layers of fat, and major muscles.




Figure 10-2. Standard electrode positioning with 3-lead set

You can use a CB_5 -lead when you want the 3-lead set connection to resemble the 5-lead set V5 connection. The red (IEC standard) or white (AAMI standard) electrode, which is shaded, is on the back.




Figure 10-3. Modified electrode positioning, CB₅, with 3-lead set

When you use routine positioning with a 5-lead set, place the chest electrode on one of the six places indicated in the picture. If the lead set follows the IEC standard, the chest electrode is white; if the set follows the AAMI standard, the chest electrode is brown.




Figure 10-4. Standard electrode positioning with 5-lead set

Cable set




Figure 10-5. ECG setup

- (1) ECG electrode(s)
- (2) Trunk cable
- (3) 3-lead set or 5-lead set

WARNING: Patient Safety. Ensure proper contact of the return electrode of the electrosurgery unit to your patient to avoid possible burns on the patient via ECG electrodes and probes.

NOTE: The module input circuits are protected against the effects of electrosurgery and defibrillation. However, the ECG trace on the monitor screen may be disturbed during electrosurgery.



Color and letter coding

3-LEAD IEC STANDARD	5-LEAD IEC STANDARD
L = yellow (left arm) F = green (foot) R = red (right arm)	R = red (right arm) N = black (neutral) L = yellow (left arm) F = green (foot) C = white (chest)

3-LEAD AAMI STANDARD	5-LEAD AAMI STANDARD
LA = black (left arm) LL = red (left leg) RA = white (right arm)	RA = white (right arm) RL = green (right leg) LA = black (left arm) LL = red (left leg) V = brown (chest)

Connecting the lead wires and the trunk cable

To make lead wires and trunk cable connections:

- 1. Attach the lead wires to the corresponding electrodes, paying attention to the color codes.
- 2. Connect the lead wire set to the trunk cable and ensure a proper connection.
- 3. Plug the trunk cable into the corresponding monitor connector.

It takes about 10 minutes for the electrodes to stabilize. During this time errors may occur in the signal.







ECG setup menu

To open the ECG setup menu:

- 1. Press the **ECG** key.
- 2. Select ECG Setup.





HR Source

Selects the heart rate source. When **ECG** is selected, the HR source is always calculated from ECG 1. If the ECG signal is affected by too much noise for a reliable heart rate calculation, heart rate can be calculated from pressure (Art and ABP) or the plethysmographic pulse waveform.

The selected heart rate source is shown above the numerical heart rate display. The heart rate color is the same as the source parameter.

The **AUTO** selection priorities for heart rate calculation are: ECG, Pressure (Art or ABP), and Plethysmographic pulse waveform. The first heart rate source available is selected.

Display with HR

Select **PR** to display combined Heart Rate and Pulse Rate next to the ECG waveform. The current HR source is displayed in a larger font size and the QRS symbol flashes next to the reading.



Filter

Filters the ECG signal high frequency noise and slow respiratory artifacts. The selections are **Monit**, **Diagn**, and **STfilt**.

- Monit (monitor) filter is used in routine monitoring. It effectively
 filters the high frequency artifacts caused by the electrosurgery
 unit and respiration, for example.
- Diagn (diagnostic) filter is used if more accurate information of the waveform is needed (e.g., of P-wave or AV block). The diagnostic filter is more susceptible to both high frequencies and baseline wander than the monitor filter.
- STfilt (ST filter) permits more accurate of ST segment information.
 It filters the high frequency artifacts caused by an electrosurgery unit but catches the slow changes in ST segment. The ST filter is more susceptible to baseline wander than the monitor filter.

5-lead Cable

This selection is visible only when a 5-lead trunk cable is connected to the monitor. Select **5elect** if you are using 5 electrodes or **3elect** if you are using 3 electrodes with the 5-lead trunk cable.

ECG Grid

Selects a background to be shown on the ECG waveforms, making it easier to evaluate the waveform. The grid scale is 0.5 mV.

Pacemaker

Selects how the pacing pulse of a cardiac pacemaker is displayed. The selections are **Show**, **Hide**, **ONR** and **Sensit**.

- **Show**—the pacer pulse is filtered away from ECG data but the pulse is displayed as a constant height marker.
- Hide—the pacer pulse is filtered away from ECG data.
- ON R—pacer pulses are not filtered away from ECG data. This
 improves ECG monitoring with A-V pacemaker patients, as QRS
 complexes are counted even if the pacer pulse hits the QRS
 complex. However, during asystole the monitor may count pacer
 pulses as heart beats.
- Sensit—uses a more sensitive pacemaker detection. A pacemaker spike is displayed on ECG.

QRS Type

QRS detection and HR calculation may be affected by the shape of the QRS complex. **Normal** QRS type detects a heart beat between 40 and 120 ms. **Wide** QRS type detects a beat between 40 and 220 ms.






Selecting a lead

The following lead selections are possible:

- With 3-lead set: I, II, III
- With 5-lead set: I, II, III, aVR, aVL, aVF and V5

To select the ECG1 lead:



Press the **ECG** key.

Select ECG1 Lead.

You can select all the leads (ECG1, ECG2 and ECG3) in the ECG menu.



The label of the lead is displayed in the ECG field and above the numeric HR.

3.

4.

NOTE: It is possible to connect a 3-lead set to a 5-lead trunk cable. The combination functions as a 3-lead set.

Viewing a cascaded ECG

If more than one ECG waveform is configured to be displayed in the waveform fields when using a 3-lead set, the same ECG is displayed in each of these fields. Thus, more QRS complexes are displayed at the same time.


With a 5-lead set, a cascaded ECG or different leads can be displayed. Select the different leads in the ECG menu.

Adjusting ECG size

Increasing or decreasing the ECG gain affects the size of the 1 mV bar at the left end of the ECG waveform, and the size of the ECG waveform accordingly.

6.

To make adjustments in ECG size:



Press the ECG key. 5. Select Size.




Adjusting beat sound volume

When the monitor detects a QRS complex or a pulse (pleth or pressure), a beep tone sounds. You can raise, lower, or turn off the beat sound volume. Adjusting the beat sound volume also affects the SpO₂ beat sound volume.

When SpO₂ is monitored, the monitor provides a variable pulse beep, so that the tone of the pulse beep rises with increasing oxygen saturation and falls as saturation decreases. This also affects the ECG tone.

To adjust the beat volume:



7. Press the ECG key.

Select Beat Sound Volume 8. and turn the ComWheel to choose the volume. The range is from 1 (silent) to 10 (loud) or 0







ST segment analysis

The ST value illustrates the electrical activity difference between ISO points and ST points. The points are based on R-wave detection.

Myocardial ischemia manifests itself in the ECG by causing the ST segment to deviate from the isoelectric line (ISO point). The ST segment generally rises above the PQ isoelectric line in the presence of transmural ischemia and is pressed below the isoelectric line in the subendocardial ischemia. Ischemic heart disease may result in myocardial infraction, fatal arrhythmias, or acute coronary insufficiency.

ST segment changes may also be affected by other factors than myocardial ischemia. These factors can be, for example, drugs, or metabolic or conduction disturbances.

Monitoring the ST segment

The monitor continuously analyzes the ST segment changes in active leads. ST numerical data is shown on the right of the second real-time ECG waveform field.

When monitoring begins or when a lead is changed, the monitor starts to learn the ST segment. During a learning period of 16 accepted beats, the ST values are displayed beat to beat.


Optimal leads for ST analysis are

- With 5-lead set: II, aVF and V5
- With 3-lead set: CB₅.

NOTE: Select the ST filter (in the ECG Setup menu) as well as the monitored leads (in the ECG menu) at the beginning of the case. Changing the filter and/or the lead affects the ST measurement values.

Accessing ST analysis view

To access the ST analysis view:



When the ST analysis view is accessed, the display shows three measured ECG leads and an averaged QRS complex for each ECG lead. In addition, a graphic trend display for each ECG lead is shown. Changing the lead causes a new learning of the ST segment. The learning period is shown as a line in the trend display.





Figure 10-6. ST analysis view

- (1) ST view showing QRS complex average for each lead with ST and ISO point cursors
- (2) Lead label, ST values
- (3) ECG waveform area showing 3 ECG leads
- (4) A trend of ST values of each selected lead
- (5) Current ST values displayed continuously

ST display in a digit field

You can also select ST numerical data to the lower digit field.



- (1) Message field
- (2) The lead that has the largest absolute ST value
- (3) ST values of ECG1, ECG2, and ECG3
- (4) Lead label

Figure 10-7. ST number field

To do this:

- 11. Press the ComWheel and select Monitor Setup from the Main Menu.
- 12. Select Screen Setup.
- 13. Select Digit Fields.






Setting ST point manually

In the monitor, the ST point is initially set 80 ms after the R detection. If this setting is changed, the original point is shown as a dashed line. If the ST point is set manually, the set value remains until the monitor is turned off or reset.

To set the ST point:




- 14. Press the **ECG** key.
- 15. Select ST Analysis.
- 16. Select **Set ST point**.
- Set the ST point by turning the ComWheel. To confirm the setting, press the ComWheel.



Setting ISO point manually

The ISO (isoelectric) point is initially set 74 ms before the R detection. If this setting is changed, the original point is shown as a dashed line. If the ISO point is set manually, the set value remains until the monitor is turned off or reset.

To set the ISO point:



- 18. Press the **ECG** key.
- 19. Select ST Analysis.
- 20. Select Set Set ISO point.
- 21. Set the ISO-point by turning the ComWheel. Confirm the setting by pressing the ComWheel.




NOTE: The ST algorithm has been tested for accuracy of the ST segment data. The significance of the ST segment changes needs to be determined by a clinician.

Setting ECG alarms

You can set the ST alarm limit and change the high and low heart rate alarms from the ECG menu (as described below). You can also make the adjustments by selecting **Adjust Limits** in the Alarms Setup menu.

To adjust ST and heart rate alarms:



22. Press the ECG key.
23. Select ECG Alarms.
24. Select Adjust Limits and make the desired adjustments.



Monitoring pacemaker patients

- 1. Press the ECG key.
- 2. Select ECG Setup Pacemaker and select one of the following:
 - Show = Pacemaker spike is displayed on ECG.
 - Sensit = Sensitive pacemaker detection; spike displayed on ECG
 - ON R = Pacemaker suppression weakened; asystole alarm may not be reliable with active pacemakers.
 - Hide = Pacemaker spike is not displayed on ECG.

NOTE: Pacemaker detector may not operate correctly during the use of high-frequency (HF) surgical equipment. The disturbances of HF surgical equipment typically cause false positive pacer detection.

NOTE: The pacemaker may change the shape of the QRS complex so much that QRS detection may be affected.

WARNING: Do not rely entirely upon rate meter alarms when monitoring patients with pacemakers. The monitor may count the pacemaker pulses as heartbeats. In this case, asystole and ventricular fibrillation may go undetected. Always keep these patients under close surveillance and monitor their vital signs carefully.





This section describes other adjustable features related to ECG measurement. A password is required for entering the Install/Service menu where the adjustments are made. If you wish to change a feature, we recommend that you contact the person responsible for the entire configuration.

Analog signal output

WARNING: Output signals are not floating and must not be connected directly to a patient.

- 25. Press the ComWheel and select Monitor Setup from the Main Menu.
- 26. Select install/Service and enter the password.
- 27. Select Installation.
- 28. Select Analog Outputs.

For more information, refer to the Technical Reference Manual.

For your notes:



Chapter 11. Impedance Respiration

	Overview	11-1
	Respiration detection	11-1
	Respiration rate calculation	
	Impedance respiration display	11-2
	Patient connections	11-3
	Activating the respiration measurement	11-4
	Turning the measurement ON	11-4
	Turning the measurement OFF	11-4
	Choosing the respiration rate source	11-5
	Improving waveform readability	11-5
	Improving respiration rate detection	11-6
	Measurement limitations	11-6
	Movement artifacts	11-6
	Electrical interference Pacemaker patients	
T	able of Figures	
	Figure 11-1. Respiration rate source indication for impedance and ${ m CO}_2$	
	Figure 11-2. Monitor display with impedance respiration measurement	
	Figure 11-3. Impedance respiration waveform and respiration rate numeric	
	Figure 11-4. Impedance respiration setup	11- 3




11. IMPEDANCE RESPIRATION

Overview

Impedance respiration is measured across the thorax. When the patient is breathing or is ventilated, the volume of air changes in the lungs, resulting in impedance changes between the electrodes. A respiration rate is calculated from these impedance changes, and a respiration waveform is displayed on the monitor screen.

Respiration detection

The respiration rate is the sum of respirations that exceed the detection limit.



The dotted lines are the zero line and the detection limit. The signal strength produced by a respiration should thus exceed this minimum limit to be included in the respiration rate calculation. Peaks within the grids are not calculated.

If **AUTO** detection mode is chosen, the grid lines are the minimum limits but the limits in use may be larger. The RR value can include fewer respirations than indicated by the gridline.

Respiration rate calculation

Respiration rate is calculated automatically when ECG is measured, unless respiration measurement is turned off or CO_2 is measured.

When CO_2 is measured (only possible with F-MXG model), the respiration rate is automatically calculated from the CO_2 .

Respiration rate calculation switches back to impedance respiration if you press the **Silence Alarms** key during an Apnea alarm.

The respiration rate source is displayed above the respiration numeric value.





Figure 11-1. Respiration rate source indication for impedance and CO₂

The impedance respiration waveform is also displayed next to the RR value when the RR value is calculated from CO_2 .

Impedance respiration display




Figure 11-2. Monitor display with impedance respiration measurement




Figure 11-3. Impedance respiration waveform and respiration rate numeric value

Patient connections

The setup for impedance respiration is the same as for ECG measurement. If you do not have the ECG setup, see *Patient connections* in the *ECG* chapter.

You can use 3-lead or 5-lead ECG sets.




Figure 11-4. Impedance respiration setup

- (1) ECG electrode(s)
- (2) Trunk cable
- (3) Lead set (3-lead or 5-lead set)

NOTE: Use of the impedance respiration measurement is recommended only for patients over three years old.

WARNING: In obstructive apnea, respiration movements and impedance variations may continue.


WARNING: Patient Safety. Ensure proper contact of the return electrode of the electrosurgery unit to avoid possible burns at sensor sites.


WARNING: Make sure that the lead set clips or snaps do not touch any electrically conductive material including earth.

Activating the respiration measurement

The respiration information is not included in the trends and the alarms are not operative unless respiration is selected in a waveform or in a digit field.

To select respiration in a waveform or a digit field:






- Press the ComWheel and select Monitor Setup from the Main Menu.
- 2. Select Screen Setup.
- Select
 Waveform Fields or Digit Fields.
- 4. Choose Resp.



Turning the measurement ON

To turn on the impedance respiration measurement:



- Press the ComWheel and select **Parameters** from the Main Menu.
- Select **Resp Setup**.
- 7. Select **Measurement** and choose **ON**.



Turning the measurement OFF

In case the impedance respiration measurement signal interferes with other measurements, such as ECG, you can turn it off:

8. Press the ComWheel and select Parameters from the Main Menu.

6.


- 9. Select Resp Setup.
- 10. Select Measurement and choose OFF.





Choosing the respiration rate source

To manually choose the respiration rate source:




- 11. Press the ComWheel and select **Parameters** from the Main Menu.
- 12. Select Resp Setup.
- 13. Select Resp Rate
 Source and choose AUTO, CO2, or Imped.



Improving waveform readability

To improve the readability of the waveform, increase its size:



- 14. Press the ComWheel and select **Parameters** from the Main Menu.
- 15. Select Resp Setup.
- 16. Select **Size** and turn the ComWheel to change the size of the waveform.

The bar at the left of the waveform always indicates a 1 Ω reference.






Improving respiration rate detection

When respirations are weak, you can adjust the detection limits (measurement sensitivity) closer to each other to ensure that all respirations are included in the RR value. In this case, the dotted line represents the absolute detection limits.

When there are numerous artifacts, you can separate smaller artifacts from larger, true respiration peaks by adjusting the grids farther apart. The small peaks fall within the grids and are not calculated; the larger peaks cross the grids and are calculated as true respirations.

During spontaneous breathing, the ventilator may support the patient's ventilation with an extra inspiration. If these ventilator inspirations are substantially larger than the spontaneous breaths, the respiration calculation may mistakenly count only the ventilator-produced inspirations and expirations. You can correct this by adjusting the detection limit.

To adjust the detection limit:



- 17. Press the ComWheel and select **Parameters** from the Main Menu.
- 18. Select Resp Setup.
- Select Detection Limit.

The percentage is the ratio to the 1 Ω reference bar, which is 100%.



Measurement limitations

Movement artifacts

Changing the patient position, moving the head or the arms, or shaking will result in movement artifacts. Also, the heart may cause noticeable movement and sometimes this may interfere with the respiration measurement.

Electrical interference

Electrical devices that emit electromagnetic disturbance, such as electrosurgery units and infrared heaters, will result in artifacts or may disable the respiration measurement.


Pacemaker patients

WARNING: Pacemaker Patients. The impedance respiration measurement may cause rate changes in Minute Ventilation Rate Responsive Pacemakers. Set the pacemaker rate responsive mode off or turn the impedance respiration measurement off on the monitor.



Chapter 12. Pulse Oximetry

Overview	12-1
Pulse oximetry options	12-1
Nellcor® sensors	12-1
Pulse oximetry display	12-2
Opening the pulse oximetry menu	12-2
Menu selections	12-3
Plethysmographic pulse wave	12-3
Patient connections	12-4
Selecting the sensor	12-4
Choosing a site	12-4
Connecting the sensor	12-5
Disconnecting the sensor	12-5
During monitoring	12-6
Measurement limitations	12-6
Interfering substances	12-6
NIBP measurement and cold infusions	
ble of Figures	
Figure 12-1. Display of SpO ₂ value and Pleth waveform	12-2





12. PULSE OXIMETRY

Overview

Oxygen saturation (SpO₂) is the percentage of saturated hemoglobin compared to total hemoglobin as measured by a two-wavelength pulse oximeter (also called functional or in vivo oxygen saturation).

 ${\rm SpO_2}$ is measured with light absorption techniques: red and infrared light are emitted from the emitter side of the sensor. The light is partly absorbed when it passes through the monitored tissue. The amount of transmitted light is detected in the detector side of the sensor. When the pulsative part of the light signal is examined, the amount of light absorbed by arterial hemoglobin is discovered and the saturation level is calculated.

The plethysmographic pulse wave is derived from variations of the intensity of the transmitted light. It reflects the blood pulsation at the measuring site. Thus, the amplitude of the waveform reflects the perfusion.

NOTE: See the pulse oximetry product specifications in chapter 1 for more details.

Pulse oximetry options

The Cardiocap/5 monitor offers three oxygen saturation measurement technologies:

- Datex-Ohmeda standard pulse oximetry.
- Datex-Ohmeda enhanced pulse oximetry (N-XOSAT option).
- Nellcor® compatible pulse oximetry (N-XNSAT option).

NOTE: Use only sensors identified for use with a specific option. For example, use only Nellcor® sensors with the N-XNSAT option.

Consult the *User's Guide* for appropriate Datex-Ohmeda sensors. Datex-Ohmeda Sat sensors and OxyTip®+ sensors are latex-free.

Nellcor® sensors

Use the sensors listed below with the N-XNSAT option (* indicates latex-free).

Oxygen Transducer Name	Patient Type	Order Number	SpO ₂ Accuracy (± 1 SD)
Oxisensor® II	Adult	D-25 * / D-25L *	
	Pediatric	D-20 *	70-100% ± 2 digits
	Infant	I-20 / I-20LF *	
	Adult Nasal	R-15	80-100% ± 3.5 digits
OxiCliq®	Adult	A *	
	Pediatric	P *	$70-100\% \pm 2.5 \text{ digits}$
	Infant	I	
Durasensor® II	Adult	DS-100A *	
Dura-Y® II Multisite	Adult, Pediatric, Infant	D-YS	70-100% ± 3 digits
Oxiband® II	Pediatric, Infant	OXI-P/I *	

Pulse oximetry display





Figure 12-1. Display of SpO₂ value and Pleth waveform

- (1) Scale of Plethysmogram
- (2) Label(s)
- (3) Oxygen saturation (SpO₂) value
- (4) Pulse oximetry message field

Opening the pulse oximetry menu

NOTE: Models with the Invasive Pressures or NMT option do not have the ${\it Pulse}$ ${\it Oximetry}$ direct access key.

To open the pulse oximetry menu:



or

2. Press the ComWheel and select Parameters from the Main Menu. Then, select Pulse Oximetry.



Menu selections

Pleth Scale The scale of the plethysmographic waveform display is set

automatically during "Pulse search." To adjust it further,

select 2, 5, 10, 20, 50, or AUTO.

NOTE: This selection is not available for the N-XNSAT or N-XOSAT

options, both of which are preset to AUTO.

When AUTO scaling is used, the scale changes automatically if the amplitude of the pleth waveform exceeds the current scale or falls below the maximum value of the next lower scale by 10% for 30 seconds or more. When the scale changes, the "Scale changed"

message is displayed.

The scale indicator number appears at the left of the waveform.

SpO_2 Response Selects the SpO $_2$ averaging time: b-to-b (beat to beat), Normal

(10 seconds), or **Slow**: (20 seconds, the default setting).

NOTE: This selection is not available for the N-XNSAT option, which is preset to 5–7 seconds, or for the N-XOSAT option, which is

preset to 12 seconds.

Beat Sound Volume Adjusts the SpO2 beat volume. This adjustment also affects the ECG

beat volume.

When SpO₂ is monitored, the tone of the pulse beep rises as oxygen

saturation increases and falls as it decreases.

HR Source Selects the heart rate source. If the ECG signal is affected by too

much noise for a reliable heart rate calculation, the heart rate can also be calculated from the invasive pressure (Art) or from the

plethysmographic pulse waveform (Pleth).

The selected heart rate source is shown above the numerical display of the heart rate. The color of the heart rate is the same as the source

parameter.

If **AUTO** is selected, the heart rate calculation priorities are:

ECG (the lead with the highest R-wave), invasive pressure (Art), then

Plethysmographic pulse waveform (Pleth).

Sp02 Alarm Select to adjust the SpO_2 alarm limits as described in the Alarms

chapter.

Plethysmographic pulse wave

If the site is poorly perfused, use smaller scale indicators to display an optimal pulse wave. A small pulse wave may be a sign of impaired circulation that may require increased attention.

Higher scale indicators together with a well-defined pulse wave indicate strong circulation and a relaxed patient.

Patient connections

Selecting the sensor

Choose a sensor that is appropriate for the patient and the monitoring requirements. Sensors are categorized as reusable (durable) or disposable (single-patient use). Examples of both types are shown and described below.

Reusable Sensors	Description
	Finger sensor. Widely-used. Quick application is possible; some may be used on toes.
	Ear sensor. Similar in appearance to finger sensor, but smaller. Used on upper or lower part of ear.
	Wrap sensor. Flexible sensor that wraps around the site (usually fingers, toes, or the fleshy part of a hand or foot). Useful for small children.
Disposable Sensors	Description
	Adhesive sensor that wraps around a finger or toe. A separate sensor cable snaps into a connector on the sensor.
SZ-0 0 0 NEILCOR	Adhesive sensor with integrated cable; wraps around a finger or toe.
	Adhesive sensor with integrated cable. The sensor is positioned inside tape that is wrapped around the site to hold the sensor in place. This sensor is used on fingers, toes, and the fleshy part of a hand or foot, depending on the patient.

A finger sensor is usually appropriate for short-term patient monitoring. For long-term monitoring, a single-patient, disposable sensor or a wrap-type sensor is commonly used.

WARNING: Patient safety. Use clean and dry sensors and cables only. Moisture and debris on connectors may affect measurement accuracy.

Choosing a site

Site selection depends on the type of sensor and the weight of the patient. When choosing a site for a sensor, refer to the instructions for that sensor.

Choose a well-perfused site on a nondominant hand or foot, or the fleshy upper or lower part of the ear when using an ear sensor.


- · Clean the application site you choose, if necessary.
- Finger or toe-remove nail polish and artificial fingernails; clip long fingernails.
- Ear—remove earrings.




Connecting the sensor

WARNING: Patient safety. To prevent erroneous readings, do not use an inflated blood pressure cuff or arterial blood pressure measurement device on the same limb as the oximeter sensor.

3. Connect the sensor (or sensor extension cable) to the SpO₂ connector on the monitor.



4. If applicable, connect the sensor cable to the extension cable.



- 5. Position the sensor and attach it to the patient. For proper sensor positioning and application, consult the instructions that accompany the sensor.
- 6. To minimize movement, attach the sensor cable to the wrist or bed clothes.

The "Pulse search" message appears. After the pulse search is completed, the plethysmographic pulse waveform and the SpO_2 reading appear on the screen.

Disconnecting the sensor

When disconnecting a sensor, grasp the connector, not the cable. Some connectors contain "buttons" on each side of the connector that you press to release the connector.

During monitoring

WARNING: Patient safety. Do not use a pulse oximeter with magnetic resonance imaging (MRI).

WARNING: Patient safety. A damaged sensor or a sensor soaked in liquid may cause burns during electrosurgery.

WARNING: Patient safety. Patient conditions (such as reddening, blistering, skin discoloration, ischemic skin necrosis, and skin erosion) may warrant changing the sensor site frequently or using a different style of sensor. For details, refer to the instructions supplied with the sensor.

Patient condition or prolonged use may require changing the sensor site periodically. Check skin integrity, circulatory status, and correct alignment.

For patients with poor peripheral blood circulation, change the site at intervals of 30 minutes to one hour. To confirm the circulatory status, observe the size of the plethysmographic waveform with a fixed pleth scale.

Be especially careful when monitoring infants.

Measurement limitations

Use Cardiocap/5 pulse oximetry only for patients weighing 5 kg (11 lb.) or more, even if the SpO_2 sensor can be used for patients weighing less than 5 kg.

Interfering substances

WARNING: Data validity. Conditions that may cause inaccurate readings and impact alarms include interfering substances, excessive ambient light, electrical interference, ventricular septal defects (VSD), excessive motion, low perfusion, low signal strength, incorrect sensor placement, poor sensor fit, and/or movement of the sensor on the patient.

The pulse oximeter cannot distinguish between oxyhemoglobin and dyshemoglobins (methemoglobin or carboxyhemoglobin, for example). Dyes that change usual arterial pigmentation, or substances containing dyes, may also cause erroneous readings.

Carboxyhemoglobin may erroneously increase readings in all pulse oximeters. The level of increase is approximately equal to the amount of carboxyhemoglobin present. Cigarette smokers and victims of smoke inhalation often have increased levels of carboxyhemoglobin. Therefore, the saturation values may be somewhat higher for smokers and for patients with carbon monoxide (CO) intoxication.

NIBP measurement and cold infusions

Do not attach the pulse oximetry sensor to a limb that is used for noninvasive blood pressure measurement or for administrating cold infusions. These circumstances reduce the pulsation and perfusion and may affect the pulse oximetry measurement.



Chapter 13. Non-Invasive Blood Pressure (NIBP)

Overview	13-1
NIBP display	13-1
Patient connections	13-2
Selecting the cuff hose	
NIBP setup menu	13-3
NIBP calibration check	
Starting NIBP measurement	13-4
During measurement	13-4
Automatic NIBP double-check	
Canceling a measurement	13-5
Autocycling	13-5
Setting the cycle time	13-5
Starting/stopping autocycling	13-5
Starting/stopping a continuous measurement (STAT)	13-6
Using venous stasis	13-6
able of Figures	
Figure 13-1. NIBP digit field display	13-1
Figure 13-2 Cuff positioning	19_9



Overview

The non-invasive blood pressure (NIBP) measurement uses the oscillometric measuring principle. The cuff is inflated with a pressure slightly higher than the presumed systolic pressure, then slowly deflated at a speed based on the patient's heart rate, collecting data from the oscillations produced by the pulsating artery. Based on this data, the unit calculates values for systolic, mean and diastolic pressures.

The measurement can be used for adults, children and infants. The monitor automatically recognizes the type of cuff hose (black hose for adults and children, white hose for infants) used for each patient type at the beginning of each measurement.

It is possible to set an automatic cycling mode to make measurements at desired time intervals. It is also possible to measure NIBP continuously for five minutes in STAT mode or take separate single measurements.

NIBP display

NIBP can be displayed in the digit field:




Figure 13-1. NIBP digit field display

- (1) Label
- (2) Systolic and diastolic pressure value of non-invasive blood pressure
- (3) Mean pressure value of NIBP
- (4) Time since the last autocycle measurement
- (5) NIBP autocycle time indicator

Patient connections




Figure 13-2. Cuff positioning

Selecting the cuff hose

Two different cuff hoses with different cuff connections are available:

- BLACK hose for adults and children
- WHITE hose for infants

The monitor automatically recognizes the hose and selects the appropriate inflation pressure. Another inflation limit can be selected in the NIBP Setup menu. Adult inflation limits cannot be selected for an infant cuff.

NOTE: The NIBP system incorporates a safety circuit to prevent over-pressurization or prolonged inflation of the cuff.

Selecting the cuff

To determine whether the cuff size is correct, check that the white index line on the outer edge of the cuff falls between the range lines. If not, use a larger or smaller cuff.

Cuff size	Color	Limb circumference	Hose
Child	Green	12-19 cm	Adult – Black
Small adult	Royal Blue	17-25 cm	Adult – Black
Adult	Navy Blue	23-33 cm	Adult – Black
Adult Long	Navy White	23-33 cm	Adult – Black
Long Adult	Wine	31-40 cm	Adult – Black
Large Adult Long	Wine	31-40 cm	Adult – Black
Thigh	Brown	38-50 cm	Adult – Black
Infant	Orange	8-13 cm	Infant – White
Neonatal #3	White	6-11 cm	Infant – White
Neonatal #4	White	7-13 cm	Infant – White
Neonatal #5	White	8-15 cm	Infant - White

The material of the reusable cuffs is soft, durable Nylon.

The material of the disposable cuffs is soft, durable, latex-free vinyl.







NOTE: When using infant cuffs, the white infant cuff hose must be used. Also make sure the **Inflation Limits** selection is set to **Infant** or **Auto**.

NIBP setup menu

To open the NIBP Setup menu, press the NIBP key and select NIBP Setup.





Ready Prompt

Gives an audible tone when the NIBP measurement is ready. Adjust the volume of the beep tone from 1 (silent) to 10 (loud), or to 0 (OFF).

Inflation Limits

The selections are: Infant, Child, Adult, and AUTO.

When the selection is **AUTO** (default), the monitor automatically identifies the cuff hose and selects the right inflation pressure and alarm limits. The **Child** selection decreases the maximum inflation pressure to 200 mmHg when using adult hoses/cuffs.

NOTE:

- When using infant cuffs, the white infant cuff hose must be used.
 The Child selection increases the maximum inflation pressure to 200 mmHg when using infant hoses/cuffs.
- When using very large adult cuffs, use Adult limits to prevent a "Cuff loose" message from appearing.


Calibration Check Enables a calibration check using an external manometer (see below).

NIBP calibration check

- 1. Remove the cuff from the connector before entering the NIBP Setup menu.
- 2. To enter the NIBP Setup menu, press the NIBP key and select NIBP Setup.
- 3. Attach an external manometer with pump to the NIBP connector.
- 4. Pump to about 200 mmHg.
- Compare the readings of the manometer and the display.
 If the difference is greater than 4 mmHg, calibration by authorized service personnel is recommended.

Starting NIBP measurement

To start a single NIBP measurement:



Press the NIBP Start/Cancel key.

Press the NIBP key and select Start Manual.

In the beginning of the measurement, the sys and dia labels are replaced with the inflation limit indication (adult, child, infant) for five seconds. The cuff pressure will be displayed in the mean pressure value field.

If motion artifacts are detected, the monitor automatically holds deflation until the motion stops (maximum of 30 seconds). If the artifacts prevent proper measurement, a new measurement is automatically started.

When the measurement is ready, a short beep appears and result numbers flash.

WARNING: The monitor sets the inflation pressure automatically according to the first measurement. Reset the case to reset the inflation limit before measuring a new patient.

During measurement

Observe the cuffed limb frequently. Measurement may impair blood circulation.

Intervals below 10 minutes and STAT measurements are not recommended for extended periods of time.

Make sure that the tubing is not bent, pressed or stretched. Otherwise, measurement may be impaired.

Blood pressure values may be affected by a change in the patient's position.

Automatic NIBP double-check

If the NIBP value exceeds the alarm limits, a new measurement is taken automatically (immediately, if Manual measurement is selected, and after 30 seconds, if AUTO measurement is selected). If an alarm situation persists, an alarm will be given.





To cancel any measurement, press the NIBP Start/Cancel key.


Autocycling

It is possible to set automatic NIBP measurement On and Off at selected intervals. Autocycling is synchronized to real time. For example, if the first measurement is at 12:02 with a 5-minute cycle time, the next measurement will be at 12:05 and again at 12:10, 12:15, etc.

Setting the cycle time

The possible intervals for autocycling are 1, 2.5, 3, 5, 10, 15, 30, or 60 minutes.

To set the cycle time:




- 1. Press the NIBP key.
- 2. Select Cycle Time and choose the time.



Starting/stopping autocycling

To start autocycling, press the NIBP key and select Start Cycling.



The bar at the bottom of the NIBP display shows the remaining time until the next measurement.

To stop autocycling:

Press the NIBP key and select Stop Cycling.


or

If you wish to cancel the present measurement but start a new one after the selected interval time, press the NIBP Start/Cancel key.

Starting/stopping a continuous measurement (STAT)

The STAT mode initiates continuous measurement for five minutes. A new NIBP measurement starts immediately after the previous one. After five minutes, the monitor automatically returns to the previously selected cycling interval or to manual mode.

To start a continuous measurement, press the NIBP key and select Start STAT.




In STAT mode, the earlier systolic value is measured and displayed until the final result of a new one is available.

To stop a continuous measurement, press the NIBP Start/Cancel key.

Using venous stasis


Venous Stasis initiates and holds the pressure in the cuff to help venous cannulation. A constant pressure of 40 (infant), 60 (child), or 80 mmHg (adult) is maintained in the cuff for one (infant) or two minutes (child, adult). The value is selected automatically after the inflation limits (adult, child or infant) are set.

To start Venous Stasis:



- 1. Press the NIBP key.
- 2. Select Start Ven. Stasis.

The pressurization time is displayed in the NIBP field and **STASIS** flashes during the last 15 seconds.



 To release the pressure before two minutes has expired, press the NIBP Start/Cancel key before two minutes has expired.



Chapter 14. Invasive Blood Pressure

Invasive blood pressure display Displaying a combined pressure waveform	
Invasive blood pressure setup1	4-2
Patient connections	
Invasive Pressures menu1	4-3
Invasive line setup1	4-4
Setup menu options	14-4
Labeling the invasive line	
Invasive blood pressure labels and descriptions	
Adjusting alarm sources and limits	14- 6
Starting with accurate values1	.4-6
Measuring Pulmonary Capillary Wedge Pressure (PCWP)1	.4-7
Starting PCWP measurement	
Adjusting PCWP	
Canceling PCWP measurement	14-9
Table of Figures	
Figure 14-1. Top: Separated waveforms Bottom: Waveforms combined in several fields	. 14-1
Figure 14-2. Invasive blood pressure setup	
Figure 14-3. Position of Swan-Ganz catheter in pulmonary artery and Pressure waveforms from balloon-tipped pulmonary catheter during insertion	





Overview

The Cardiocap 5 monitor with the Invasive Pressures option (N-XP) enables you to measure and monitor two invasive blood pressures at the same time.

During the invasive blood pressure measurement, the transducer converts the pressure variations into electrical signals. The electrical signals are amplified and displayed as numeric pressure values and waveforms.

WARNING: All invasive procedures involve risks to the patient. Use aseptic technique. Follow catheter manufacturer's instructions.

Invasive blood pressure display

You can display the two invasive pressure waveforms separately (one above the other) or you can combine them. Combining the waveforms increases the size of the waveform.




Figure 14-1. Top: Separated waveforms

Bottom: Waveforms combined in several fields

- (1) Invasive blood pressure waveforms with zero and reference lines
- (2) Selected pressure label
- (3) Selected pressure scale
- (4) Field for messages and alarm limit settings
- (5) Systolic, diastolic and mean pressure values of invasive blood pressures

Displaying a combined pressure waveform

To display combined pressure waveforms:

- 1. Press the ComWheel and select Monitor Setup from the Main Menu.
- 2. Select Screen Setup.
- 3. Select Waveform Fields.
- Select Combine Pressures and YES.

Invasive blood pressure setup




Figure 14-2. Invasive blood pressure setup

- (1) Heparinized fluid bag with pressure infusor
- (2) Flushing set
- (3) Transducer
- (4) Adapter cable for using disposable transducers

NOTE: Patient connections made according to the picture above using Datex-Ohmeda specified accessories are defibrillator proof.

Refer to transducer manufacturer's instructions on how to remove trapped air from the transducer.

WARNING: Mechanical shock to invasive blood pressure transducer may cause severe shifts in zero balance and calibration, and cause erroneous readings.

WARNING: Patient Safety: Ensure proper contact of the return electrode of the electrosurgery unit to your patient to avoid possible burns at sensor sites.

WARNING: Make sure that no part of the patient connections touches any electrically conductive material including earth.

WARNING: Use only defibrillator-proof transducers and cables.

Patient connections


Do the connections as follows:

- Connect the pressure transducer to the transducer cable and the cable to the monitor's red connector.
- Prepare the transducer kit according to the manufacturer's instructions. Mount the kit with the transducer zeroing port at mid-heart level.
- 3. Ensure that there is no air in the transducer dome or in the catheter line. Refer to transducer manufacturer's instructions on how to remove trapped air from the transducer.
- Connect the patient catheter to the pressure line.
- 5. Connect the patient catheter to the pressure line.
- 6. Open the dome stopcock to room air.
- 7. Zero the transducer. See Starting with accurate values later in this chapter.
- Open the dome stopcock to pressure catheter and check the quality of the waveform.

Invasive Pressures menu

NOTE: Monitors with the NMT option do not have the Invasive Pressures key.

To open the Invasive Pressures menu:



Press the Invasive Pressures key.

0

Press the ComWheel and select **Parameters**. Then select **Invasive Press**.

Wedge Pressure Used for Pulmonary Capillary Wedge Pressure Measurement (PCWP).

See Measuring Pulmonary Capillary Wedge Pressure later in this chapter.


P 1 'Art' Setup and P2 'CVP' Setup

Used to set the label, scale, and other attributes of the P1 or P2 channel. See *Invasive line setup*.

Ventilation Mode Respiration causes artifacts in invasive pressures. The artifact is smallest at the end of expiration. You can select **Spont** for spontaneous respiration or **Contri** for controlled ventilation.

Invasive line setup

To open the setup menu for an invasive pressure line:



9. Press the Invasive Pressures key.
or
Press the ComWheel and select Parameters.
Then select Invasive Press.

10. Select P1 'Art" Setup or P2 'CVP' Setup.

Setup menu options

Label The label sets the features of the pressure channel. See Labeling the

invasive line.

Scale The scale is assigned when the channel is labeled. The scale can be

adjusted from 10 mmHg to 300 mmHg in steps of 10.

Digit Format With the numeric display format, you can choose to display the

Systolic/Diastolic (S/D) numbers or the Mean pressure value in a large

font size.

Filter Frequency Filters the measured signal to remove noise and artifacts. The filter

can be adjusted from 4 to 22 Hz.

HR Source If the ECG signal is affected by too much noise for a reliable heart

rate calculation, heart rate can be calculated mechanically from the pressure (Art) or plethysmographic pulse waveform. The heart rate source is shown above the numerical heart rate display. The color

used for the heart rate is the same as the source parameter.

The **AUTO** selection prioritizes the heart rate calculation in a specified

order: ECG (the lead with highest R-wave), pressure (Art), and

plethysmographic pulse waveform.


P1'Art' Alarm Alarms can be adjusted in this menu or in the Alarms Setup menu.

See Adjusting alarm sources and limits later in this chapter.



When a pressure channel is labeled, the display scales, waveform color, filter, alarm source, and alarm limits are also changed. The label sets the other features.

To change the label:



11. Press the Invasive Pressures key.

Press the ComWheel and select **Parameters**. Then select **Invasive Press**.

12. Select **P1 'Art" Setup** or **P2** 'CVP' Setup.

13. Select **Label** and choose the label.



Invasive blood pressure labels and descriptions

The labels are as follows:

P1, P2 Standard, startup labels

Art Arterial pressure

CVP Central venous pressure

PA Pulmonary arterial pressure

RAP Right atrial pressure

RVP Right ventricular pressure

LAP Left atrial pressure

ICP Intracranial pressure

ABP Arterial Blood Pressure

Both Art and ABP labels are available for situations when two arterial lines are desired but you want to use different settings or alarm labels.

Assigning a label automatically changes the other pressure settings as shown below:

LABEL	P1, Art, ABP	P2, CVP	RAP, LAP	ICP	PA	RVP
Scale	200	20	20	20	60	60
Digit Format	Sys/Dia	Mean	Mean	CPP	Sys/Dia	Sys/Dia
Filter	22	9	9	9	9	9
Alarm source	Sys	off	off	off	off	off
Color	Red	Blue	White	White	Yellow	White



Adjusting alarm sources and limits

You can adjust or turn off pressure alarm limits.

To do this:

14. Press the Invasive Pressures key.

or

Press the ComWheel and select Parameters. Then select Invasive Press.

15. Select P1 'Art' Setup and P1 'Art' Alarm.

or

Select P2 'CVP' Setup and P2 'CVP' Alarm.

You can choose **Sys** (Systolic), **Dia** (Diastolic), or **Mean** as a source for each pressure channel. The limits are on for only one source at a time.

For more information about alarms, see the Alarms chapter.

Starting with accurate values

Pressure transducers generally produce a small signal even when no pressure is applied to them. It is necessary to zero the monitor with the transducer to establish an accurate electrical zero point.

The position of the transducer affects the accuracy of the measurement. An error of 10 mmHg of static pressure will be introduced for every 13.6 cm (5.4 inches) difference in height between the mid-heart and the transducer.

To zero the transducer:

Invasive Pr	essures
Zero ALL	
Zero P1 'Art'	
Zero P2 'CVP'	
Wedge Pressure	STATE OF THE STATE OF T
P1 'Art' Setup	
P2 'CVP' Setup	
Ventilation Mode	Contrl
Parameters Menu	
	1

16. Open the transducer to air.

17. Pressures key.

or

Press the ComWheel and select **Parameters**. Then select **Invasive Press**.

Press the Invasive

18. Zero each channel.

Zero ALL Zeros all transducers (except ICP) to atmospheric pressure.

Zero P1... Zeros P1 transducer only.

Zero P2 ... Zeros P2 transducer only.



The "Zeroing" message is displayed during the zeroing process. After the transducer has been zeroed, the "Zeroed" message appears in the digit field.

After each channel has been zeroed, the time when zeroing occurred appears in the menu.

NOTE: Check the zero level after power interruptions.





Measuring Pulmonary Capillary Wedge Pressure (PCWP)

Pulmonary capillary wedge pressure (PCWP) is the occlusion pressure of the pulmonary artery. PCWP can be measured with a pulmonary artery flotation catheter. When the balloon at the tip of the catheter is inflated, the pulmonary artery is occluded. The pressure (PCWP) indirectly reflects the filling pressure of the left ventricle.

The monitor uses an algorithm to automatically detect the pressure values at the end of expiration when the artifact caused by respiration is minimal. The mean pressure value is PCWP.




Figure 14-3. Position of Swan-Ganz catheter in pulmonary artery and Pressure waveforms from balloon-tipped pulmonary catheter during insertion

The speed of the waveforms in the Wedge Pressure menu is 12.5 mm/s.

NOTE: During wedge pressure measurement, PA values are not trended and PA alarms are disabled.


Starting PCWP measurement

NOTE: Because the PCWP measurement site is in an extremely delicate area, only specially qualified medical personnel should perform the insertion of the Swan-Ganz catheter. Follow the catheter manufacturer's instructions.

- 19. Position the Swan-Ganz catheter in the pulmonary artery. Continuous monitoring of the pressures along the route of the catheter tip will help to identify the location of the tip. Use the distal lumen for the pressure line.
- 20. PCWP measurement is dedicated to the invasive pressure channel labeled **PA**. To check the label:
 - Press the Invasive Pressures key.
 or
 Press the ComWheel and select Parameters. Then select Invasive Press.
 - Check that the wedge pressure channel is labeled **PA**. If it is not, label the channel as described in *Labeling the invasive line* earlier in this chapter.
- 21. Check that the monitor has correct ventilation status information for the patient:

In the Invasive Pressures menu, select **Ventilation Mode** and choose **Spont** (spontaneous) or **Contri** (controlled).

- 22. Select Wedge Pressure.
- 23. Select Measurement.



24. Inflate the catheter balloon when the "Inflate the balloon" message is displayed in the PA waveform field.

The monitor searches the end expiratory mean pressure value. The pressure is displayed in the waveform field and menu. After 20 seconds, the waveform is automatically frozen and the "Deflate the balloon" message is displayed.

25. Deflate the balloon when the "Deflate the balloon" message is displayed in the PA waveform field.

The pressure waveform will stay frozen until you accept the PCWP level.

26. Adjust the PCWP level by turning the ComWheel. Press the ComWheel to accept the PCWP level that represents the true PCWP level.

After accepting the PCWP level, normal pressure monitoring continues.


NOTE: You can also manually freeze the waveform before 20 seconds has elapsed by selecting Freeze/Adjust.



Adjusting PCWP

To manually adjust the PCWP pressure level:

- 27. In the Wedge menu, select Freeze/Adjust.
- 28. Turn the ComWheel to move the cursor to a point on the waveform that represents the true PCWP level.



- 29. Press the ComWheel.
- 30. Select Confirm.

Canceling PCWP measurement

If the measurement did not succeed, select Cancel.

This cancels the whole measurement. The first menu appears so you may start a new measurement.




Chapter 15. Temperature

Overview	15-1
Temperature display	15-1
Automatic temperature test	
Patient connections	15-1
Temperature menu	15-2
Changing the temperature label	
Changing the temperature units	15-2
Table of Figures	
Figure 15-1. Temperature digit display	15-1
Figure 15-2. Temperature setup	







15. TEMPERATURE

Overview

You can use Datex-Ohmeda temperature probes to measure esophageal, nasopharyngeal, rectal, and skin temperature, for example.

Monitoring perioperative body temperature is recommended when inducing hypothermia or if unexpected temperature changes occur.

WARNING: Patient Safety. To prevent patient injury, use only Datex-Ohmeda temperature probes.

Temperature display




Figure 15-1. Temperature digit display

- (1) Label
- (2) Temperature measurement value

Automatic temperature test

The temperature measurement function is tested periodically. The "Performing temp test" message appears during the test. If the test fails, an error message is displayed.

Patient connections





Figure 15-2. Temperature setup

To make the connections:

- 1. Attach the temperature probe to the patient.
- 2. Connect the temperature probe to monitor connector T1 or T2.

Temperature menu




NOTE; The **Temp Alarms** selection in the Temp Setup menu opens a menu in which you can adjust the temperature alarm limits. For instructions, see the *Alarms* chapter.

Changing the temperature label

- 1. Press the ComWheel and select Parameters from the Main Menu.
- 2. Select Temp Setup.
- 3. Select T1 Label.

or

Select T2 Label.



4. Select the label you wish to use for the temperature measurement site:

Eso	Esophageal temperature	Skin	Skin temperature
Naso	Nasopharyngeal temperature	AirW	Airway temperature
Tymp	Tympanic temperature	Room	Room temperature
Rect	Rectal temperature	Муо	Myocardial temperature
Blad	Bladder temperature	Core	Core temperature
Axil	Axillary temperature	Surf	Surface temperature

Changing the temperature units

- 5. Press the ComWheel and select Parameters from the Main Menu.
- 6. Select Temp Setup.
- 7. Select Unit and choose °C or °F.







Overview	16-1
Display of gasesAgent mixture	
Patient connections	16-2
Airway gases setup alternatives	16-3
Sample gas exhaust connections	16-5
Scavenging through the ventilator reservoir	16-6 16-6
CO ₂ setup	16-8
0 ₂ setup	16-9
N ₂ O setup	16-10
Agent setup	16-11
MAC values	16-12
Automatic agent identification	16-13
Agent mixtures	16-13
Selecting alarm sources	16-13
Gas calibration	16-14
Calibrating the monitor Calibration adjustments Autozeroing intervals	16-16
Interfering gases	16-16
Unit conversions	16-17
Other adjustable features	16-17
Changing units	
Changing parameter colors	16-17

Table of Figures

Figure	16-1. Airway gas waveform display	16-1
Figure	16-2. Lower digit field for gases	16-1
Figure	16-3. Airway gases setup	16-2
Figure	16-4. Normal endotracheal intubation	16 - 3
Figure	16-5. Tracheostomy	16-3
Figure	16-6. Mask ventilation	16-4
Figure	16-7. Endotracheal intubation of pediatric patients	16-4
Figure	16-8. Scavenging through ventilator reservoir	16-5
Figure	16-9 Connecting sample gas outlet directly to a anesthesia gas scavenging	
	system	16-6
Figure	16-10. Gas return to patient circuit in ADU	16-7
Figure	16-11. Attaching the regulator to the gas container 1	6-14
Figure	16-12. Connecting a sampling line to the gas valve and feeding gas	6-15







16. AIRWAY GASES

Overview

With the Datex-Ohmeda Cardiocap/5, you can measure and monitor the gases delivered to an anesthetized patient and exhaled by the patient through the anesthesia circuit.

Respiratory rate is the frequency of peak (end tidal) CO_2 measurements per minute. A breath is defined as a change in the CO_2 signal that exceeds 1% (8 mmHg). All concentrations are measured and displayed breath by breath.

Display of gases




Figure 16-1. Airway gas waveform display

- (1) Gas waveforms
- (2) Scale
- (3) Message field for gases
- (4) Gas label
- (5) Symbol to indicate that FiO₂ low alarm limit is set below 21%
- (6) Digit field for ET and FI gas values, and FI-ET difference
- (7) Respiration rate
- (8) MAC value, based on measured AA and N2O ET values

Gases			
FT A	02%	N20%	Enf%
E1	16	77	2.3
FI	21	77	3.0

Figure 16-2. Lower digit field for gases

Agent mixture

During an agent mixture situation, the "Agent mixture" message appears in the lower digit field for gases. The labels of both agents and their concentrations are displayed.

Patient connections

CAUTION: Keep the monitor horizontal. Tilting the monitor may cause erroneous results in the readings and damage the monitor.

CAUTION: Use only cables and accessories approved by Datex-Ohmeda. Other cables and accessories may damage the system or interfere with measurement. Single-use accessories are not designed to be reused.

Take the gas sample as close to the patient's airway as possible.

- 1. Before making patient connections, check that:
 - The airway adapter connections are tight and the adapter is operating properly.
 - The D-fend container is empty and properly attached. The container should be changed or cleaned between patients and emptied when it is more than half full.
- 2. Connect the sample gas outlet to gas scavenging if N2O or volatile agents are used.
- 3. Attach the gas sampling line to the connector on the D-fend water trap.

 NOTE: The "Sampling line blocked" message may appear if you attach the sampling line to the water trap after you turn on the monitor.
- 4. Turn on the monitor. A self-check is performed and automatic agent identification is activated.
- 5. When the "Calibrating gas sensor" message disappears, make the patient connections as described below.




Figure 16-3. Airway gases setup

- (1) Gas sampling line
- (3) Airway adapter with sampling line connector
- (2) Y-piece
- (4) Heat and moisture exchanger with filter HMEF
- 6. Connect the sampling line to the patient's airway adapter. Position the adapter's sampling port upwards to prevent condensed water from entering the sampling line.



NOTE: If you are administering medication using a nebulizer, disconnect the HMEF and airway adapter while the nebulizer is in use.

CAUTION: Remove the airway sampling line from the patient airway while nebulized medications are being delivered.

Airway gases setup alternatives




Figure 16-4. Normal endotracheal intubation

- (1) Intubation tube with 15 mm connector
- (2) Airway adapter




Figure 16-5. Tracheostomy

- (1) Tracheostomy tube with 15 mm connector
- (2) Heat and Moisture Exchanger (HME)
- (3) Airway adapter




Figure 16-6. Mask ventilation

- (1) Mask
- (2) Bacteria filter
- (3) Airway adapter




Figure 16-7. Endotracheal intubation of pediatric patients

- (1) Endotracheal tube
- (2) Pediatric airway adapter (Replaces regular pediatric endotracheal tube connector. For sizes 2.5 to 4.0 endotracheal tubes.)
- (3) Pediatric heat and moisture exchanger (HME)
- (4) Fresh gas inlet

Sample gas exhaust connections

When N_2O or volatile anesthetics are used, pollution of the operation room by these gases should be prevented. Connect the sample gas outlet of the monitor to the scavenging system or return it to the patient circuit.

Connect the sample gas outlet of the monitor to the scavenging system either

- · through the ventilator, or
- · directly to the vacuum scavenging system.

CAUTION: Strong scavenging suction may change the operating pressure of the monitor and cause inaccurate readings or internal damage.

Connect sample gas outlet only to open scavenging system where gas is removed in room pressure.

Scavenging through the ventilator reservoir

To scavenge through the ventilator reservoir:

- 1. Connect an exhaust line to the sample gas outlet on the rear panel of the monitor.
- 2. Attach the other end of the line to the ventilator reservoir. Make sure that the reservoir tube diameter is at least 2-3 times larger than the exhaust line.




Figure 16-8. Scavenging through ventilator reservoir

Scavenging through the anesthesia gas scavenging system

Anesthesia machines are equipped with an anesthesia gas scavenging system (AGSS), and in some machines you can connect the sample gas outlet directly to that.

For example, connect the sample gas outlet to the Datex-Ohmeda S/5 Avance:




Figure 16-9 Connecting sample gas outlet directly to a anesthesia gas scavenging system

Note: Refer to the anesthesia machine's documentation to find out where and how the sample gas can be connected.

Connecting directly to a vacuum scavenging system

To scavenge through a direct connection:

- 1. Connect the exhaust line (733195, 5/pkg) to the sample gas outlet on the monitor.
- Connect the exhaust line only to an open scavenging system where gas is removed at room pressure. Do not connect the monitor directly to a vacuum scavenging system.



Returning gas to the patient circuit

In some anesthesia machines, you can return the sample gas to the patient circuit, refer to the anesthesia machines manuals. For example, if you use the Datex-Ohmeda S/5 Anesthesia Delivery Unit (ADU), connect an optional adapter (881644, 5/pkg) to the patient breathing tubes.





Figure 16-10. Gas return to patient circuit in ADU

CO₂ setup

- Press the ComWheel and select Parameters from the Main Menu.
- 2. Select Airway Gas.
- Select CO2 Setup.



To change the scale, select Scale and choose the scale.

The 0 - 6% scale is used in normal situations. The 0 - 10% and 0 - 15% scales are used if hypercarbia is expected.


5. To change the source used for the respiration rate, select Resp Rate Source.

You can choose to have the rate calculated from the ECG or the CO_2 measurement. If **AUTO** is selected, the respiration rate is calculated from CO_2 , when CO_2 is measured, and from the ECG measurement otherwise.

Respiration rate calculation switches back to impedance respiration if you press the **Silence Alarms** key during an Apnea alarm.


6. Select CO2 Alarm or Resp Rate Alarm to open the corresponding alarms adjustment menu where you can change the CO_2 or respiration rate alarm limits.





To open the O_2 Setup menu:

- Press the ComWheel and select Parameters from the Main Menu.
- Select Airway Gas.
- 3. Select 02 Setup.



4. To select the scale for the waveform oxygram, select **Scale**.

The monitor measures inspiratory and expiratory O_2 continuously and displays an O_2 waveform oxygram. It also calculates and displays the difference of inspired and expired O_2 . The scale allows you to select the following ranges for oxygram: difference scales **DIFF 6**, **DIFF 10**, **DIFF 15**, **DIFF 30**, **10 - 60%**, **0 - 100%**.

The monitor determines the maximum measured O_2 concentration and automatically sets the reference level for the difference scales. DIFF 30 and 10 - 60% are used in the recovery room when the patient is oxygenated, for example. Scale 0 - 100% enables you to inspect the full range of O_2 . It is used during pre-oxygenation, for example.

5. Select Measurement and choose ON or OFF.

You can turn off O_2 measurement wen no O_2 alarms or digit information is desired.



If **OFF** is selected, this symbol is displayed in the digit field and the "Measurement OFF" message is displayed in the O_2 waveform field.

6. Select **02 Alarm** to open the O_2 Alarms Adjustment menu where you can change the O_2 alarm limits as described in the *Alarms* chapter.




This symbol is displayed beside the O_2 value when the FiO_2 low alarm limit is set below 21%.

N₂O setup

NOTE: N₂O and Agent adjustments are made in the same menu. To adjust for agent measurement, see *Agent setup* later in this chapter.

To open the Agent/N2O setup menu:

- 1. Press the ComWheel and select Parameters from the Main Menu.
- 2. Select Airway Gas.
- 3. Select Agent/N20 Setup.
- Select N20 Measurement and choose ON or OFF.



- When the measurement is **ON**, the FiN $_2$ O alarm limit is FI x 82%.
- In cases where no N_2O alarms or digit information is desired, you can turn off the N_2O measurement.




If OFF is selected, this symbol is displayed in the digit field and the "Measurement OFF" message is displayed in the N_2O waveform field.

Agent setup

NOTE: Agent and N_2O adjustments are made in the same menu. To adjust for N_2O measurement, see N_2O setup earlier in this chapter.

To open the Agent/N₂O Setup menu:

- 1. Press the ComWheel and select Parameters from the Main Menu.
- 2. Select Airway Gas.
- Select Agent/N20 Setup.



4. To select the anesthetic agent waveform scale, select **Agent Scale** and choose 0 - 1.2%, 0 - 2.5%, 0 - 5%, 0 - 10% or 0 - 20%.

Scale 0 - 1.2% is used with concentrations < 1%. Scales 0 - 2.5% and 0 - 5% are normal scales, 0 - 5% being used in induction, for example. The maximum scale is 0 - 20%.

5. Select Agent Measurement and choose ON or OFF.

You can turn off agent measurement when no agent alarms or digit information is desired.



If **OFF** is selected, this symbol is displayed in the digit field and the "Measurement OFF" message is displayed in the agent waveform field.

6. Select **Agent Alarm** to open the Agent Alarms Adjustment menu where you can change agent alarms.

MAC values

To display the MAC value on the screen:

- 1. Press the ComWheel and select Monitor Setup from the Main Menu.
- 2. Select Screen Setup.
- 3. Select Digit Fields.
- 4. Select one of the lower digit fields and choose MAC.

1 MAC (Minimum Alveolar Concentration) is the alveolar concentration (end-tidal) of the agent at which 50% of individuals fail to move in response to a noxious stimulus, such as a surgical incision.

WARNING: MAC values are empirical, not absolute. Cardiocap/5 MAC values correspond to those of healthy adults and cannot be applied to children. Age and other individual factors influencing the effect of volatile agents are not taken into account.

The monitor calculates MAC values according to the following formula:

$$MAC(AA) = \frac{\%(EtAA)}{x(AA)} + \frac{\%EtN2O}{100}$$

where x(AA) is Hal=0.75%, Enf=1.7%, Iso=1.15%, Sev=2.05%, Des=6.0%

Table 16-1. MAC numbers of different anesthetics in oxygen ¹

	HAL	ENF	ISO	SEV	DES	N ₂ 0
1 MAC	0.75%	1.7%	1.15%	2.05%	6%	100%
2 MAC	1.5%	3.4%	2.3%	4.1%	12%	
3 MAC	2.3%	5.1%	3.45%	6.2%	18%	

Table 16-2. MAC numbers of different anesthetics in 65% N_20^{-1}

	HAL	ENF	ISO	SEV	DES	N ₂ O
1 MAC	0.3	0.7	0.45	0.8	2.8-4	_

¹⁾ Quasha AL. Eger EI II, Tinker JH. Determination and application of MAC. Anesthesiology 1980; 53: 315.





The agent identification option automatically identifies and selects Halothane, Enflurane, Isoflurane, Sevoflurane and Desflurane on the basis of the identification. Automatic agent identification is activated after power up.

When the monitor identifies and selects the agent, it displays the "XXX selected" message (XXX being HAL, ENF, ISO, SEV or DES). The message is displayed for 60 seconds. The inspiratory and expiratory concentrations of the agent are shown in the digit field.

The minimum concentration for the identification is 0.15 vol%. The agent selection remains active even if the concentration decreases below 0.15 vol% during the case.

Automatic agent identification operates after the normal warm up of the monitor (approximately five minutes).

Agent mixtures

The agent identification warns you if a mixture of anesthetic agents is detected. The mixture warning is activated at the latest when the concentration of the minor agent is greater than 0.3 vol% and more than 15% of the total anesthetic agent concentration.

When a mixed agent is detected, the "Agent mixture" message is displayed and an audible alarm sounds. The message remains active as long as the situation persists.

When there is an agent mixture, the anesthetic agent concentrations and labels of two anesthetic agents are displayed.

When changing the anesthetic agent, the monitor will detect an agent mixture until the first agent is washed out of the patient and the circuit. When the second agent begins to dominate, the measurement is based on the new agent. The "Agent mixture" message disappears when the concentration of the first agent becomes insignificant. When this event occurs, the exact limits depend on the mixture of agents.

Selecting alarm sources

Alarm sources can be selected in the Alarm Sources submenu under the Alarms Setup menu. The selections are **FI** or **ET** as the high- and low- alarm limits. For more information, see the *Alarms* chapter.

Gas calibration

Perform a gas calibration every six months or whenever there are indications of errors in the gas readings. The time of the last calibration is shown at the bottom of the Gas Calibration menu.

Calibration is not available during the first five minutes of monitoring and during a gas sampling warning.

NOTE:

- During gas calibration, % units are always used for CO₂ regardless of selected measuring units.
- Anesthetic agent is always calibrated with Desflurane.

Equipment and use

- Use only Datex-Ohmeda calibration gas to calibrate the gas measurement.
 Otherwise, the calibration will not succeed.
- The calibration gas container may be used until the pressure indicator reaches the red zone.
- If separate gas containers are used, each gas must be calibrated separately.
- If you use an older brass regulator, the feeding pressure should be adjusted between 5 and 7 psi. The use of an old regulator with the new aerosol containers requires an adapter available from Datex-Ohmeda. For ordering details see the Datex-Ohmeda Supplies and Accessories Catalog.

Calibrating the monitor

To calibrate the monitor:

- Turn on the monitor. For maximum accuracy, let the monitor warm up for 30 minutes.
- 2. Attach the regulator to the gas container.





Figure 16-11. Attaching the regulator to the gas container








Figure 16-12. Connecting a sampling line to the gas valve and feeding gas

- 4. Press the ComWheel and select Parameters from the Main Menu.
- 5. Select Airway Gas.

The **Gas Calibration** selection remains gray (inactive) until the "Calibrating gas sensor" message is no longer displayed.

6. Select Gas Calibration.





7. Wait until the "Zero ok" and then "Feed gas" messages appear on the screen after each gas.

If the "Zero error" message is displayed, press the **Normal Screen** key and repeat the calibration procedure. If the problem persists, contact authorized service personnel.

8. Open the regulator and feed the calibration gas until "Adjust" appears, then close the valve.


NOTE: When the monitor is in automatic agent ID mode and it detects a calibration gas and no anesthetic agent (from Datex-Ohmeda Quick Cal calibration gas, for example), the "Cal gas found" message appears. If the identification sensor fails, the "Agent id inop" message will be displayed.

Check that the displayed gas values match the values on the calibration gas container.

Calibration adjustments

Calibration adjustments may be required if the gas values displayed during calibration do not match the values on the calibration gas container.

If adjustments are required:



- 10. Select **Gas Calibration** in the Airway Gas menu.
- 11. Turn the ComWheel to highlight the first gas to be adjusted. Press the ComWheel to open an adjustment window.
- Turn the ComWheel until the displayed value matches the desired value. Press the ComWheel to confirm the adjustment.
- 13. Repeat steps 2 and 3 for each gas.

Autozeroing intervals

Autozeroing intervals after startup are: 5 minutes, 5 minutes, 5 minutes, 15 minutes, 15 minutes, and every hour thereafter. Autozeroing may also occur after agent selection.

Interfering gases

Known effects of gases NOT MEASURED by the monitor:

Helium (He) Affects the CO_2 measurement, decreases the CO_2 readings. For example: 5% CO_2 , 30% O_2 , balance He: CO_2 reading decreases 8.5% relative.

Nitrogen (N₂) The monitor compensates for the effect of nitrogen in the gas measurement. Nitrogen is used as the balance gas in gas measurement.



Unit conversions

Relationship between gas concentration and its partial pressure:

Reading in mmHg (dry gas) =

(ambient pressure in mmHg * gas concentration in %) / 100.

Reading in mmHg (water vapor saturated gas) =

[(ambient pressure in mmHg – 47 mmHg) * gas concentration in %] / 100.

Reading in kPa (dry gas) =

(ambient pressure in mmHg * gas concentration in %) / 750.

Reading in kPa (water vapor saturated gas) =

[(ambient pressure in mmHg - 47 mmHg) * gas concentration in %] / 750.

NOTE: 47 mmHg is the partial pressure of the saturated water vapor at 37°C.

Other adjustable features

This section describes other adjustable features related to measuring airway gases. A password is required for entering the Install/Service menu where the adjustments are made. If you wish to make changes, we recommend that you contact the person responsible for the entire configuration.

Changing units

To change the CO₂ measurement unit:

- Press the ComWheel and select Monitor Setup from the Main Menu.
- 2. Select Install/Service and enter the password.
- Select Installation.
- 4. Select Units.
- 5. Select Parameters.
- 6. Select CO2 and choose the unit (%, kPa, or mmHg).

Changing parameter colors

You can choose yellow, white, green, red, or blue as the color used in the waveform and digit fields for CO_2 , O_2 , or agent gas readings. For gases, you can also select orange or violet.

The factory default colors for anesthetic agents are:

Desflurane: Blue

Halothane: Red

Sevoflurane: Yellow

Enflurane: Orange

Isoflurane: Violet

To change a color:

- 1. Press the ComWheel and select Monitor Setup from the Main Menu.
- 2. Select Install/Service and enter the password.
- Select Colors.
- 4. Select Gases and choose the color.

For your notes:

Chapter 17. Patient Spirometry

Overview	17-1
Measured parameters	17-1
Measurement principles	
Display of patient spirometry	17-2
Continuous waveforms	17-2
Lower digit field display	17-3
Split screen display	17-3
Patient connections	17-4
Selecting the sensor	17-5
Adult D-lite sensor	17-5
Pediatric Pedi-lite sensor	17-5
Spirometry setup	17-6
Spirometry loops	17-7
Spirometry selections	
Scaling of loops and waveforms	
Calibrating	17-10
Other adjustable features	17-11
Changing parameter colors	17- 11
Specifying measurement conditions	
TV measuring conditions	17-11
Table of Figures	
Figure 17-1. Patient spirometry waveform display	17-2
Figure 17-2. Digit fields for Paw, Paw + TV, Flow, and Compl. + F	
Figure 17-3. Spiro1 (left) and Spiro2 (right) split screen options	17-3
Figure 17-4. Patient spirometry setup	17-4
Figure 17-5. Connector positioning	
Figure 17-6. D-lite sensor	
Figure 17-7. Pedi-lite sensor	
Figure 17-8. Vol Scale, Paw Scale, and Flow Scale options	17-10

17. PATIENT SPIROMETRY

Overview

The Patient Spirometry option (N-XV) enables monitoring of the patient's airway pressures, volumes, lung mechanics, and the ventilator operation.

In Patient Spirometry measurement, the airway pressures are measured as close to the patient as possible (between patient circuit and intubation tube), using the D-lite and Pedi-lite sensors. The same sensors are used for gas sampling.

The D-lite and Pedi-lite sensors measure kinetic pressure by a two-sided Pitot tube. Pressure is transferred to the monitor through a spirometry tube and measured by a pressure transducer. The pressure difference across a flow restrictor, together with the gas concentration information, is used to calculate flow. The volume information is obtained by integrating the flow signal.

Measured parameters

- Expiratory and inspiratory tidal volumes (TVinsp/exp)
- Expiratory and inspiratory minute volumes during controlled mechanical ventilation and during spontaneous breathing (MVinsp/exp)
- Airway pressures:

Peak pressure (Ppeak)

Plateau pressure (Pplat, plateau)

Mean pressure (Pmean)

Real-time pressure waveform (Paw)

- End expiratory pressure (PEEP)
- Flow real-time waveform (Flow)
- Compliance (Compl)
- · Airway resistance (Raw)
- Ratio of the inspiratory and expiratory time (I:E)
- Pressure-volume loop (Paw-Vol. loop)
- Flow-volume loop (Flow-Vol. loop)

NOTE: Airway resistance and PEEP are not measured with spontaneous breaths compliance or with pressure-supported breaths.

Measurement principles

Pplat is the pressure at the reversal point of the flow, at the end of the inspiration phase, after the inspiratory pause.

Ppeak is the maximum pressure during one breath.

PEEP is the pressure in the lungs at the end of the expiration, measured at the moment when the expiratory phase changes to inspiratory flow.

Compliance (Compl) tells how big a pressure difference is needed to deliver a certain volume of gas into the patient. Compliance is calculated for each breath from the following equation:

$$Compl = \frac{TV \exp}{Pplat - PEEP}$$

Raw, the airway resistance, is calculated from an equation that describes the kinetics of the gas flow between the lungs and the flow sensor. The pressure at the sensor can be derived at any moment of the breath cycle from the following equation:

$$p(t) = Raw \cdot V'(t) + V(t) / Compl + PEEP$$

where p(t), V'(t) and V(t) are pressure, flow, and volume measured at the sensor at a certain time (t).

Display of patient spirometry

Continuous waveforms





Figure 17-1. Patient spirometry waveform display

- (1) Pressure and flow waveforms
- (2) Label
- (3) Scale
- (4) Digit field for Ppeak, Pplat, and PEEP values
- (5) Digit field for tidal or minute volume

Lower digit field display



Paw TV		
cmH20	Adı	ult
Ppeak	25	TVexp ml
PEEPtot	7	790

Flow		
TV ml	Adult	
insp	exp	
800	790	
	• • •	_

Compl	Raw
mI/cmiH20	cmH20/1/s
55	8

Figure 17-2. Digit fields for Paw, Paw + TV, Flow, and Compl. + Raw

You can choose to display tidal volume (TV) or minute volume (MV) in the Spirometry Setup menu. This selection also affects the **Spiro1** split screen display (see below).

Split screen display




Figure 17-3. Spiro1 (left) and Spiro2 (right) split screen options

- (1) Scale
- (3) Real-time loop
- (2) Label
- (4) Digit field

Spiro1, the basic spirometry view, shows Ppeak, Pplat, PEEP, Compliance, TVinsp, Tvexp, MVinsp, and MVexp.

Spiro2 shows the same parameters as the basic view plus Pmean, I:E, and Raw.

For both views, TVinsp/exp and MVinsp/exp values are displayed according to the spirometry setup selections.

To select the spirometry values for the split screen display:

- Press the ComWheel and select Monitor Setup from the Main Menu.
- Select Screen Setup.
- 3. Select Split Screen and choose from the options.

Patient connections

To make the connections:

- Connect a new spirometry tube to the flow sensor by inserting the angle connectors into the sensor connectors.
- 2. Connect the other end of the spirometry tube to the pressure connectors on the monitor.
- 3. Connect a gas sampling line to the luer connector on the other side of the flow sensor.
- 4. Connect the other end of the gas sampling line to the sampling line connector on the D-fend water trap. Confirm that the connections are secure.
- Select the type of sensor you use, D-lite (the factory default) or Pedi-lite:
 Press the ComWheel and select Parameters. Then, select Airway Gases and Spirometry
 Setup. Select Sensor Type and choose the sensor.
- 6. Complete the other patient connections as shown below.





Figure 17-4. Patient spirometry setup

- (1) Y-piece
- (2) Spirometry tube
- (3) D-lite sensor
- (4) Bacterial filter
- (5) Gas sampling line



To prevent condensed water from entering the lines, position the gas sampling line and the spirometry tube connections so that they head upward from the patient.

Figure 17-5. Connector positioning

The gas sampling line can be aligned along the groove of the spirometry tube for a more serviceable setup.



NOTE: If you are using a nebulizer to administer medication, disconnect the HME-F filter and the flow sensor while the nebulizer is in use.

NOTE: When using a Bain circuit, the inspiratory volumes are erroneously high. This can be corrected by using a filter.

Selecting the sensor

Adult and pediatric patients are measured with different sensors.

To select the correct sensor type:

- 7. Press the ComWheel and select Parameters from the Main Menu.
- 8. Select Airway Gas.
- 9. Select Spirometry Setup.
- 10. Select Sensor Type and choose Adult or Pedi.

Both D-lite and Pedi-lite sensors also work as airway adapters with a connector for airway gas monitoring.

Adult D-lite sensor

D-lite sensors are reusable (transparent yellow) or disposable (non-transparent yellow).




Figure 17-6. D-lite sensor

The D-lite sensor is for patients with a tidal volume of 150 to 2000 ml.

The filter smoothes the flow profile and prevents mucus or humidity from entering the D-lite.

The outer diameter of the D-lite at the patient end allows connection of the D-lite sensor to a mask.

Pediatric Pedi-lite sensor

Pedi-lite sensors are reusable (transparent yellow).





Figure 17-7. Pedi-lite sensor

The Pedi-lite sensor is intended for patients having a tidal volume from 15 to 300 ml.

Spirometry setup

To open the Spirometry Setup menu:

- 11. Press the ComWheel and select Parameters from the Main Menu.
- 12. Select Airway Gas.
- 13. Select Spirometry Setup.



NOTE: For detailed information about the **Scaling** selection, refer to *Scaling of loops and waveforms* later in this chapter.

- 14. Select **Sensor Type** and choose **Adult** (the factory default) or **Pedi** (pediatric) according to the flow sensor in use.
- 15. Select **TV or MV** and choose tidal volume (**TV**) or minute volume (**MV**) for display in the digit field. Your selection also affects the Spiro1 Split Screen.
- 16. Select **Paw Alarms** to open the Paw Alarms Adjustment menu. You can adjust Ppeak and PEEP alarms in the same menu.
- 17. Select MVexp Alarm to open the MVexp Alarms Adjustment menu.





Spirometry loops


Spirometry parameters are viewed in the Spirometry display.

- If CO₂ is detected and the respiration rate is below 15/minute, the loops are updated every breath.
- If no CO₂ is detected or the respiration rate is 15/minute or higher, the loops are updated every other breath.

The values are always updated breath by breath.

To open this menu:

- 18. Press the ComWheel and select Parameters from the Main Menu.
- 19. Select Airway Gas.
- 20. Select Spirometry Loops.



Spirometry selections

Print Saved Prints the saved loops.

You can also print all saved loops by pressing the ComWheel and selecting Record/Print from the Main Menu. Then select Print Loops.

Paw-Vol Loop Displays the pressure-volume loop. A 45° angle is equal to a compliance

of 30 ml/cmH₂O for adults and 10 ml/cmH₂O for pediatrics.

With pediatric measurement, the ratio of the loop axis is changed to maintain the normalized 45° loop slope also in pediatric measurement.

Flow-Vol Loop Displays the flow-volume loop.

Scaling Changes the size of the Paw-Vol and Flow-Vol loops. See Scaling of

loops and waveforms later in this chapter.

Save Loop


Saves up to six pairs of reference loops in memory. Both loops are saved at the same time. When more than six pairs are saved, the most recent one is erased from the memory and the "Saving next will override last loop" message is displayed.

The time the saved loop pair is saved appears at the upper right of the spirometry view. A frame around the time of saving indicates which loop is currently displayed.

Reference Loop

Recalls a selected reference loop from memory. One real-time loop and one reference loop are displayed simultaneously.

To recall, highlight the stamp of the loop with the ComWheel and press the ComWheel. The real-time loop is drawn with a blue line, the reference loop is drawn with a white, dotted line.



Reference Off

If you do not want the reference loop to be displayed, choose this option.

Erase Loop

Erases the selected reference loop from memory.

To erase, turn the ComWheel to highlight the number of the loop. The loop is drawn with a blue line. Press the ComWheel to erase.


To return to the menu without erasing a loop, select **Exit** instead of the number of the loop.



Scaling of loops and waveforms

If the flow or pressure waveform, or flow, volume, or pressure axis of the loop is not drawn to suit your needs you can change the scaling.

To change the scaling:



- Press the ComWheel and select Parameters from the Main Menu.
- 22. Select Airway Gas.
- 23. Select Spirometry Loops or Spirometry Setup.
- 24. Select Scaling.

Scaling

AUTO (the factory default) automatically adjusts the scaling of parameters to suit the situation. Complete loops and waveforms are always displayed.

Vol results in interdependent scales. A change to one scale affects the others so that the ratio between pressure and volume and flow and volume scales remains constant.


Indep. allows independent and separate changes to the volume, pressure, and flow scales. A change in one scale does not affect the scaling of other scales. This selection may be needed, for example, with pediatric or ARDS patients whose airway pressure is high but the tidal volume low.


Scaling Speed

Changes the speed at which loops and waveforms are rescaled during **AUTO** scaling.

Slow rescales to a higher scale after two breaths are above the selected scale. Rescaling to a lower scale occurs after four breaths are within the limits of the lower scale.

Fast rescales to a higher scale after one breath is above the selected scale. Rescaling to a lower scale occurs after two breaths are within the limits of the lower scale. For example, select **Fast** as the scaling speed for manual ventilation.








Figure 17-8. Vol Scale, Paw Scale, and Flow Scale options

Calibrating

The flow measurement should be calibrated once a year, or when there is a permanent difference between inspiratory and expiratory volume.

Calibration should be performed by trained service personnel using a special spirometry tester designed by Datex-Ohmeda. For more information, see the *Technical Reference Manual*.





This section describes other adjustable features related to patient spirometry measurement. A password is required for entering the Install/Service menu where the adjustments are made. If you wish to make changes, we recommend that you contact the person responsible for the entire configuration.

Changing parameter colors

You can choose yellow, white, green, red, or blue as the color used in the waveform and digit fields for CO_2 , O_2 , N_2O , agent gas readings, Paw, and Flow. You can also select orange or violet for gases.

To change the color:

- 25. Press the ComWheel and select Monitor Setup from the Main Menu.
- 26. Select Install/Service and enter the password.
- 27. Select Colors.
- 28. Select Gases or select More Parameters and Paw or Flow.
- 29. Choose the color.

Specifying measurement conditions

You can specify measurement conditions of pressure, temperature, and humidity for volume measurement.

- 30. Press the ComWheel and select Monitor Setup from the Main Menu.
- 31. Select Install/Service and enter the password.
- 32. Select installation.
- 33. Select Units.
- 34. Select Parameters.
- 35. Select TV based on and choose ATPD, NTPD, BTPS, or STPD.

TV measuring conditions

Condition	Temperature	Pressure	Relative Humidity
ATPD	Ambient	Ambient	0%
NTPD	20 °C (68 °F)	760 mmHg	0%
BTPS	37 °C (98.6 °F)	Ambient	100%
STPD	0 °C (32 °F)	760 mmHg	0%

For your notes:

Chapter 18. NeuroMuscular Transmission (NMT)

Overview	18-1
Displaying NMT	
Selecting NMT on a digit field	18-1
Selecting NMT on a waveform field	18-1
Patient connections	18-2
Stimulating electrodes	18-2
ElectroSensor setup	
Grounding electrode	
Recording electrodes	
MechanoSensor setup	
Cable set	
Starting measurement	18-6
Supramaximal current search	18-6
Increasing pulse width	18-7
Determining response reference level	18-7
Changing the stimulus mode	18-8
Train of Four (TOF) stimulation	
Double Burst Stimulation (DBS)	
Single Twitch (ST) stimulation	18-9
Changing the cycle time	18-10
Changing the stimulus current	18-10
Enabling recovery note	
Activation principle	
Stopping and restarting	18-11
Measuring deep relaxation with PTC	18-12
Regional block stimulation (Plexus stimulation)	18-13
Measurement limitations and artifacts	18-14
Stimulation artifact	
Movement artifact	
Electrosurgery	
Verifying reliability	

Other adjustable features	18-15
Selecting NMT trend	18-15
Table of Figures	
Figure 18-1. Ulnar nerve and the corresponding muscles	18-2
Figure 18-2. Stimulating electrodes	18-3
Figure 18-3. Grounding electrode of EMG measurement	18-3
Figure 18-4. Measurement from the adductor pollicis muscle	18-4
Figure 18-5. NMT setup with ElectroSensor	18-4
Figure 18-6. MechanoSensor attachment	
Figure 18-7. NMT setup with MechanoSensor	
Figure 18-8. Perioperative stimulus response	
Figure 18-9. Sensor cable, regional block adapter, and needle with syringe	







18. NEUROMUSCULAR TRANSMISSION (NMT)

Overview

The Cardiocap/5 monitor with the N-XNMT option delivers stimulating electrical pulses to a motor nerve and measures the muscle response to these stimulations.

Two electrodes are needed for electrical stimulation of a peripheral nerve. The resulting response can be measured with a MechanoSensor or an ElectroSensor.

- The MechanoSensor is attached between the thumb and index finger. It is easier to
 use and more suitable for routine clinical use. It measures movements with a
 piezoelectric wafer.
- The ElectroSensor uses three recording electrodes for electromyography (EMG) and is mainly intended for research purposes.

Displaying NMT

You can choose three different NMT views on the display: current numeric values, waveforms illustrating the progression of the stimulation responses during measuring, and graphical trends.

NOTE: When waveforms are selected for the display, the digit field with the NMT measurement disappears automatically.

Selecting NMT on a digit field

To select an NMT digit field:

- 1. Press the ComWheel and select Monitor Setup from the Main Menu.
- 2. Select Screen Setup.
- 3. Select Digit Fields.
- Select one of the fields and choose NMT.

Selecting NMT on a waveform field

During the NMT stimulation cycle, you can observe the progression of the stimulation responses as waveforms. The waveform display is mainly for research purposes.

To select an NMT waveform field:

- 1. Press the ComWheel and select Monitor Setup from the Main Menu.
- Select Screen Setup.
- 3. Select Waveform Fields.
- 4. Select one of the fields and choose NMT.

The first waveform is saved as a reference waveform automatically. It appears on the background of the display during other measurements. The reference waveform is a different color than the real-time NMT waveform.

Patient connections

Before connecting the electrodes, clean the application site of oil or dirt. Avoid placement over excessive body hair or lesions. Pay attention to the condition and placement of the electrodes. Good signal quality requires good skin contact with the electrodes. Only Datex-Ohmeda NMT electrodes are recommended.

Stimulating electrodes

To stimulate the motor nerve correctly on standard conditions, the stimulating electrodes (white and brown) are recommended to be placed along the ulnar nerve on a site where the nerve is as near to the skin as possible.

To locate the nerve, locate the ulnar artery which runs side by side with the ulnar nerve. The wrist area is the easiest site for locating the artery, and thus the nerve. The wrist area is a site where the nerve is near the skin. Fat or muscle does not hinder detection at this site.





Figure 18-1. Ulnar nerve and the corresponding muscles

WARNING: Electrical Shock Hazard. Do not place the NMT stimulation electrodes on the patient's chest.

WARNING: Electrical Shock Hazard. Always stop the NMT measurement before handling the stimulation electrodes.

WARNING: Patient Safety. Ensure proper contact of the return electrode of the electrosurgery unit to your patient to avoid possible burns at sensor sites.

WARNING: Patient Safety. Make sure that the lead set clips do not touch any electrically conductive material including earth.



The electrodes should not touch each other.

Figure 18-2. Stimulating electrodes

ElectroSensor setup

Grounding electrode

For measuring EMG response, the grounding electrode (black) may be placed where convenient, but placing it between the stimulation and recording electrodes at the palmar groove may reduce stimulation artifact.




Figure 18-3. Grounding electrode of EMG measurement

Recording electrodes

Place one green recording electrode on the adductor pollicis. Place a second recording electrode on the muscle's tendon or insertion site. The hand should be immobilized.




Figure 18-4. Measurement from the adductor pollicis muscle

The adductor pollicis muscle is customarily used for the measurement. When the thenar is preferred, electrode placement becomes crucial. The adductor pollicis and flexor pollicis brevis muscles lie next to each other and may be recorded on the radial surface of the palm between the first and second metacarpals. Interference due to median nerve stimulation is more likely than on the hypothenar side.

Cable set




Figure 18-5. NMT setup with ElectroSensor

- (1) NMT sensor cable
- (2) ElectroSensor

MechanoSensor setup

The MechanoSensor is attached between the thumb and index finger with a piece of tape. Make sure that the thumb can move freely.




Figure 18-6. MechanoSensor attachment

Cable set




Figure 18-7. NMT setup with MechanoSensor

- (1) NMT sensor cable
- (2) MechanoSensor

Starting measurement

NMT monitoring should be started before a muscle relaxant drug is administered. It is, however, advisable to begin after the induction of sleep to prevent voluntary muscle contraction and tension from interfering with the reference search.

If the patient is already relaxed, the reference level is not usually found, however, monitoring can proceed without a reference level. In some cases, the reference level may be found but may be erroneous.

To start a measurement:

- Press the NMT key.
- Select Start-Up.


Supramaximal current search

The monitor begins by searching for the supramaximal stimulation current, which is needed for activating all the fibers of the stimulated (recorded) muscles. When a stimulation pulse is delivered, a star (*) is displayed in the digit field for each pulse. Optionally, a beep will sound at the same time. The delivered stimulus current is measured and the star is displayed if the current is the same as adjusted.

The search for the supramaximal stimulation current begins with a 10 mA stimulus and the response is measured. The current is increased by 5 mA steps until the increase in current does not increase the response. This maximal current is automatically increased by 15%, resulting in a supramaximal current.

NOTE: When the ElectroSensor is used, a reference level for the stimulation responses is determined after the search.

For adults, the supramaximal current is usually between 35 and 55 mA. The current is displayed in the digit field during the search for supramaximal current. It is also displayed in the NMT Setup menu directly after the current selection.



Defasciculating doses of nondepolarizing neuromuscular blocking agents (precurarisation) do not normally affect the supramaximal current or the reference value.

Supramaximal current may not be found if the patient is already relaxed. If the supramaximal current is not found or if the response is too weak for searching a supramaximal current, the current is set to 70 mA.

If the monitor is unable to find the supramaximal stimulation current, the fault is usually in the electrode placement—the nerve is outside the dense current flow or both the ulnar and median nerves are stimulated. Progressively more muscle activity is detected as the increasing stimulation current activates new motor units.








Increasing pulse width

To help find the supramaximal current, you can increase the pulse width to increase the effect of the stimulation:

300 µs.



- Press the NMT key.
 Select NMT Setup.
 Select Pulse Width and
- 10. To start a new measurement, select **Start-up** in the NMT menu.


Determining response reference level

The supramaximal current search is followed by determination of a response reference level.

The monitor delivers four stimuli of supramaximal current at 1-second intervals. A mean of four stimulation responses is calculated and is determined as a 100% reference level.

- With the ElectroSensor, this reference is used to calculate the T1% (T1/Tref *100), the ratio of the first twitch's response to the determined reference level.
- With the MechanoSensor, the reference is used only for scaling the bar graph.

Each stimulation response is displayed as a bar in the digit field. If the bars are not similar, the responses are not repeatable and the reference level will not be accepted.



If the patient is paralyzed at the time of start-up, it does not make the measurement impossible or the results useless. When nondepolarizing blocking agents are used, the responses will fade and the reference value is rejected, however, values other than T1% are available.


If the block is depolarizing at start-up, T1% should not be used because the reference level is incorrect.



Changing the stimulus mode

NOTE: To ensure a correct reference value, the stimulus mode should be changed before monitoring starts.

To change the stimulus mode:




- 11. Press the NMT key.
- 12. Select Stimulus Mode.
- 13. Select the stimulus mode (**TOF**, **DBS**, or **ST**).


Train of Four (TOF) stimulation

Train of Four (TOF), the factory default setting, is recommended for most cases.


During Train of Four stimulation, four stimulation pulses are generated at 0.5-second intervals. The response is measured after each stimulus and the ratio of the fourth to the first response of the TOF sequence is calculated, resulting in TOF%.



If you use the ElectroSensor, T1% is displayed below the stimulation pulse bars, and a scale is included if the reference is successfully found. Scale markers represent 0%, 25%, 50%, 75% and 100% reference values.



When no reference is available, no T1% is displayed and the displayed bars are not scaled.



When relaxation deepens, the TOF% declines until the fourth response disappears and no TOF% is calculated.

When no TOF% is available, the degree of neuromuscular block is estimated from the number of responses, or Counts. The Count tells how many responses have been detected to the 4 stimuli. The fewer responses detected, the deeper the relaxation.

Number of responses (Count)	Neuromuscular Block	Muscle Power (% of Control)
1	95%	5%
2	90%	10%
3	85%	15%
4	75%	25%

Perioperatively, this can be seen in the trends where the bars represent the TOF%.





Figure 18-8. Perioperative stimulus response

Adequate clinical recovery from competitive block is normally reached when the TOF% is over 70.

Double Burst Stimulation (DBS)


Double Burst Stimulation (DBS) enables better visual observation of fading responses. It is practical when the MechanoSensor cannot be used with infants, for example.

DBS includes two separate bursts where each burst consists of three pulses directly after each other, at 50 Hz frequency. The response ratio of the second to the first burst is calculated, resulting in DBS%. DBS% is equivalent to TOF%.



Single Twitch (ST) stimulation


In single twitch stimulation, one pulse is generated and its response is measured. Single Twitch is practical when using depolarizing relaxants, since TOF% does not give additional information about the state of the patient.



Changing the cycle time

Check that the desired cycle time is selected.

To change the cycle time:




- 14. Press the NMT key.
- 15. Select Cycle Time.
- 16. Select the time interval.

NOTE: The TOF and DBS cycle times are as shown. The ST cycle times are 1, 10, and 20 seconds.

Changing the stimulus current

Check that the correct stimulus current strength is set. The current is either supramaximal (automatic start-up search) or manually selected between 10 and 70 mA. Smaller currents may be desirable for children.

To change the stimulus current:



- 17. Press the NMT key.
- 18. Select NMT Setup.
- 19. Select Current and choose the current.






Enabling recovery note

The Recovery Note alarms when the count reaches the set value. It indicates that the patient is responding more clearly to the stimuli and the neuromuscular block is decreasing. The note can be used, for example, to help maintain a certain relaxation level.

To make the note operative and set the limit for the alarming count number:



- 20. Press the NMT key.
- 21. Select Recovery Note.
- 22. Select the count limit for activating the note.

When the note is activated, you will hear a note sound and a beep. The "Block recovery" message is displayed.

Activation principle

Activation of the note depends on the Count number set as the recovery note and the cycle time.

When the cycle time is less than one minute, the monitor requires that the Count is below the Count limit in two successive measurements before the Recovery Note is enabled. Then, to give the note, the monitor requires that the number of Counts exceeds the limit, or is the same as the limit, twice in one minute. The note disappears when two measurements are below the limit again.

If the cycle time is one minute or more, or if the measurement is done manually, the monitor requires that at least one Count is below the chosen limit before the note is enabled. Then, to give the note, the monitor requires one Count to exceed or to be the same as the Count limit. The note disappears when one measurement is below the limit again.

Stopping and restarting


- To discontinue the NMT measurement, press the NMT key and select Stop.
- To continue with same patient, press the NMT key and select Continue.

Do not select **Start-up** if you want to preserve the current and reference values.

Measuring deep relaxation with PTC

When neuromuscular block deepens, different parameters are needed to measure the response. At first, when the response to the fourth TOF stimulation pulse disappears, or the first twitch is very weak, the TOF% is not available and only the number of detected counts can be observed.


Relaxation Meter



When stimulation pulses no longer give any stimulation response, the Count is not available either. To monitor the relaxation level, you can start tetanic stimulation and estimate the relaxation level from the Post Tetanic Count (PTC).

Tetanic stimulation is a continuous, 5-second stimulation. The stimulation makes the muscle more responsive. After the tetanic stimulation, single-twitch stimulations are generated. The number of detected responses to the stimulations is counted and expressed as PTC. The fewer responses detected, the deeper the relaxation.

For example, when the patient is relaxed with pancuronium, the TOF response will return in about 5 minutes when PTC is 8, and in 15 minutes when PTC is 5. These values are general guidelines and may differ from case to case.



If the responses do not fade away, a maximum of 20 responses are counted and >20 is displayed.

After tetanic stimulation, NMT measurements are stopped for one minute. After this the monitor continues with the previously selected cycle automatically.

Regional block stimulation (Plexus stimulation)

Regional block stimulation is used for locating a nerve.

Regional block requires its own adapter cable. When the regional block adapter cable is connected, the Regional Block menu can be selected from the NMT menu. Once the cable is connected, the digit field displays only the stimulation current value.

The monitor delivers single stimulation pulses at a selected rate until manually stopped. The responses to these stimulations are observable visually, the monitor does not measure them. You adjust the current manually.

To manually adjust the current:

23. Set up the regional block cable, adapter, and needle as shown below.





Figure 18-9. Sensor cable, regional block adapter, and needle with syringe

- 24. Connect the cable to the NMT connector on the monitor.
- 25. Press the NMT key.
- 26. Select **Cycle Time** and choose the cycle time.
- Select Regional Block to open the Regional Block menu.
- 28. Select Run.

The monitor is automatically in single twitch mode.

When the needle approaches the motor nerve, reduce the current so that the response stays small. The closer the needle is to the nerve, the lower the current that is needed to give a response.



Note that the pulse width is 40 μ s and somewhat higher currents may be needed when compared to other similar systems. However, shorter pulse width is less painful to the patient.

29. To stop the stimulation, select **Stop** in the Regional Block menu.

Measurement limitations and artifacts

Stimulation artifact

If electrodes are placed incorrectly, the wrong nerves are stimulated. Consequently, the wrong muscles are stimulated and the measuring electrodes are improperly placed to measure the stimulation response.

When multiple nerves are stimulated, the measured response may be affected by the electrical activity of other muscles.

If the stimulation electrodes are placed very close to the palm of the hand, the muscles are stimulated directly by the stimulation pulses.

If the current is too strong, the muscles may be stimulated directly by the current.

Movement artifact

All movements of the hand are measured as a response to the stimulation. Therefore, moving or touching the patient during a measurement may cause incorrect results. If possible, immobilize the hand during NMT measurements.

Electrosurgery

The measurement results may be incorrect during electrosurgery.

Verifying reliability

The reliability of the response can be estimated by looking at the bar graph or the NMT trend.

The bars of the graph should be in a smoothly descending order from left to right, and the NMT trend should indicate that the T1% has remained steady. If this is not the case, the newest response is unreliable. Relaxation level does not usually decrease greatly in one minute even with short-acting relaxants.

To verify the observations, start a new measurement manually right after completing the previous measurement. Sometimes the selected cycle time has been so long (five minutes, for example) that the relaxation level has changed considerably between the measurements.





This section describes other adjustable features related to NMT measurement. A password is required for entering the Install/Service menu where the adjustments are made. If you wish to make changes, we recommend that you contact the person responsible for the entire configuration.

Selecting NMT trend

If you want the NMT trend in a trend field or you want to change the NMT trend page number, do the following:

Graphical Trends
Page 4
Field 1 ECG1 -More-
Field 2 P1+HR NMT
Field 3 Sp+Ple Event1
Field 4 T1+T2 Event2
Field 5 NMT Event3
Field 6 OFF Event4
Previous Menu Event5
Event6
Event7
-More-
Neuromuscular transmission graphical trends: T1% Ratin% Count and PTC
1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1

- 30. Press the ComWheel and select Monitor Setup from the Main Menu.
- 31. Select Install/Service and enter the password.
- 32. Select Trends & Snapshot.

 33. Select Graphical
- Trends.

 34. Select the page you wish.
- 35. Select the field and choose **NMT**.

For your notes:



