

\*  
\* ee203 BOOTLOADER  
\*

ee203-bl-20151006-A  
-----

Dismount EE203 flash drive when the application launches.

ee203-bl-20150915-A  
-----

Increased USB CDC max packet size to improve serial performance.  
Reload volume directory after loading a new firmware image.

ee203-bl-20150621-C  
-----

Initial production release.

\*  
\* ee203 APPLICATION  
\*

ee203-app-20151008-A  
-----

requires: ee203-bl-20151006-A  
Remove NUL byte sent at the end of some strings.  
Add [V]ersion command.

ee203-app-20150915-A  
-----

requires: ee203-bl-20150915-A  
Modified calibration algorithm to saturate rather than discard out-of-range samples.  
Modified range switching algorithm to minimize glitching on LO to HI range switch.

ee203-app-20150728-A  
-----

requires: ee203-bl-20150621-C  
Improved calibration algorithm to make calibration constants more stable.  
Suppressed task overflow check on startup.  
Fixed text error on 'S' command.

ee203-app-20150621-A  
-----

requires: ee203-bl-20150621-C  
Initial production release.