

# **CLIA CATEGORIZATION: WAIVED** URINE SCREENING TEST RESULTS IN 5 MINUTES

The Alere iCup® Rx Drugs of Abuse Cup tests for the possible use of Amphetamine (AMP), Barbiturates (BAR), Benzodiazepines (BZO), Buprenorphine (BUP), Cocaine (COC), Marijuana (THC), Methadone (MTD), Methamphetamine (MET), Methylenedioxymethamphetamine - ecstasy (MDMA), Morphine 300 (MOP/OPI300), Morphine 2000 (OPI), Oxycodone (OXY), Phencyclidine (PCP), and Tricyclic Antidepressants (TCA).

The Alere iCup® Rx test cup is a rapid, screening test for the qualitative detection of multiple drugs in human urine at specified cutoff levels.

For Healthcare professionals including professionals at point of care sites.

For in vitro diagnostic use only.

# WHAT IS THE ALERE ICUP® RX DRUGS OF ABUSE CUP?

The Alere iCup® Rx Drugs of Abuse Cup is an immuno-chromatographic assay for the qualitative determination of the presence of drugs listed in the table below. The test you purchased may test for any combination of drugs

## WHAT IS THE CUTOFF VALUE AND APPROXIMATE DETECTION TIME?

Drug(Identifier)	Cutoff level	Minimum detection time	Maximum detection time
Marijuana (THC)	50 ng/mL	2 hours	Up to 5+ days
Cocaine (COC)	300 ng/mL	1-4 hours	2-4 days
Amphetamine (AMP)	1000 ng/mL	2-7 hours	2-4 days
Methamphetamine (MET)	1000 ng/mL	2-7 hours	2-4 days
Morphine (MOP/OPI300)	300 ng/mL	2 hours	2-3 days
Morphine 2000 (OPI)	2000 ng/mL	2 hours	2-3 days
Methylenedioxymethamphetamine (MDMA)	500 ng/mL	2-7 hours	2-4 days
Barbiturates (BAR)	300 ng/mL	2-4 hours	1-3 weeks
Benzodiazepines (BZO)	300 ng/mL	2-7 hours	1-4 days
Methadone (MTD)	300 ng/mL	3-8 hours	1-3 days
Oxycodone (OXY)	100 ng/mL	1-3 hours	1-2 days
Phencyclidine (PCP)	25 ng/mL	4-6 hours	7-14days
Tricyclic Antidepressants (TCA)	1000 ng/mL	8-12hours	2-7 days
Buprenorphine(BUP)	10 ng/mL	4 hours	1-3 days

# WARNINGS AND PRECAUTIONS

- . This kit is for external use only.
- . Do not use if pouch is punctured or not sealed.
- . Discard after first use. The test cannot be used more than once.
- Do not use test kit beyond expiration date.
- . Keep out of the reach of children.
- . Do not read results after 5 minutes

# MATERIAL PROVIDED

- 25 Test devices, one test cup per pouch. Each pouch contains a test cup with integrated test card.
- · Desiccant Pouch Remove desiccant from cup. The desiccant is for storage purposes only.
- · One (1) Instruction Sheet.
- · One (1) Adulteration color comparison chart for interpretation of adulteration test results (if equipped).
- · 25 Security seals (if provided).

### STORAGE AND STABILITY

- Store at 4-30 °C (39-86 °F) in the sealed pouch up to the expiration date.
- . Keep away from direct sunlight, moisture and heat.
- DO NOT FREEZE.

# PRINCIPLE

The Alere iCup® Rx Drugs of Abuse Cup is a competitive immunoassay that is used to screen for the presence of various drugs in urine. It is chromatographic absorbent device in which drugs in a urine sample competitively combined to a limited number of drug monoclonal antibody (mouse) conjugate binding sites.

When the test is activated, the urine is absorbed into each test strip by capillary action, mixes with the respective drug monoclonal antibody conjugate, and flows across a pre-coated membrane. When the drug levels within the urine sample are below the detection level of the test, respective drug monoclonal antibody conjugate binds to the respective drug-protein conjugate immobilized in the Test Region (T) of the test strip. This produces a colored Test line in the Test Region (T) of the strip that regardless of its intensity indicates a negative test result.

When the sample drug levels are at or above the detection level of the test, the free drug in the sample binds to the respective drug monoclonal antibody conjugate, preventing the respective drug monoclonal antibody

conjugate from binding to the respective drug-protein conjugate immobilized in the Test Region (T) of the device. This prevents the development of a distinct colored band in the test region, indicating a preliminary positive

To serve as a procedure control, a colored line will appear at the Control Region (C), of each strip, if the test has been performed properly.

## SPECIMEN COLLECTION AND PREPARATION

## WHEN TO COLLECT URINE FOR THE TEST?

Collect urine sample after minimum detection time following suspected drug use. Urine collection time is very important in detecting any drug of abuse. Each drug is cleared by the body and is detected in the urine at different times and rates. Please refer to the section "WHAT IS THE CUTOFF VALUE AND APPROXIMATE DETECTION TIMES?" contained within this package insert, to determine the minimum/maximum detection times, and cutoff level for each drug.

## HOW TO COLLECT URINE?

- 1. Remove a test cup from the foil pouch by tearing at the notch. Use it as soon as possible. Instruct the donor to remove the cap from the test cup, and void directly into the test cup. Instruct the donor to fill the cup to the 30 mL mark. It is acceptable to collect an extra sample.
- 2. Immediately read the temperature to verify that the urine temperature is within the acceptable range 32-38°C

### HOW TO PERFORM THE TEST?

Test must be performed at room temperature, 18-30°C (65-86°F).

- 1. After the urine has been collected, tighten lid, and place the test cup on a flat surface.
- 2. Read the temperature immediately to verify that the urine temperature is within the acceptable range 32-38°C (90-100°F)
- 3. Remove cup label, and verify that adulterant pad colors are within acceptable range according to adulteration guide. (If your cup is equipped with specimen validity/adulterant pads.)
- 4. Remove the label and read the results. Wait 5 minutes to determine a positive result.
- 5. Do not read results after 5 minutes. Results after more than 5 minutes may be not accurate and should not be



# READING THE RESULTS


# Preliminary positive (+)

A rose-pink band is visible in each control region. If no color band appears in the appropriate Test Region (T), a preliminary positive result is indicated for the corresponding drug of that specific test zone.

If a rose-pink band is visible in each control region and the appropriate Test Region (T), it indicates that the concentration of the corresponding drug of that specific test zone is absent or below the detection limit of the test.

If a color band is not visible in the Control Region (C) or a color band is only visible in the Test Region (T), the test is invalid. Another test should opened and run to re-evaluate the specimen. If test still provides an invalid result, please contact the distributor from whom you purchased the product. When calling, be sure to provide the lot number of the test.

Note: There is no meaning attributed to line color intensity or width. Any visible line is considered to be a line.



Certain lines may appear lighter or thinner than other lines. ANY COLORED LINE VISIBLE IN THE TEST REGION (T), NO MATTER HOW DARK OR FAINT, SHOULD BE INTERPRETED AS A **NEGATIVE** RESULT. IMPORTANT: This assay provides only a preliminary analytical test result. A more specific alternative chemical method must be used in order to obtain a confirmed analytical result. Gas chromatography-mass spectrometry (GC-MS) and liquid chromatography-tandem mass spectrometry (LC-MS/MS) are the preferred confirmatory methods. Clinical consideration and professional judgment should be applied to any drug test result, particularly when preliminary positive results are indicated.

#### What Is A False Positive Test?

The definition of a false positive test would be an instance where a substance is identified incorrectly by the Alere iCup® Rx Drugs of Abuse Cup. The most common causes of a false positive test are cross reactants. Certain foods, medicines, diet plan drugs and nutritional supplements may cause a false positive test result with this product.

# What Is A False Negative Test?

The definition of a false negative test is that the initial drug is present but isn't detected by Alere iCup® Rx Drugs of Abuse Cup. Diluted or adulterated urine specimens may cause a false negative result.

#### TEST LIMITATIONS

- 1. This test has been developed for testing urine samples only. No other fluids have been evaluated. DO NOT use this device to test substances other than urine.
- 2. Adulterated urine samples may produce erroneous results. Strong oxidizing agents such as bleach (hypochlorite) can oxidize drug analytes. If a sample is suspected of being adulterated, obtain a new sample in a different, unused cup.
- 3. This test is a qualitative screening assay. It is not designed to determine the quantitative concentration of drugs or the level of intoxication.

# ASSISTANCE

If you have any question regarding to the use of this product, please contact 1-800-340-4029.

Amphetamine/Methamphetamine (AMP/MET) and their metabolites are potent central nervous system stimulants. Acute doses induce euphoria, alertness, and sense of increased energy and power. Responses from chronic use can include anxiety, paranoia, psychotic behavior, and cardiac dysrhythmias. Methamphetamine and amphetamine are excreted in the urine as unchanged drug along with deaminated or hydroxylated derivatives. Methamphetamine also metabolize to amphetamine in the body. As a result, urine specimens from most methamphetamine users contain both the unchanged parent drug and the amphetamine metabolite.

Barbiturates (BAR) are classified as central nervous system depressants. These products produce a state of intoxication that is similar to alcohol intoxication. Symptoms include slurred speech, loss of motor coordination and impaired judgment. Depending on the dose, frequency, and duration of use, tolerance, physical dependence and psychological dependence can occur. Barbiturates are taken orally or by intravenous and intramuscular injections. Members of the barbiturate drug class typically excrete in urine as parent compound and metabolites.

Benzodiazepines (BZO) are central nervous system (CNS) depressants commonly prescribed for the short-term treatment of anxiety and insomnia. In general, benzodiazepines act as hypnotics in high doses, as anxiolytics in moderate doses, and as sedatives in low doses. The use of benzodiazepines can result in drowsiness and confusion. Psychological and physical dependence on benzodiazepines can develop if high doses of the drug are given over a prolonged period. Benzodiazepines are taken orally or by intramuscular or intravenous injection, and are extensively oxidized in the liver to metabolites. Most benzodiazepines are excreted in the urine as conjugates and metabolites.

Buprenorphine (BUP) is a potent analgesic often used in the treatment of opioid addiction. The drug is sold under the trade names Subutex™, Buprenex™, Temgesic™ and Suboxone™; all of which contain Buprenorphine HCl alone or in combination with Naloxone HCl. Therapeutically, Buprenorphine is used as a substitution treatment for opioid addicts. A substitution treatment is a form of medical care offered to opiate addicts (primarily heroin addicts) based on a similar or identical substance to the drug normally used. In substitution therapy. Buprenorphine is as effective as Methadone but demonstrates a lower level of physical dependence. The plasma half-life of Ruprenorphine is 2-4 hours. While complete elimination of a single-dose of the drug can take as long as 6 days, the detection window for the parent drug in urine is thought to be approximately 3 days.

Cocaine (COC) is a potent central nervous system stimulant and a local anesthetic found in the leaves of the coca plant. The psychological effects induced by using cocaine are euphoria, confidence and sense of increased energy. These psychological effects are accompanied by increased heart rate, dilation of the pupils, fever, tremors and sweating. Cocaine is excreted in the urine primarily as benzoylecgonine in a short period of time. Benzoylecgonine has a biological half-life of 5 to 8 hours, which is much longer than that of cocaine (0.5 to 1.5 hour), and can be generally detected for 24 to 60 hours after cocaine use or exposure.

Marijuana (THC) is generally accepted to be the principle active component in marijuana. When ingested or smoked, it produces euphoric effects. Abusers exhibit central nervous system effects, altered mood and sensory perceptions, loss of coordination, impaired short term memory, anxiety, paranoia, depression, confusion, hallucinations and increased heart rate. The primary metabolite of marijuana excreted in the urine is 11-nor-\*-9-tetrahydrocannabinol-9-carboxylic acid. The elimination of THC and metabolites in urine is highly dependent on frequency of drug use and physiology of the user.

Methadone (MTD) is a synthetic analgesic drug originally used for the treatment of narcotic addiction and pain management. The psychological effects induced by using methadone are analgesia, sedation, and respiratory depression. Overdose of methadone may cause coma or even death. Methadone is taken orally or intravenously and is metabolized in the liver and has a biological half-life of 15-60 hours.

Methylenedioxymethamphetamine (MDMA) is classified as both a stimulant and a hallucinogen. Like methamphetamine, adverse effects of 3,4-methylenedioxymethamphetamine use include jaw clenching, teeth grinding, dilated pupils, perspiring, anxiety, blurred vision, vomiting, and increased blood pressure and heart

rate. Overdose of 3,4-methylenedioxymethamphetamine may cause heart failure or extreme heat stroke. 3.4-methylenedioxymethamphetamine is taken orally in tablets or capsules and excreted in urine as parent compound and metabolites including methylenedioxyamphetamine (MDA).

Opiates (OPI/MOP/OPI300) such as heroin, morphine, and codeine, are central nervous system (CNS) depressants. The use of opiates at high doses produces euphoria and release from anxiety. Physical dependence is apparent in users and leads to depressed coordination, disrupted decision making, decreased respiration, hypothermia and coma. Heroin is quickly metabolized to 6-acetylmorphine (6-AM), morphine, and morphine glucuronide. Codeine also partially metabolizes to morphine and morphine glucuronide. Thus, the presence of morphine or morphine glucuronide in the urine can indicate heroin, morphine, and/or codeine use.

Oxycodone (OXY) is a semi-synthetic opioid with a structural similarity to codeine. It produces potent euphoria, analgesic and sedative effects, and has a dependence liability similar to morphine. Oxycodone is most often administered orally and is metabolized by demethylation to noroxycodone and oxymorphone followed by glucuronidation. The window of detection for oxycodone in urine is expected to be similar to that of other opioids such as morphine.

Phencyclidine (PCP), commonly known as "angel dust" and "crystal cyclone", is an arylcyclohexylamine that is originally used as an anesthetic agent and a veterinary tranquilizer. The drug is abused by oral or nasal ingestion, smoking, or intravenous injection. It produces hallucinations, lethargy, disorientation, loss of coordination, trance-like ecstatic states, a sense of euphoria and visual distortions. It is well absorbed following all routes of administration. Unchanged PCP is excreted in urine in moderate amounts (10% of the dose).

Tricyclic Antidepressants (TCA) have been prescribed for depression and compulsive disorders. Because of the possibility of causing serious cardiac complications, TCAs can be lethal if misused at high doses. TCAs are taken orally or sometimes by injection. TCAs and their metabolites are excreted in urine (mostly in the form of metabolites) for up to ten days.

# QUALITY CONTROL

Users should follow the appropriate federal state, and local guidelines concerning the frequency of assaying external quality control materials. Though there is an internal procedural control line in the test device of Control region, the use of external controls is strongly recommended as good laboratory testing practice to confirm the test procedure and to verify proper test performance. Positive and negative control should give the expected results. When testing the positive and negative control, the same assay procedure should be adopted.

### PERFORMANCE CHARACTERISTICS

# Accuracy

1120 (eighty of each drug) clinical urine specimens were analyzed by GC/MS and by each corresponding Alere iCup® Rx Drugs of Abuse Cup. Each test was read by three viewers. Samples were divided by concentration into five categories: drug-free, less than half the cutoff, near cutoff negative, near cutoff positive, and high positive. Results were as follows:

					Near Cutoff Negative	Near Cutoff Positive	High Positive	
Drug	Alere iCup®		Drua	Less than half the	(Between 50% below	(Between the cutoff	(greater than 50%	%Agreement with
Test	Rx	Rx Result		cutoff concentration	the cutoff and the cutoff		above the cutoff	GC/MS
	Result			by GC/MS analysis	concentration)	cutoff concentration)	concentration)	(95%CI)
AMP		+	0	0	2	11	29	100% (84.5% - 100%)
	Viewer A	-	10	18	10	0	0	95% (79.5% - 100%)
	Viewer B	+	0	0	2	11	29	100% (84.5% - 100%)
	viewei b	-	10	18	10	0	0	95% (79.5% - 100%)
	Viewer C	+	0	0	1	11	29	100% (84.5% - 100%)
	viewei C	-	10	18	11	0	0	97.5% (82% - 100%)
BAR	Viewer A	+	0	0	2	20	20	100% (84.5% - 100%)
	viewei A	-	10	10	18	0	0	95% (79.5% - 100%)
	Viewer B	+	0	0	2	20	20	100% (84.5% - 100%)
	viewer B	-	10	10	18	0	0	95% (79.5% - 100%)
	Viewer C	+	0	0	1	20	20	100% (84.5% - 100%)
	viewei C	-	10	10	19	0	0	97.5% (82% - 100%)
BZO		+	0	0	1	20	20	100% (84.5% - 100%)
	Viewer A	-	10	10	19	0	0	97.5% (82% - 100%)
	Viewer B	+	0	0	1	20	20	100% (84.5% - 100%)
		-	10	10	19	0	0	97.5% (82% - 100%)
		+	0	0	2	20	20	100% (84.5% - 100%)
	Viewer C	-	10	10	18	0	0	95% (79.5% - 100%)
coc	Viewer A	+	0	0	1	11	29	100% (84.5% - 100%)
		-	10	10	19	0	0	97.5% (82% - 100%)
	Viewer B	+	0	0	2	11	29	100% (84.5% - 100%)
	Viewer B	-	10	10	18	0	0	95% (79.5% - 100%)
		+	0	0	2	11	29	100% (84.5% - 100%)
	Viewer C	-	10	10	18	0	0	95% (79.5% - 100%)
THC	Viewer A	+	0	0	2	18	22	100% (84.5% - 100%)
		-	10	12	16	0	0	95% (79.5% - 100%)
	\	+	0	0	1	18	22	100% (84.5% - 100%)
	Viewer B	-	10	12	17	0	0	97.5% (82% - 100%)
		+	0	0	1	18	22	100% (84.5% - 100%)
	Viewer C	-	10	12	17	0	0	97.5% (82% - 100%)
MET	Viewer A	+	0	0	1	20	20	100% (84.5% - 100%)
		-	10	16	13	0	0	97.5% (82% - 100%)
	Viewer B	+	0	0	2	20	20	100% (184.5% - 100%)
	viewer B	-	10	16	12	0	0	95% (79.5% - 100%)
		+	0	0	1	20	20	100% (84.5% - 100%)
	Viewer C	-	10	16	13	0	0	97.5% (82% - 100%)

MDMA		Ι.	0	0	2	20	20	100% (84.5% - 100%)
IDIVIA	Viewer A	+	10	10	18	0	0	
		-		0				95% (79.5% - 100%)
	Viewer B	+	0		2	20	20	100% (84.5% - 100%)
		-	10	10	18	0	0	95% (79.5% - 100%)
	Viewer C	+	0	0	1	20	20	100% (84.5% - 100%)
		-	10	10	19	0	0	97.5% (82% - 100%)
IOP/	Viewer A	+	0	0	1	20	20	100% 84.5% - 100%)
PI300		-	10	19	10	0	0	97.5% (82% - 100%)
	Viewer B	+	0	0	2	20	20	100% (84.5% - 100%)
	V.0.00. D	-	10	19	9	0	0	95% (79.5% - 100%)
	Viewer C	+	0	0	1	20	20	100% (84.5% - 100%)
	1101101		10	19	10	0	0	97.5% (82% - 100%)
ITD	Viewer A	+	0	0	2	19	21	100% (84.5% - 100%)
	VICWCI A	-	10	12	16	0	0	95% (79.5% - 100%)
	Viewer B	+	0	0	1	19	21	100% (84.5% - 100%)
	viewei b	-	10	12	17	0	0	97.5% (82% - 100%)
	Viewer C	+	0	0	2	19	21	100% (84.5% - 100%)
	viewei C	-	10	12	16	0	0	95% (79.5% - 100%)
PI	Viewer A	+	0	0	1	18	22	100% (84.5% - 100%)
			10	20	9	0	0	97.5% (82% - 100%)
	Viewer B	+	0	0	1	18	22	100% (84.5% - 100%)
			10	20	9	0	0	97.5% (82% - 100%)
	Viewer C	+	0	0	1	18	22	100% (84.5% - 100%)
		١.	10	20	9	0	0	97.5% (82% - 100%)
CP		+	0	0	2	18	22	100% (84.5% - 100%)
	Viewer A		10	13	15	0	0	95% (79.5% - 100%)
		+	0	0	2	18	22	100% (84.5% - 100%)
	Viewer B	Ė	10	13	15	0	0	95% (79.5% - 100%)
		+	0	0	2	18	22	100% (84.5% - 100%)
	Viewer C		10	13	15	0	0	95% (79.5% - 100%)
CA		+	0	0	1	10	30	100% (84.5% - 100%)
•	Viewer A	Ė	10	19	10	0	0	97.5% (82% - 100%)
		+	0	0	1	10	30	100% (84.5% - 100%)
	Viewer B	Ė	10	19	10	0	0	97.5% (82% - 100%)
		+	0	0	1	10	30	100% (84.5% - 100%)
	Viewer C	-	10	19	10	0	0	97.5% (82% - 100%)
XY		-	0	0	2	19	21	100% (84.5% - 100%)
^1	Viewer A	-	10	20	8	0	0	95% (79.5% - 100%)
		÷	0	0	2	19	21	100% (84.5% - 100%)
	Viewer B	-	10	20	8	0	0	95% (79.5% - 100%)
		-	0	0	0 1	19	21	100% (84.5% - 100%)
	Viewer C	+	10	20	9	0	0	97.5% (82% - 100%)
	-	-	0	0	1	16	24	
UP	Viewer A	+	_			16	0	100% (84.5% - 100%)
		-	10	18	11			97.5% (82% - 100%)
	Viewer B	+	0	0	1	16	24	100% (84.5% - 100%)
		-	10	18	11	0	0	97.5% (82% - 100%)
	Viewer C	+	0	0	1	16	24	100% (84.5% - 100%)
	VICWCI C		10	18	11	0	0	97.5% (82% - 100%)

# Precision and Sensitivity

To investigate the precision and sensitivity, each drug sample was analyzed at the following concentrations: cutoff - 100%, cutoff -75%, cutoff -50%, cutoff -25%, cutoff, cutoff +25%, cutoff +50%, cutoff +75% and the cutoff +100%. All concentrations were confirmed with GC/MS. The study was performed 2 runs /day and lasted 25 days using three different lots of the corresponding Alere iCup® Rx Drugs of Abuse Cup. In total, 3 operators participated in the study of the corresponding Alere iCup® Rx Drugs of Abuse Cup. Each of the 3 operators tested 2 aliquots at each concentration for each lot per day (2 runs /day), for a total of 50 determinations per concentration per lot of the corresponding Alere iCup® Rx Drugs of abuse Cup.

Drug test	Approximate concentration	Number of	Result (Negative/Positive)		
	of sample (ng/mL)	determinations per lot	Lot 1	Lot 2	Lot 3
AMP	0	50	50/0	50/0	50/0
	250	50	50/0	50/0	50/0
	500	50	50/0	50/0	50/0
	750	50	50/0	50/0	50/0
	1000	50	5/45	5/45	4/46
	1250	50	0/50	0/50	0/50
	1500	50	0/50	0/50	0/50
	1750	50	0/50	0/50	0/50
	2000	50	0/50	0/50	0/50
BAR	0	50	50/0	50/0	50/0
	75	50	50/0	50/0	50/0
	150	50	50/0	50/0	50/0
	225	50	50/0	50/0	50/0
	300	50	7/43	5/45	5/45
	375	50	0/50	0/50	0/50
	450	50	0/50	0/50	0/50
	525	50	0/50	0/50	0/50
	600	50	0/50	0/50	0/50
BZO	0	50	50/0	50/0	50/0
	75	50	50/0	50/0	50/0
	150	50	50/0	50/0	50/0
	225	50	50/0	50/0	50/0
	300	50	7/43	6/44	5/45
	375	50	0/50	0/50	0/50
	450	50	0/50	0/50	0/50
	525	50	0/50	0/50	0/50
	600	50	0/50	0/50	0/50

coc	0	50	0/50	0/50	0/50
	75	50	50/0	50/0	50/0
	150	50	50/0	50/0	50/0
	225				
		50	50/0	50/0	50/0
	300	50	5/45	5/45	5/45
	375	50	0/50	0/50	0/50
	450	50	0/50	0/50	0/50
	525	50	0/50	0/50	0/50
	600	50	0/50	0/50	0/50
ТНС	0	50	50/0	50/0	50/0
-					
	12.5	50	50/0	50/0	50/0
	25.0	50	50/0	50/0	50/0
	37.5	50	50/0	50/0	50/0
	50.0	50	5/45	6/44	5/45
	62.5	50	0/50	0/50	0/50
	75.0	50	0/50	0/50	0/50
<del></del>					
	87.5	50	0/50	0/50	0/50
	100.0	50	0/50	0/50	0/50
MET	0	50	50/0	50/0	50/0
	250	50	50/0	50/0	50/0
	500	50	50/0	50/0	50/0
	750	50	50/0	50/0	50/0
	1000	50	4/46	5/45	5/45
	1250	50	0/50	0/50	0/50
	1500	50	0/50	0/50	0/50
	1750	50	0/50	0/50	0/50
	2000	50	0/50	0/50	0/50
MDMA	0	50	50/0	50/0	50/0
	125	50	50/0	50/0	50/0
<del></del>					
<u> </u>	250	50	50/0	50/0	50/0
<u> </u>	375	50	50/0	50/0	50/0
L	500	50	6/44	5/45	6/44
	625	50	0/50	0/50	0/50
	750	50	0/50	0/50	0/50
<del>-</del>	875	50	0/50	0/50	0/50
-					
	1000	50	0/50	0/50	0/50
MOP/OPI300	0	50	50/0	50/0	50/0
	75	50	50/0	50/0	50/0
	150	50	50/0	50/0	50/0
	225	50	50/0	50/0	50/0
	300	50	5/45	6/44	5/45
	375	50	0/50	0/50	0/50
	450	50	0/50	0/50	0/50
	525	50	0/50	0/50	0/50
	600	50	0/50	0/50	0/50
MTD	0	50	50/0	50/0	50/0
	75	50	50/0	50/0	50/0
	150	50	50/0	50/0	50/0
	225	50	50/0	50/0	50/0
	300	50	6/44	4/46	5/45
	375	50	0/50	0/50	0/50
	450	50	0/50	0/50	0/50
	525	50	0/50	0/50	0/50
	600	50	0/50	0/50	0/50
OPI	0	50	50/0	50/0	50/0
UPI					
	500	50	50/0	50/0	50/0
	1000	50	50/0	50/0	50/0
	1500	50	50/0	50/0	50/0
	2000	50	5/45	5/45	6/44
	2500	50	0/50	0/50	0/50
	3000	50	0/50	0/50	0/50
	3500	50	0/50	0/50	0/50
	4000	50	0/50	0/50	0/50
PCP	0	50	50/0	50/0	50/0
L	6.25	50	50/0	50/0	50/0
	12.5	50	50/0	50/0	50/0
	18.75	50	50/0	50/0	50/0
	25	50	5/45	4/46	5/45
<u> </u>					
<del>-</del>	31.25	50	0/50	0/50	0/50
<u> </u>	37.5	50	0/50	0/50	0/50
<u> </u>	43.75	50	0/50	0/50	0/50
	50	50	0/50	0/50	0/50
				50/0	50/0
TCA	0	50	50/0		
TCA	0 250	50 50	50/0	50/0	50/0
TCA	0	50			50/0
TCA	0 250	50 50	50/0	50/0	
TCA	0 250 500	50 50 50 50	50/0 50/0 50/0	50/0 50/0 50/0	50/0 50/0
TCA	0 250 500 750 1000	50 50 50 50 50	50/0 50/0 50/0 5/45	50/0 50/0 50/0 6/44	50/0 50/0 5/45
TCA	0 250 500 750 1000 1250	50 50 50 50 50 50	50/0 50/0 50/0 50/0 5/45 0/50	50/0 50/0 50/0 6/44 0/50	50/0 50/0 5/45 0/50
ICA	0 250 500 750 1000 1250 1500	50 50 50 50 50 50 50	50/0 50/0 50/0 50/0 5/45 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50	50/0 50/0 5/45 0/50 0/50
TCA	0 250 500 750 1000 1250	50 50 50 50 50 50	50/0 50/0 50/0 5/45 0/50 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50	50/0 50/0 5/45 0/50 0/50 0/50
	0 250 500 750 1000 1250 1500 1750 2000	50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 5/45 0/50 0/50 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50	50/0 50/0 5/45 0/50 0/50 0/50 0/50
	0 250 500 750 1000 1250 1500	50 50 50 50 50 50 50	50/0 50/0 50/0 5/45 0/50 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50	50/0 50/0 5/45 0/50 0/50 0/50
	0 250 500 750 1000 1250 1500 1750 2000	50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 5/45 0/50 0/50 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 50/0	50/0 50/0 5/45 0/50 0/50 0/50 0/50 50/0
	0 250 500 750 1000 1250 1500 1750 2000 0	50 50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 5/45 0/50 0/50 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 50/0 50/0	50/0 50/0 5/45 0/50 0/50 0/50 0/50 50/0 50/
	0 250 500 750 1000 1250 1500 1750 2000 0 25	50 50 50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 5/45 0/50 0/50 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 0/50 50/0 50/0	50/0 50/0 5/45 0/50 0/50 0/50 0/50 50/0 50/
	0 250 500 750 1000 1250 1500 1750 2000 0 25 50	50 50 50 50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 5/45 0/50 0/50 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 50/0 50/0 50/0	50/0 50/0 5/45 0/50 0/50 0/50 0/50 50/0 50/0 50/0 50/0
	0 250 500 750 1000 1250 1500 1750 2000 0 25	50 50 50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 5/45 0/50 0/50 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 0/50 50/0 50/0	50/0 50/0 5/45 0/50 0/50 0/50 0/50 50/0 50/
	0 250 500 750 1000 1250 1500 1750 2000 0 25 50 75	50 50 50 50 50 50 50 50 50 50 50 50 50 5	50/0 50/0 50/0 5/45 0/50 0/50 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 50/0 50/0 50/0 5	50/0 50/0 5/45 0/50 0/50 0/50 0/50 0/50 50/0 50/0 50/0 50/0 50/0 50/0
	0 250 500 750 1000 1250 1500 1750 0 0 25 50 75	50 50 50 50 50 50 50 50 50 50 50 50 50 5	50/0 50/0 50/0 5/45 0/50 0/50 0/50 0/50 50/0 50/0 50/0 50/0 6/44 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 50/0 50/0 50/0 50/0 6/44	50/0 50/0 50/15 5/45 0/50 0/50 0/50 0/50 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0
	0 250 500 750 1000 1250 1500 1750 2000 0 25 50 75 100 125	50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 50/0 5/45 0/50 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 50/0 50/0 50/0 6/44 0/50 0/50	50/0 50/0 50/0 5/45 0/50 0/50 0/50 50/0 50/
	0 250 500 750 1000 1250 1500 1750 2000 0 25 50 75 100 125 50	50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 50/0 5/45 0/50 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 50/0 50/0 50/0 5	50/0 50/0 50/0 5/45 0/50 0/50 0/50 50/0 50/
OXY	0 250 500 750 1000 1250 1500 1750 2000 0 25 50 75 100 125	50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 50/0 5/45 0/50 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 50/0 50/0 50/0 6/44 0/50 0/50	50/0 50/0 50/0 5/45 0/50 0/50 0/50 50/0 50/
OXY	0 250 500 750 1000 1250 1250 1250 1250 1250 1250 12	50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 50/0 50/0 5/45 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 50/0 50/0 50/0 50/0 6/44 0/50	50/0 50/0 50/45 0/50 0/50 0/50 0/50 50/0
OXY	0 250 500 750 1000 1250 1500 1750 2000 0 25 50 75 100 125 100 1750 2000 0 25 50 75 100 125 100 175 100 175 100 100 100 100 100 100 100 100 100 10	50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 50/0 50/0 51/45 0/50 0/50 0/50 0/50 0/50 50/0 50/0 5	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 0/50 50/0 50/0 5	50/0 50/0 50/0 50/0 50/0 0/50 0/50 0/50
OXY	0 250 500 750 1000 1250 1500 1750 2000 0 25 50 75 100 125 150 175 200 0 2.5 50 7.5 100 125 150 150 125 150 125 150 125 150 125 150 125 150 125 150 125 150 125 150 175 200 0	50 50 50 50 50 50 50 50 50 50 50 50 50 5	50/0 50/0 50/0 50/0 50/0 5/45 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 0/50 50/0 50/0 6/44 0/50 0/50 0/50 0/50 50/0 50/0 6/45 0/50 0/50 50/0 6/44 0/50 0/50 0/50 0/50 50/0 6/44 0/50 0/50 0/50 0/50 0/50 50/0 6/44 0/50	50/0 50/0 50/0 50/0 50/45 0/50 0/50 0/50
OXY	0 250 500 750 1000 1250 1500 1750 2000 0 25 50 75 100 125 100 1750 2000 0 25 50 75 100 125 100 175 100 175 100 100 100 100 100 100 100 100 100 10	50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 50/0 50/0 51/45 0/50 0/50 0/50 0/50 0/50 50/0 50/0 5	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 0/50 50/0 50/0 5	50/0 50/0 50/0 50/0 50/0 0/50 0/50 0/50
DXY	0 250 500 750 1000 1250 1500 1750 0 0 2000 0 25 50 75 100 125 150 175 200 125 200 125 200 200 25 50 75 100 200 200 25 50 200 200 25 50 200 200 2	50 50 50 50 50 50 50 50 50 50 50 50 50 5	50/0 50/0 50/0 50/0 50/0 5/45 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 0/50 50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 50/0 50/0 6/45 0/50 50/0 6/44 0/50 0/50 0/50 0/50 50/0 6/44 0/50 0/50 0/50 0/50 50/0 6/44 0/50 0/50 0/50 0/50 0/50 50/0 6/44 0/50	50/0 50/0 50/45 0/50 0/50 0/50 0/50 50/0 50/
OXY	0 250 500 750 1000 1250 1500 1750 2000 0 0 25 50 75 100 125 150 175 200 0 225 50 75 50 75 50 75 50 77 50 50 77 50	50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 50/0 50/0 50/0 5/45 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 0/50 50/0 50/0 5	50/0 50/0 50/0 50/5 0/50 0/50 0/50 0/50
OXY	0 250 500 750 1000 1250 1500 1750 2000 0 25 50 176 100 125 50 75 100 125 50 175 50 175 50 175 50 775 50 775 50 775 50 775 50 775 50 775 50 775 50 775 50 775 50 775 50 775 50 775 50 775 50	50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 50/0 50/0 5/45 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 0/50 50/0 50/0 50/0 50/0 6/44 0/50 0/50 50/0	50/0 50/0 51/45 0/50 0/50 0/50 0/50 0/50 0/50 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0
OXY	0 250 500 750 1000 1250 1500 1750 0 0 25 50 75 100 125 50 75 100 125 50 75 100 125 50 75 100 125 50 75 100 125 50 75 100 100 100 100 100 100 100 100 100 10	50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 50/0 51/45 0/50 0/50 0/50 0/50 0/50 0/50 0/50 0	50/0 50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 0/50 50/0 50/0 5	50/0 50/0 50/0 50/0 50/0 50/0 0/50 0/50
OXY	0 250 500 750 1000 1250 1500 1750 2000 0 25 50 176 100 125 50 75 100 125 50 175 50 175 50 175 50 775 50 775 50 775 50 775 50 775 50 775 50 775 50 775 50 775 50 775 50 775 50 775 50 775 50	50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 50/0 50/0 5/45 0/50 0/50	50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 0/50 50/0 50/0 50/0 50/0 6/44 0/50 0/50 50/0	50/0 50/0 51/45 0/50 0/50 0/50 0/50 0/50 0/50 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0 50/0
OXY	0 250 500 750 1000 1250 1500 1750 0 0 25 50 75 100 125 50 75 100 125 50 75 100 125 50 75 100 125 50 75 100 125 50 75 100 100 100 100 100 100 100 100 100 10	50 50 50 50 50 50 50 50 50 50	50/0 50/0 50/0 50/0 51/45 0/50 0/50 0/50 0/50 0/50 0/50 0/50 0	50/0 50/0 50/0 50/0 6/44 0/50 0/50 0/50 0/50 0/50 50/0 50/0 5	50/0 50/0 50/0 50/0 50/0 50/0 0/50 0/50

# Specificity and cross reactivity

To test the specificity of the test, the test device was used to test various drugs, drug metabolites and other components of the same class that are likely to be present in urine. All the components were added to drug-free normal human urine. The following structurally related compounds produced positive results with the test when tested at levels equal to or greater than the concentrations listed below.

Item	Concentration (ng/ml)	Item	Concentration (ng/ml)
Amphetamine (AMP)	(g)	Methamphetamine (MET)	(g)
d-Amphetamine	1.000	D(+)-Methamphetamine	1,000
d.I-Amphetamine	3,000	D-Amphetamine	50,000
1-Amphetamine	50.000	Chloroquine	50.000
(+/-) 3,4-methylenedioxyamphetamine (MDA)	5.000	(+/-)-Ephedrine	50.000
Phentermine	3.000	(-)-Methamphetamine	25.000
d-methamphetamine	>100.000	(+/-)3,4-methylenedioxumethamphetamine(MDMA)	2.000
I-methamphetamine	>100,000	8-Phenylethylamine	50.000
3.4-Methylenedioxyethylamphetamine(MDE)	100,000	Trimethobenzamide	10.000
(+/-)3,4-methylenedioxumethamphetamine	100,000	Methylenedioxymethamphetamine (MDMA)	,
Barbiturates (BAR)		3,4-Methylenedioxymethamphetamine HCI	500
Secobarbital	300	3,4-Methylenedioxyamphetamine HCI (MDA)	3.000
Amobarbital	300	3.4-Methylenedioxyethylamphetamine (MDE)	300
Alphenol	150	Morphine (MOP/OPI300)	
Aprobarbital	200	Morphine	300
Butabarbital	75	Codeine	300
Butathal	100	Ethyl Morphine	300
Butalbital	2,500	Hydrocodone	5,000
Cyclopentobarbital	600	Hydromorphone	5,000
Pentobarbital	300	Morphine-3-β-d-glucuronide	1,000
Phenobarbital	10,000	Thebaine	30,000
Benzodiazepines (BZO)		Opiate (OPI2000)	
Oxazepam	300	Morphine	2,000
Alprazolam	200	Codeine	2,000
a-Hydroxyalprazolam	1,500	Ethylmorphine	5,000
Bromazepam	1,500	Hydrocodone	12,500
Chlordiazepoxide	1,500	Hydromorphine	5,000
Clonazepam HCI	800	Levorphanol	75,000
Clobazam	100	σ-Monoacetylmorphine	5,000
Clonazepam	800	Morphine 3-β-D-glucuronide	2,000
Clorazepate dipotassium	200	Norcodeine	12,500
Delorazepam	1,500	Normorphone	50,000
Desalkylflurazepam	400	Oxycodone	25,000
Diazepam	200	Oxymorphine	25,000
Estazolam	2,500	Procaine	150,000
Flunitrazepam	400	Thebaine	100,000
D,L-Lorazepam	1,500	Phencyclidine (PCP)	
Midazolam	12,500	Phencyclidine	25
Nitrazepam	100	4-Hydroxyphencyclidine	12,500
Norchlordiazepoxide	200	Tricyclic Antidepressants (TCA)	
Nordiazepam	400	Nortriptyline	1,000
Temazepam	100	Nordoxepin	1,000
Triazolam	2,500	Trimipramine	3,000
Marijuana (THC)		Amitriptyline	1,500
11-nor-Δ9-THC-9-COOH	50	Promazine	1,500
11-nor-∆8-THC-9-COOH	30	Desipramine	200
11-hydroxy-Δ9-Tetrahydrocannabinol	2,500	Imipramine	400
Δ8- Tetrahydrocannabinol	7,500	Clomipramine	12,500
Δ9- Tetrahydrocannabinol	10,000	Doxepin	2,000
Cannabinol	100,000	Maprotiline	2,000
Cannabidiol	100,000	Promethazine	25,000
Cocaine (COC)		Oxycodone(OXY)	
Benzoylecgonine	300	Oxycodone	100
Cocaine HCI	750	Dihydrocodeine	20,000
Cocaethylene	12,500	Codeine	100,000
Ecgonine	32,000	Hydromorphone	100,000
Methadone (MTD)		Morphine	>100,000
Methadone	300	Acetylmorphine	>100,000
Doxylamine	50,000	Buprenorphine	>100,000
Buprenorphine(BUP)		Ethylmorphine	>100,000
Buprenorphine	10		
Buprenorphine -3-D-Glucuronide	15		
Norbuprenorphine	20		
Norbuprenorphine 3-D-Glucuronide	200		

## Adulteration/Specimen Validity Test

Cr: Creatinine reacts with a creatinine indicator in an alkaline medium to form a purplish-brown color complex if creatinine in the urine is present at the normal level. The color intensity is directly proportional to the concentration of creatinine. A urine sample with creatinine concentration of less than 20 mg/dl produces a very light, or no pad color change, which indicates adulteration in the form of specimen dilution.

S.G.: The specific gravity test is based on the pKa change of certain pretreated polyelectrolytes in relation to the ionic concentration. The pad colors will change from dark blue to blue-green in urine of low ionic concentration to green and yellow-green in urine of higher ionic concentration. A urine specific gravity below 1.003 or above 1.025 is considered abnormal.

OX: Bleach or other oxidizing agents react with an oxidant indicator to form a color complex. Observation of a blue-green, medium to dark brown, or orange color indicates adulteration with bleach or other oxidizing agents.

pH: Normal urine pH ranges from 4.5 to 8.0. Values below pH 4.0 or above pH 9.0 are indicative of adulteration.

GI.: Glutaraldehyde is not a natural component of human urine and it should not be present in normal urine. The presence of glutaraldehyde in the urine sample indicates the possibility of adulteration. However, false positive may result when ketone bodies are presence in urine. Ketone bodies may appear in urine when a person is in ketoacidosis, starvation or other metabolic abnormalities.

Ni.: Although nitrite is not a normal component of urine, nitrite levels of up to 3.6 mg/dl may be found in some urine specimens due to urinary tract infections, bacterial contamination or improper storage. In this adulteration control, nitrite level above 7.5 mg/dl is considered abnormal.

Pyridinium Chlorochromate: The presence of any chromate in urine is indicative of adulteration as chromate is not a normal constituent of urine.

## Interfering substances

Clinical urine samples may contain substances that could potentially interfere with the test. The following compounds were added to drug-free urine, with a drug concentration 25% below the cutoff, and urine with a drug concentration 25% above the cutoff and were tested with the Alere iCup® Rx Drugs of Abuse Cup. All potential interferents were added at a concentration of 100 µg/mL. None of the urine samples showed any deviation from the expected results.

Acetominophen (4-Acetamidophenol) (except OXY test) Oxolinic acid Acetophenetidin Furosemide Oxymetazoline N-Acetylprocainamide (except OXY test) Gentisic acid Papaverine Acetylsalicylic acid Hydralazine (except BZO test) Penicillin-G Aminopyrine Hydrochlorothiazide (except BZO test) Pentobarbital (except BZR, OXY test) Amoxicillin Hydrocodone (except Perphenazine BZO,MOP/OPI300,OXY test) Ampicillin Hydrocortisone Apomorphine O-Hydroxyhippuric acid Phencyclidine(except PCP, OXY tests) 3-Hydroxytyramine Prednisone Atropine (except BAR test) Ibuprofen (except OXY test) Procaine (except BZO, MOP/OPI300, OXY D.I.-Isoproterenol (except AMP, BAR test) DI -Propranolol Renzilic acid D-Propoxyphene (except OXY, test) Benzoic acid Benzoylecgonine (except COC,OXY test) Ketamine (except OXY test) D-Pseudoephedrine(except AMP, BAR tests) Riliruhin Cannabidiol (except THC, OXY tests) Labetalol Ranitidine Chloralhydrate Loperamide Salicylic acid Maprotiline (except TCA, OXY tests) Secobarbital (except BAR, OXY tests) Chloramphenicol Meneridine (except THC, OXY tests) Chlorothiazide Serotonin (5- Hydroxytyramine) Chlorpromazine Sulfamethazine Methadone (except MTD, OXY tests) Chlorquine Sulindac Tetrahydrocortisone, 3-acetate (except AMP, BAR, OXY tests) Tetrahydrocortisone 3-(β-Dglucuronide) (except AMP, BAR, OXY tests) Morphine -3-β-d-glucuronide (except BZO, Clonidine MOP/OPI300 tests Codeine (except MOP/OPI300 BZO OXY Nalidivic acid Tetrahydrozoline tests) Thiamine Cortisone (-) Cotinine Naltrexone Thioridazine Creatinine Naproxen Triamterene Deoxycorticostero Niacinamide DL-Tyrosine Dextromethorphan Nifedipine Trifluoperazine Norcodein (except MOP/OPI300, BZO, OXY Diclofenac Trimethonrim tests) Diflunisal Norethindrone D L-Tryptophan (except AMP, BAR tests) Tyramine (except AMP, BAR tests) Digoxin D-Norpropoxyphene Diphenhydramine Noscapine Uric acid Ecgonine methyl ester D,L-Octopamine Erythromycin (except BZO test) Oxalic acid Zomepirac Oxazepam (except BZO, OXY tests) β-Estradiol (except BZO test)

# BIBLIOGRAPHY OF SUGGESTED READING

Baselt, R.C. Disposition of Toxic Drugs and Chemicals in Man. Biomedical Publications, Davis, CA, 1982. Ellenhorn, M.J. and Barceloux, D. G Medical Toxicology. Elservier Science Publishing Company, Inc., New York,

Gilman, A. G., and Goodman, L. S. The Pharmacological Fluids, in Martin WR(ed): Drug Addiction I, New York, Spring - Verlag, 1977.

Harvey, R.A., Champe, P.C. Lippincotts Illustrated Reviews. Pharmacology. 91-95, 1992.

Hawwks RL, CN Chiang. Urine Testing for drugs of Abuse. National Institute for Drug Abuse (NIDA), Research

Hofmann F.E., A Handbook on Drug and Alcohol Abuse: The Biomedical Aspects, New York, Oxford University

McBay, A. J. Clin. Chem. 33,33B-40B, 1987.

# INDEX OF SYMBOLS



Keep away from sunlight



Store between 4-30°C (39-86°F)





Keep dry



Do not re-use

© 2017 Alere All rights reserved. The Alere Logo. Alere and iCup. are trademarks of the Alere group of companies. All other trademarks

Alere Toxicology Services - Products Division

Manufactured for:

Portsmouth, VA 23704 USA