TEST REPORT: 7191138366-CHM16-02-RC

Date:

20 MAY 2016

Tel: +65 68851345 Fax: +65 67732912

Client's Ref: 4500002083

Email: Randy.CHIN@tuv-sud-psb.sg

Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.



PSB Singapore

SUBJECT

Choose certainty.
Add value.

Microbiological Analysis

CLIENT

Corlison Private Limited 7030 Ang Mo Kio Avenue 5 #06-36/39 Northstar@Amk Singapore 569880

<u>Attn</u>

:Mr. Joe Koh

SAMPLE COLLECTION DATE / TEST DATE

16 May 2016 / 16 May 2016

DESCRIPTION OF SAMPLE

1 sample of Mouthrinse was submitted and labelled as below

No.	Source
1	Pearlie White Fluorinze Anti-bacterial Fluoride Mouthrinse 750ml Batch Number: PFN 0170 Exp: APR 2019

METHOD OF TEST

The United States Pharmacopeia, USP 38 (2015)

<61> Microbiological Examination Of Nonsterile Products: Microbial Enumeration Tests



Laboratory: TÜV SÜD PSB Pte, Ltd. No.1 Science Park Drive Singapore 118221 Phone: +65-6885 1333 Fax: +65-6776 8670 E-mail: testing@tuv-sud-psb.sg www.tuv-sud-psb.sg Co. Reg: 199002667R Regional Head Office: TÜV SÜD Asia Pacific Pte. Ltd. 1 Science Park Drive, #02-01 Singapore 118221 TÜV®

TEST REPORT: 7191138366-CHM16-02-RC 20 MAY 2016



RESULTS

Test	Total Aerobic Microbial Count, CFU/g
Pearlie White Fluorinze Anti-bacterial Fluoride Mouthrinse 750ml Batch Number: PFN 0170 Exp: APR 2019	Not Detected*

Notes:

CFU: Colony Forming Unit TSA: Tryptone Soya Agar

* : Lowest detection limit is 10 CFU/g

Remarks:

The above test results relate to the sample as received.

MS AW HWEE YING HIGHER TECHNICAL

EXECUTIVE

MR RANDY CHIN KOK FEI PRODUCT MANAGER MICROBIOLOGY

CHEMICAL & MATERIALS

- - - -