

■ Panadent Corporation

580 S. Rancho Avenue • Colton, California 92324, USA Tel: (909) 783-1841 • USA & Canada (800) 368-9777

CE

These instructions apply to the following items: REF 2850, 1610, 1620, 1210, 1211, 1230, 1231 Includes: 2860%

Magna-Split II System Instructions

1




Fasten Metal Mounting Plate to Articulator with fastening screws.

3



Place plastic index plate to metal magnetic mounting plate.

5



To prevent losing the magnet assembly, <u>Do not</u> try to remove your casts by unscrewing the knob, the magnet housing and magnet will fall out or stay with the mounted casts.

2




Attach Magnet to Mounting Plate with mounting plate screw.

4



Add plaster to plastic index plate and study cast to mount models in usual manner.

6



To prevent casts from separating from mounting plate, grasp the cast and mounting plate with thumb and fingers. Tilt the mounted cast assembly sideways to release cast from mounting plate magnet. (Do not pull on mounted casts.)

Check to make sure your magnet is still on your articulator, without a magnet, your stone transfer plates will not connect.