



EG – Conformity declaration

The producer: **TecSense GmbH**
Teslastraße 4
A - 8074 Grambach
Tel.: +43 (0)316 40 35 80

Declares herewith that the following products:

Product notation: TecPen / TecWeld – Optochemical Oxygen Sensors
Type designation: TecPen / TecWeld
Serial number:
Year of manufacture: from 2017 ongoing

are in agreement with the regulations of the EG instruction (2014/30/EU) about the electromagnetic compatibility.

Following harmonised norms were applied:

ÖVE/ÖNORM EN 61326-1 Elektrische Mess-, Steuer-, Regel- und Laborgeräte – EMV-Anforderungen;
Teil 1: Allgemeine Anforderungen laut Report:
CE_Testreport_TecLab_TecPac_EN 61326_2013

ÖVE/ÖNORM EN 61326-1 Requirements on electrical measurement, steering, control and laboratory equipment
Part 1: General requirements according to the report:

CE_Testreport_TecLab_TecPac_EN 61326_2013

The electromagnetic compatibility was tested by Seibersdorf Labor GmbH.

Name of the person responsible for this document: Dr. Johannes Krottmaier
Address: identical with the address of manufacturer

TECSENSE GmbH
Teslastraße 4
A-8074 Grambach
Tel.: +43 (0)316 40 35 80
office@tecsense.com

Grambach / 15.12.2017
Location / Date /

/ Dr. Johannes Krottmaier, CEO TecSense
Person