

# CHILD TORSO PHANTOM 1-YEAR-OLD CHILD

Age  
Category

Child

Body  
Region

Torso

Target  
Modality


CT, X-ray

Diagnostic  
FeaturesVasculature, soft and  
bone tissue

This phantom simulates a 1-year-old child's contrast medium enhanced thorax, abdomen and pelvis in portal venous phase. It covers the sixth cervical vertebra to the perineum. The colon is filled with contrast medium as after rectal administration.

The phantom can be used in CT (including CBCT) and X-ray imaging to evaluate and optimize imaging performance and post-processing applications, including AI-enabled applications. It is also suited for training purposes.

The phantom provides a detailed and realistic simulation of soft and bone tissue. Air voids including those of the lungs are filled with a cellulose-polymer composite of approx. -160 HU.



# CHILD TORSO PHANTOM 1-YEAR-OLD CHILD




## Specifications

Size	Approx. 210 x 145 x 313 mm
Weight	Approx. 4400 g
Base material	Cellulose-polymer composite
Optimal tube voltage	100 kVp (cf page 3) - adaptable upon request -

## Diagnostic features

Realistic simulation of vasculature, bone and soft tissues, including the lungs, heart, liver, gallbladder, pancreas, spleen, adrenals, kidneys, stomach, small intestine, colon and bladder.



For more information visit  
[www.phantomx.de](http://www.phantomx.de)

# CHILD TORSO PHANTOM 1-YEAR-OLD CHILD


## General indications

- The phantom is made of a cellulose-polymer composite material with properties similar to hardwood. If handled carefully, it will last a long time.
- The phantom is coated with a protective layer. If the protective layer is undamaged, the phantom can be cleaned using a damp cloth (water or mild detergent).
- Protect from direct sunlight.
- Maintain a storage temperature of 10 °C to 30 °C. If the phantom is exposed to temperatures below -10 °C or above 45 °C, it can be severely damaged.
- The phantom is not equipped for dose measurements with dosimeters and it is not suited for material characterization with dual energy CT.
- The phantom is not certified as medical device.
- Air voids are filled with cellulose-polymer composite of approx. -160 HU.


## Attenuation properties

### Soft Tissue

Linear attenuation coefficients [cm<sup>-1</sup>] (calculated)




### Hounsfield units (calculated)




### Bone Tissue

Linear attenuation coefficients [cm<sup>-1</sup>] (calculated)



### Hounsfield units (calculated)



Tissue Reference: Woodard HQ, White DR. The composition of body tissues. Br J Radiol. 1986.