

EUROFINS Medical Device Testing

2425 New Holland Pike Lancaster, PA 17601 Telephone: 717-656-2300

Fax: 717-656-3772

FINAL REPORT: Analytical Report for Particulate Filtration Efficiency, N95 Masks

Prepared for:

Breathh, Inc. 9830 Siempre Viva Rd. #15 San Diego CA - 92154 USA

Eurofins Medical Device, Study #: STUUY20AA0174



Page: 1 / 6

Analytical Report Nr. Sample code Nr. Date

AR-20-YL-007675-01 560-2020-00006491 12/11/2020

ANALYTICAL REPORT

Client Information

Eurofins Medical Device Testing 2425 New Holland Pike Lancaster USA +1 717 725 8250 KellyConner@eurofins.com

For the attention of Kelly Conner

Sample Information

Order Code: EUAA70-00008339

23-Sep-2020 Reception Date: **Analysis Starting Date:** 23-Sep-2020 Analysis Ending Date: 12-Nov-2020

Sample code Nr. 560-2020-00006491

Sample described as: Masks

Requirements and decision rule

Customer requirements: No requirements **Decision Rule:** Not applicable.

Information provided by the customer*

Client Reference: EN 14863-19 & ASTM F2100-19

Sample Description: On behalf of Breathh Inc/ Lutema Products Worldwide

Purchase Order Number:



Analytical Report Nr. Sample code Nr. Date Page: 2 / 6 AR-20-YL-007675-01 560-2020-00006491 12/11/2020

SAMPLE PICTURE





Page: 3 / 6

 Analytical Report Nr.
 AR-20-YL-007675-01

 Sample code Nr.
 560-2020-00006491

 Date
 12/11/2020

CONCLUSION:

TEST PROPERTY	PASS	FAIL	REMARKS
◆ Determination of face mask materials efficiency to penetration by particulates (latex spheres) ASTM F2299 / F2299M - 03(2017)			
A			REFER RESULT

Remark: Test has been performed as per application request



Page: 4 / 6

Analytical Report Nr. Sample code Nr. AR-20-YL-007675-01 560-2020-00006491

Date

12/11/2020

COMPONENT LIST:

COMPONENT ID	COMPONENT NAME	MATERIAL DESCRIPTION	COLOR	REMARKS
CUST 01	A	Mask	Pink	

AR-20-YL-007675-01



Page: 5 / 6

Analysis Ending Date: 12/11/2020

Analytical Report Nr. Sample code Nr.

560-2020-00006491

Date

12/11/2020

MASKS TESTING CAS No. RESULTS UNC. LOQ GUIDELINES

99.82 %

Analyses on:A

Determination of face mask materials efficiency to penetration by particulates (latex spheres)

ASTM F2299 / F2299M - 03(2017)

Filtration Efficiency

Particle diameter: 0.1 $\mu m \pm 0.015~\mu m$ Complete test data reported at Annex.



Analytical Report Nr. Sample code Nr. Date Page: 6 / 6 AR-20-YL-007675-01 560-2020-00006491 12/11/2020

Signed for and on behalf of Eurofins Textile Testing Spain:



Report electronically validated by

Maria Jesus Martinez Puig Chemical Lab manager

EXPLANATORY NOTE

- ◆ Test not covered by ENAC accreditation scope
- Test is subcontracted within Eurofins group and is accredited
- Test is subcontracted within Eurofins group and is not accredited
- Test is subcontracted outside Eurofins group and is accredited
- □ Test is subcontracted outside Eurofins group and is not accredited N/A = Not Applicable

Eurofins Textile Testing Spain S.L.U is not responsible of the information supplied by the costumer and reported as section "Information provided by the costumer".

Eurofins General Sales Terms and Conditions Applied.

Results obtained refer only to samples, products or material received in Laboratory, as described in section "Sample information" and tested in conditions shown in present report.

Test uncertainties not reported are at customer disposal, for those tests in which it is possible to evaluate the test uncertainty.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k = 2, which for a normal distribution provides a level of confidence of approximately 95%.

Reproduction of this document is only valid if it is done completely and under the written permission. Any printed version of this document will be considered as a copy.

If you happen to have any comments, please do it by sending email to textile_spain@eurofins.com and referring to this report number.

End Of Report



Test report Annex Analytical Report Nr. AR-20-YL-007675-01 Sample code Nr. 560-2020-00006491

DETERMINATION OF THE INITIAL EFFICIENCY OF MATERIALS USED IN MEDICAL FACE MASKS TO PENETRATION BY PARTICULATES USING LATEX SPHERES

Standard: ASTM F2299 / F2299M - 03(2017)

Number of test specimens: 5

Thickness: 1,43 mm

Unit area weight: 228,40 g/m² Exposed specimen area: 23,8 cm²

Size of test specimen was facing towards the challenge aerosol: Inner

Face velocity: 6 cm/s

Pressure drop: -0,3 mmH₂O

Particle diameter: 0.1 µm ± 0.015 µm

Conditioning and test conditions: Ta between 18°C and 24°C. HR% between 30 and 50 % Hr

Pretreatment techniques used: No

Test duration: Three one-minute counts per specimen

Controls used: Three one-minute control counts were performed, without testing sample, before and

after each 5 specimens. Control counts are averaged.

Testing date: 12/11/2020

Deviation from the stated method: Test aerosol is non-neutralized according to the FDA guidance

document on surgical face masks.

Test results:

Specimen	Filtration Efficiency (%)		
1	99,86		
2	99,83		
3	99,83		
4	99,76		
5	99,82		
Average	99,82		
Standard Deviation	0.04		