

N95 Particulate Respirator

Foldable Non-sterile

BYD CARE 

- ⊕ Very close facial **fit**
- ⊕ **≥95%** filtrations efficiency against solid and liquid aerosols free of oil.



* Picture is for reference only, actual product may have differences.

Description

The Particulate Respirator DE2322 is designed to help provide respiratory protection for the user. This product has a filter efficiency of 95% or greater against solid and liquid aerosols free of oil. This product contains no components made from natural rubber latex.

Use instructions

- Suitable for protecting the mouth, nose, and chin to airborne particles and block liquid aerosols.
- For adults use only.
- Do not wash.
- Store in a cool, dry, clean place away from fire and contamination.
- Store between -20°C and 38°C (-4°F and 100°F) with relative humidity below 80%.
- Expiration date: 2 years after production.

3-Step check before use

1. Do not use if packaging has been opened or damaged.
2. Check if elastic bands are in good condition and are not damaged.
3. Check that the metallic strip is not broken.

Wearing instructions

1. Unfold the mask and hold it with both hands, with the metallic strip facing up.
2. Place elastic bands around the neck and head respectively. Put the mask against your face covering both nose and mouth.
3. Adjust the metallic strip cover bridge of the nose using two fingers to press down until achieving a close fit.
4. Perform a fit check according to instructions.




Fit check

1. Place both hands over mask.
2. Take a deep breath and hold your breath for a few seconds, making sure the mask collapses inward as you inhale.
3. Then exhale and hold your breath for a few seconds, making sure the mask bulges outward as you exhale.
4. If air leaks, reposition the mask, tighten the metallic strip and reposition the elastic bands for a better fit. Follow steps 1-3 again until a tight seal has been achieved.
5. Masks that have passed the fit check are now safe to use.
6. Facial hair or other facial features may prevent the mask from achieving a proper fit. If a tight seal is not achievable, please DO NOT use this product.
7. If you experience difficulty breathing, dizziness, or other symptoms, leave the contaminated area immediately.

N95 Particulate Respirator

Foldable Non-sterile

Specifications



| Name | Parameter |
|-------------------------|--|
| Product name | N95 Particulate Respirator |
| Materials | <p>The following materials are used in the production of N95 Particulate Respirator.</p> <p>Inner layer: polypropylene spunbond nonwoven fabric; Middle layer: polypropylene melt-blown nonwoven + hot air cotton; Outer layer: polypropylene spunbond nonwoven fabric ; Elastic bands: polyester/nylon spandex blend; Nose clip: aluminum</p> |
| Product model | DE2322 (Foldable) |
| Packaging specification | 1pc/bag; 20pcs/box; 960pcs/case |

WARNINGS

1. The mask does not eliminate the risk of contracting any disease or infection.
2. Improper use may lead to illness and even death.
3. Use this product immediately after package is opened.
4. Not recommended for more than 8 hours of use.
5. DO NOT sleep while wearing the mask.
6. Avoid hand contact within the interior part of the mask.
7. DO NOT use masks if they expired.
8. For one-time use only. Dispose the mask according to regulations.
9. DO NOT use if package is damaged.



Time use limitation

If the mask becomes damaged, soiled, or breathing becomes difficult, leave the contaminated area and replace and refit the new mask.

Important notice

1. To the extent permitted by law, BYD shall not be liable for any loss or damage including any loss of business, loss of profits, or for any indirect, special, incidental or consequential loss or damage arising from reliance upon any information herein provided by BYD. Nothing in this statement will be deemed exclude or restrict BYD's liability for death or personal injury arising from its negligence.
2. Staples are used for strap fastening, it is normal that small punctures be found around the staples. Product sampling has been processed to ensure the respirators meet Part 84 approval requirement.
3. The process of the strap fastening has been strictly controlled. Any enlarged holes resulting from ripped filter material around staple punctures are considered as damage.
4. Filtering facepieces are to be inspected prior to each use to assure there are no holes in the breathing zone other than punctures around staples and no damage has occurred.
5. Before occupational use of this respirator, a written respiratory protection program must be implemented meeting all the local government requirements. In the United States, employers must comply with OSHA 29 CFR 1910.134 which includes medical evaluation, training, and fit testing.

N95 Particulate Respirator

Foldable

Non-sterile



This respirator is approved only in the following configuration:

| TC- | Protection ¹ | Respirator | Cautions and Limitations ² |
|----------|-------------------------|------------|---------------------------------------|
| | | DE2322 | |
| 84A-9221 | N95 | X | ABCJMNOP |

1. Protection

N95 – Particulate Filter (95% filter efficiency level).
Effective against particulate aerosols free of oil; time use restrictions may apply.

2. Cautions and Limitations

- A- Not for use in atmospheres containing less than 19.5% oxygen.
- B- Not for use in atmospheres immediately dangerous to life or health.
- C -Do not exceed maximum use concentrations established by regulatory standards.
- J -Failure to properly use and maintain this product could result in injury or death.
- M-All approved respirators shall be selected, fitted, used, and maintained in accordance with MSHA, OSHA, and other applicable regulations.
- N-Never substitute, modify, add, or omit parts. Use only exact replacement parts in the configuration as specified by the manufacturer.
- O- Refer to User Instructions and/or maintenance manuals for information on use and maintenance of these respirators.
- P- NIOSH does not evaluate respirators for use as surgical masks.

Approval and Standard



GB 19083-2010 standard



6 Precision Lane, Notting Hill, VIC 3168
<https://www.ausmedhealth.com.au/>
contact@ausmedhealth.com.au



AusMed Health Pty Ltd

Mobile Enquiries:

Paul 0414 565 840

Frank 0419 511 731