

## Enuresis/Bed Wet Sensor

Part No. **QT635**



### Overview

Designed to trigger a call when moisture is detected in a patient's bed.

Connects to Quantec via the jack socket on the QT637 Enuresis/general purpose interface socket.

Typically positioned across and over a mattress, directly underneath the bed sheet.

Protects against patient discomfort and reduces the number of incontinence checks staff are required to make.

Measures 80cm x 12cm and is supplied with a 2m lead and normally open right angled jack plug.

Wipe clean design - can be re-used many times.

### Technical Specifications

Compatibility	Must be used in conjunction with a QT637 Quantec Enuresis Sensor/General Purpose Interface Socket.
Indicators	Enuresis Active indicator on QT637 will flash red every four seconds when the QT635 detects moisture.
Connections	6.35mm (quarter inch) jack plug.
Product dimensions (mm)	800mm W x 111mm H. Includes a 2m lead.
Construction & finish	Plastic.
IP Rating	TBC.
Weight	140g
Operating conditions/temperature	-5°C to +40°C. Max relative humidity: 95%.

# Comsec Protection Systems



T 01942 322744 E [sales@c-tec.co.uk](mailto:sales@c-tec.co.uk)  
F 01942 829867 W [www.c-tec.com](http://www.c-tec.com)



C-TEC, Challenge Way, Martland Park, Wigan, WN5 0LD United Kingdom