



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

San Diego Scale A Data Support Company

4510 Federal Boulevard, San Diego, CA 92102

*(Hereinafter called the Organization) and hereby declares that Organization is accredited
in accordance with the recognized International Standard:*

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the
operation of a laboratory quality management system
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Mass Force and Weighing and Mechanical Calibration (As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

October 13, 2023

Issue Date:

October 13, 2023

Expiration Date:

January 31, 2026

Accreditation No.:

108285

Certificate No.:

L23-749

*The validity of this certificate is maintained through ongoing assessments based on a
continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjilabs.com*



Certificate of Accreditation: Supplement

San Diego Scale Co.

4510 Federal Boulevard, San Diego, CA 92102
Contact Name: Josh Banayan Phone: 619-262-7000

Accreditation is granted to the facility to perform the following calibrations:

Mass Force and Weighing

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Micro-Balances and Ultra-Micro-Balances ^{FO}	Up to 50 g	$(1 \times 10^{-4} + 6.2 \times 10^{-7}Wt) \text{ g}$	Class E1 weights 1 mg to 50 g (EURAMET 18 V4)
Semi-Micro Balances ^{FO}	Up to 120 g	$(1 \times 10^{-4} + 6.32 \times 10^{-7}Wt) \text{ g}$	Class E1 weights 1 mg to 120 g (EURAMET 18 V4)
Analytical Balances ^{FO}	Up to 500 g	$(2 \times 10^{-3} + 1.63 \times 10^{-6}Wt) \text{ g}$	Class E2 weights 1 mg to 500 g (EURAMET 18 V4)
Top Loading Laboratory Balances ^{FO}	Up to 30 kg	$(1.2 \times 10^{-3} + 5.74 \times 10^{-6}Wt) \text{ g}$	Class F1 weights 1 mg to 10 kg Class F1 weight 20 kg (EURAMET 18 V4)
Top Loading Industrial Scales ^{FO}	Up to 30 kg	$(0.116 + 1 \times 10^{-10}Wt) \text{ g}$	Class F1 weights 1 mg to 10 kg Class F1 weights 20kg (EURAMET 18 V4)
Industrial scales ^{FO}	Up to 60 kg	$(0.116 + 1.67 \times 10^{-5}Wt) \text{ g}$	Class F2 weights 10 kg, 20 kg, 50 kg (EURAMET 18 V4)

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Pipettes ^{FO}	0.5 μL to 2 μL	0.076 μL	Balance SOP-33 Pipettes
	2 μL to 10 μL	0.078 μL	
	10 μL to 20 μL	0.08 μL	
	20 μL to 100 μL	0.094 μL	
	100 μL to 200 μL	0.16 μL	
	200 μL to 500 μL	0.4 μL	
	500 μL to 1 000 μL	0.51 μL	
	1 000 μL to 5 100 μL	2.5 μL	

- The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
- The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.



Certificate of Accreditation: Supplement

San Diego Scale Co.

4510 Federal Boulevard, San Diego, CA 92102
Contact Name: Josh Banayan Phone: 619-262-7000

Accreditation is granted to the facility to perform the following calibrations:

3. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.

