



# Versana Balance™

## Transducer Guide





4C-RS



L6-12-RS



12L-RS



E8C-RS



E8CS-RS



3SC-RS



6S-RS



RAB2-6-RS

	Applications	Footprint	Frequency Range	FOV	Biopsy Guide
	Abdominal, OB/GYN, Vascular, Urological, Thoracic, Pediatric, MSK	18.3 x 66.2 mm	2.0 – 5.0 MHz	58°	E8385NA
	Vascular, Small parts, Pediatrics, MSK, Thoracic	11.4 x 47.0 mm	4.0 – 13.0 MHz	38.4 mm	H40432LC
	Vascular, Small parts, Pediatrics, MSK, Thoracic	12.7 x 47.1 mm	4.2 – 13.0 MHz	38.4 mm	H48392LT H48392LL
	OB/GYN, Urological, Transvaginal, Transrectal	16.9 x 21.2 mm	4.2 – 10.0 MHz	128°	E8385MJ or H40412LN
	OB/GYN, Urological, Transvaginal, Transrectal	18.6 x 13.9 mm	4.0 – 10.0 MHz	168°	E8385MJ or H40412LN
	Cardiac, Vascular, Transcranial, Thoracic, Abdominal	23.7 x 18.4 mm	1.7 – 4.0 MHz	120°	H46222LC
	Cardiac, Vascular, Transcranial, Pediatrics	23.5 X 16.8 mm	2.5 – 7.0 MHz	120°	not available
	Abdominal, OB/GYN, Urological	62.2 x 34.0 mm	2.0 – 6.0 MHz	70° (B), 70° x 84° (Volume scan)	H48681ML



8C-RS



LK760-RS

Applications	Footprint	Frequency Range	FOV	Biopsy Guide
Pediatrics, Pediatric Cardiac, MSK	12.0 x 22.0 mm	4.2 – 10.0 MHz	131°	not available
MSK	67 x 13 mm	3.5 – 10.0 MHz	59.9 mm	not available



## Imagination at work

Product may not be available in all countries and regions. Full product technical specification is available upon request. Contact a GE Healthcare Representative for more information. Please visit [www.gehealthcare.com/promotional-locations](http://www.gehealthcare.com/promotional-locations).

Data subject to change.

© 2019 General Electric Company.

GE, the GE Monogram, imagination at work and Versana Active are trademarks of General Electric Company.

Reproduction in any form is forbidden without prior written permission from GE. Nothing in this material should be used to diagnose or treat any disease or condition. Readers must consult a healthcare professional.

August 2019  
DOC2296679