PRODUCT SIZING				Ref	Ref No. 2307-XX	
Measure length of lateral epicondyle to knuckle.						
Size	XX	Min	Max	Min	Max	
XS	10	8"	10"	20cm	25cm	
S	20	10"	12"	25cm	30cm	
Μ	30	12"	14"	30cm	35cm	
L	40	14"	16"	35cm	40cm	
XL	50	16"	18"	40cm	46cm	

## ACCOMMODATOR SHOULDER IMMOBILIZER

## **INDICATIONS:**

## Patient Population: Adults and Children with identified indications

- · Post operative support and positioning
- Injuries requiring immobilization


WARNING: This product is intended for single patient use at the direction of a medical professional and will not prevent or reduce the risk of injury. Do Not Over Tighten. If you experience a sudden increase in pain, numbness, or unusual activity with this device, discontinue use immediately and consult your healthcare provider.

CARE: Hand wash using a mild detergent. Rinse thoroughly. Air Dry only. Do not tumble dry.

54 Hugh Adams Rd., Defuniak Springs, FL 32435 National: (800) 234-9004 International: (850) 892-5731 Fax: (800) 274-9006







- 1. Loosen straps before application.
- 2. Place elbow in the back of envelope and thumb in thumb loop.
- **3.** Position waist and shoulder straps for ease of removal.
- 4. Adjust excess waist and shoulder strap lengths from back; trim as necessary.

Scan for application and accessories.

