PRODU	CT SIZING	Ref	No. 460-XX		
Measure circumference of leg 6" above mid-patella with leg fully extended					
Size	XX	Min		Max	
XS	10	13"	33cm	16"	41cm
S	20	16"	41cm	19"	48cm
M	30	19"	48cm	21"	53cm
L	40	21"	53cm	24"	61cm
XL	50	24"	61cm	27"	69cm
XXL	60	27"	69cm	30"	76cm
XXXI	70	30"	76cm	33"	84cm

© 2023 Professional Products, Inc. www.ezywrap.com Made in the USA



54 Hugh Adams Rd., Defuniak Springs, FL 32435 National: (800) 234-9004 International: (850) 892-5731 Fax: (800) 274-9006



SUPERIOR PATELLA STABILIZER

INDICATIONS:

Patient Population: Adults and Children with identified indications

- Chondromalacia
- · Patellofemoral Malalignment
- · Acute Patella Injuries and Chronic Conditions


WARNING: This product is intended for single patient use at the direction of a medical professional and will not prevent or reduce the risk of injury. Do Not Over Tighten. If you experience a sudden increase in pain, numbness, or unusual activity with this device, discontinue use immediately and consult your healthcare provider.

CARE: Hand wash using a mild detergent. Rinse thoroughly. Air Dry only. Do not tumble dry.




1.





3.





- 1. Place silicone buttress to support top of patella
- 3. Refine silicone buttress alignment with top of the patella

Scan for application and accessories.

