



**Report Number:** 23-010974/D002.R000

Report Date: 09/21/2023 ORELAP#: OR100028

Purchase Order: SCH10-Sept2023 Received: 09/14/23 15:34

**Customer:** CJ2, LLC (DBA Zentopia)

Product identity: SCH10-HB-090723-16:13 (Huckleberry Lemonade)

Client/Metrc ID:

Laboratory ID: 23-010974-0006

**Summary** 

Potency:					
Analyte per 16fl oz CBD per 16fl oz	Result	Limits	Units mg/16fl oz	Status	CBD-Total per Serving Size 54.9 mg/16fl oz
			J		THC-Total per Serving Size <loq< td=""></loq<>
					(Reported in milligrams per serving)





Report Number: 23-010974/D002.R000

**Report Date:** 09/21/2023 **ORELAP#:** OR100028

**Purchase Order:** SCH10-Sept2023 **Received:** 09/14/23 15:34

Customer: CJ2, LLC (DBA Zentopia)

541-801-3097 92 Centennial Loop Eugene Oregon 97401

United States of America (USA)

**Product identity:** SCH10-HB-090723-16:13 (Huckleberry Lemonade)

Client/Metrc ID: .

Sample Date:

**Laboratory ID:** 23-010974-0006

Evidence of Cooling: No
Temp: 24.4 °C
Relinquished by: client
Serving Size #1: 473.65 g
Density: 1.001 g/ml

# Sample Results

Potency per 16fl oz	Method: J AOAC 2015 V	98-6 (mod) <sup>þ</sup>	Units mg/se Batc	<b>h:</b> 2311002	<b>Analyze:</b> 9/16/23 12:16:00 AM
Analyte	Result	Limits	Units	LOQ	Notes
CBC per 16fl oz	< LOQ		mg/16fl oz	0.473	
CBC-A per 16fl oz	< LOQ		mg/16fl oz	0.473	
CBC-Total per 16fl oz	< LOQ		mg/16fl oz	0.888	
CBD per 16fl oz	54.9		mg/16fl oz	0.473	
CBD-A per 16fl oz	< LOQ		mg/16fl oz	0.473	
CBD-Total per 16fl oz	54.9		mg/16fl oz	0.888	
CBDV per 16fl oz	< LOQ		mg/16fl oz	0.473	
CBDV-A per 16fl oz	< LOQ		mg/16fl oz	0.473	
CBDV-Total per 16fl oz	< LOQ		mg/16fl oz	0.883	
CBE per 16fl oz	< LOQ		mg/16fl oz	0.473	
CBG per 16fl oz	< LOQ		mg/16fl oz	0.473	
CBG-A per 16fl oz	< LOQ		mg/16fl oz	0.473	
CBG-Total per 16fl oz	< LOQ		mg/16fl oz	0.883	
CBL per 16fl oz	< LOQ		mg/16fl oz	0.473	
CBL-A per 16fl oz	< LOQ		mg/16fl oz	0.473	
CBL-Total per 16fl oz	< LOQ		mg/16fl oz	0.888	
CBN per 16fl oz	< LOQ		mg/16fl oz	0.473	
CBT per 16fl oz	< LOQ		mg/16fl oz	0.473	
$\Delta 8$ -THCV per 16fl oz	< LOQ		mg/16fl oz	0.473	
$\Delta 10$ -THC-9R per 16fl oz	< LOQ		mg/16fl oz	0.473	
$\Delta 10$ -THC-9S per 16fl oz	< LOQ		mg/16fl oz	0.473	
$\Delta 10$ -THC-Total per 16fl oz	< LOQ		mg/16fl oz	0.946	
$\Delta 8$ -THC per 16fl oz	< LOQ		mg/16fl oz	0.473	
$\Delta 9$ -THC per 16fl oz	< LOQ		mg/16fl oz	0.473	
delta-9-THCP per 16fl oz	< LOQ		mg/16fl oz	0.473	
exo-THC per 16fl oz	< LOQ		mg/16fl oz	0.473	
THC-A per 16fl oz	< LOQ		mg/16fl oz	0.473	
THC-Total per 16fl oz	< LOQ		mg/16fl oz	0.888	

www.columbialaboratories.com

Page 2 of 8





**Report Number:** 23-010974/D002.R000

**Report Date:** 09/21/2023 ORELAP#: OR100028

Purchase Order: SCH10-Sept2023 09/14/23 15:34 Received:

Potency per 16fl oz	Method: J AOAC 2015 V	98-6 (mod) <sup>þ</sup>	Units mg/se Batc	<b>h:</b> 2311002	<b>Analyze:</b> 9/16/23 12:16:00 AM
Analyte	Result	Limits	Units	LOQ	Notes
THCV per 16fl oz	< LOQ		mg/16fl oz	0.473	
THCV-A per 16fl oz	< LOQ		mg/16fl oz	0.473	
THCV-Total per 16fl oz	< LOQ		mg/16fl oz	0.889	
Total Cannabinoids per 16f	l oz 54.9		mg/16fl oz		





**Report Number:** 23-010974/D002.R000

**Report Date:** 09/21/2023 **ORELAP#:** OR100028

**Purchase Order:** SCH10-Sept2023 **Received:** 09/14/23 15:34

#### **Abbreviations**

Limits: Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220, CCR title 16-division 42. BCC-section 5723

**Limit(s) of Quantitation (LOQ):** The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

b = ISO/IEC 17025:2017 accredited method.

### Units of Measure

g = g g/ml = Gram per milliliter mg/16fl oz = Milligram per 16fl oz % = Percentage of sample $% wt = \mu g/g divided by 10,000$ 

Approved Signatory

Derrick Tanner General Manager





**Report Number:** 23-010974/D002.R000

**Report Date:** 09/21/2023 ORELAP#: OR100028

Purchase Order: SCH10-Sept2023 Received: 09/14/23 15:34

Revision: 4 Document ID: 7148 Legacy ID: Worksheet Validated 04/20/2021

#### Laboratory Quality Control Results

JAOAC2015 V986	Batch ID: 2311002							
Laboratory Control S	ample							
Analyte	LCS	Result	Spike	Units	%Rec	Limits	Evaluation	Notes
CBDVA	2	0.00101	0.0009	%	109	80.0 - 120	Acceptable	
CBDV	2	0.00102	0.0009	%	109	80.0 - 120	Acceptable	
CEE	2	0.00106	0.0010	%	108	80.0 - 120	Acceptable	
CBDA	1	0.00116	0.00109	%	106	90.0 - 110	Acceptable	
CBGA	1	0.00118	0.00110	%	107	80.0 - 120	Acceptable	
CBG	1	0.00126	0.00118	%	107	80.0 - 120	Acceptable	
CBD	1	0.00122	0.00114	%	107	90.0 - 110	Acceptable	
THCV	2	0.0007	0.0006	%	109	80.0 - 120	Acceptable	
d8THCV	2	0.0009	0.0008	%	109	80.0 - 120	Acceptable	
THCVA	2	0.0010	0.0009	%	107	80.0 - 120	Acceptable	
CBN	1	0.00120	0.00111	%	108	80.0 - 120	Acceptable	
exo-THC	2	0.0010	0.0009	%	108	80.0 - 120	Acceptable	
d9THC	1	0.00128	0.00115	%	111	90.0 - 110	Acceptable	Q1
d8THC	1	0.00127	0.00117	%	108	90.0 - 110	Acceptable	
9Sd10THC	1	0.00124	0.00115	%	108	80.0 - 120	Acceptable	
CB.	2	0.00104	0.0009	%	111	80.0 - 120	Acceptable	
9R-d10THC	1	0.0004	0.0004	%	105	80.0 - 120	Acceptable	
CBC	2	0.00107	0.00100	%	107	80.0 - 120	Acceptable	
THCA	1	0.00118	0.00111	%	106	90.0 - 110	Acceptable	
CBCA	2	0.00103	0.0010	%	108	80.0 - 120	Acceptable	
CBLA	2	0.00208	0.00194	%	107	80.0 - 120	Acceptable	
d9THCP	2	0.00101	0.0009	%	106	80.0 - 120	Acceptable	
CBI	2	0.00102	0.0009	%	111	80.0 - 120	Acceptable	
Method Blank					·			· · · · · · · · · · · · · · · · · · ·

Analyte	Result	LOQ	Units	Limits	Evaluation	Notes
CBDVA	<l0q< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	< 0.0001	Acceptable	
CBDV	<loq< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></loq<>	0.0001	%	< 0.0001	Acceptable	
CEE	<loq< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></loq<>	0.0001	%	< 0.0001	Acceptable	
CBDA	<l0q< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	< 0.0001	Acceptable	
CBGA	<loq< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></loq<>	0.0001	%	< 0.0001	Acceptable	
CBG	<loq< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></loq<>	0.0001	%	< 0.0001	Acceptable	
CBD	<loq< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></loq<>	0.0001	%	< 0.0001	Acceptable	
THCV	<l0q< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	< 0.0001	Acceptable	
d8THCV	<l0q< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	< 0.0001	Acceptable	
THCVA	<l0q< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	< 0.0001	Acceptable	
CBN	<loq< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></loq<>	0.0001	%	< 0.0001	Acceptable	
exo-THC	<l0q< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	< 0.0001	Acceptable	
d9THC	<l0q< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	< 0.0001	Acceptable	
d8THC	<l0q< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	< 0.0001	Acceptable	
9Sd10THC	<l0q< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	< 0.0001	Acceptable	
CB.	<l0q< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	< 0.0001	Acceptable	
9R-d10THC	<l0q< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	< 0.0001	Acceptable	
CBC	<l0q< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	< 0.0001	Acceptable	
THCA	<loq< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></loq<>	0.0001	%	< 0.0001	Acceptable	
CBCA	<loq< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></loq<>	0.0001	%	< 0.0001	Acceptable	
CBLA	<l0q< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	< 0.0001	Acceptable	
d9THCP	<loq< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></loq<>	0.0001	%	< 0.0001	Acceptable	
CBL	<loq< td=""><td>0.0001</td><td>%</td><td>&lt; 0.0001</td><td>Acceptable</td><td></td></loq<>	0.0001	%	< 0.0001	Acceptable	

Abbreviations

ND - None Detected at or above MRL RPD - Relative Percent Difference

LOQ- Limit of Quartitation

Q1 - Quality control result biased high. Only non-detect samples reported.
Units of Measure:
%- Percent





**Report Number:** 23-010974/D002.R000

**Report Date:** 09/21/2023 ORELAP#: OR100028

Purchase Order: SCH10-Sept2023 Received: 09/14/23 15:34

Revision: 4 Document ID: 7148 Legacy ID: Worksheet Validated 04/20/2021

#### Laboratory Quality Control Results

JAOAC2015 V986				·	Ba	atch ID: 2311002		
Sample Duplicate					San	nple ID: 23-010919	90001	
Analyte	Result	Org. Reult	LOQ	Units	RPD	Limits	Evaluation	Notes
CBDVA	<loq< td=""><td><loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<>	0.0001	%	NA	< 20	Acceptable	
CBDV	0.0001	0.0001	0.0001	%	3.68	< 20	Acceptable	
CEE	<loq< td=""><td><l0q< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></l0q<></td></loq<>	<l0q< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	NA	< 20	Acceptable	
CBDA	<loq< td=""><td><loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<>	0.0001	%	NA	< 20	Acceptable	
CBGA	<loq< td=""><td><loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<>	0.0001	%	NA	< 20	Acceptable	
CBG	<loq< td=""><td><loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<>	0.0001	%	NA	< 20	Acceptable	
CBD	0.0154	0.0150	0.0001	%	2.57	< 20	Acceptable	
THCV	<loq< td=""><td><l0q< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></l0q<></td></loq<>	<l0q< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	NA	< 20	Acceptable	
d8THCV	<loq< td=""><td><loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<>	0.0001	%	NA	< 20	Acceptable	
THCVA	<loq< td=""><td><loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<>	0.0001	%	NA	< 20	Acceptable	
CBN	<loq< td=""><td><loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<>	0.0001	%	NA	< 20	Acceptable	
exo-THC	<loq< td=""><td><loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<>	0.0001	%	NA	< 20	Acceptable	
d9THC	0.00718	0.00706	0.0001	%	1.74	< 20	Acceptable	
d8THC	0.00203	0.00201	0.0001	%	1.05	< 20	Acceptable	
9Sd10THC	<loq< td=""><td><l0q< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></l0q<></td></loq<>	<l0q< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	NA	< 20	Acceptable	
CB.	<loq< td=""><td><loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<>	0.0001	%	NA	< 20	Acceptable	
9R-d10THC	<loq< td=""><td><loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<>	0.0001	%	NA	< 20	Acceptable	
CBC	<loq< td=""><td><l0q< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></l0q<></td></loq<>	<l0q< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	NA	< 20	Acceptable	
THCA	<loq< td=""><td><l0q< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></l0q<></td></loq<>	<l0q< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	NA	< 20	Acceptable	
CBCA	<loq< td=""><td><l0q< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></l0q<></td></loq<>	<l0q< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	NA	< 20	Acceptable	
CB.A	<loq< td=""><td><loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<>	0.0001	%	NA	< 20	Acceptable	
d9THCP	<loq< td=""><td><l0q< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></l0q<></td></loq<>	<l0q< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></l0q<>	0.0001	%	NA	< 20	Acceptable	
CBT	<loq< td=""><td><loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.0001</td><td>%</td><td>NA</td><td>&lt; 20</td><td>Acceptable</td><td></td></loq<>	0.0001	%	NA	< 20	Acceptable	

Abbreviations

ND - None Detected at or above MRL RPD - Relative Percent Difference LOQ- Limit of Quantitation

Units of Measure:

%- Percent





23-010974/D002.R000 **Report Number:** 

Report Date: 09/21/2023 ORELAP#: OR100028

Purchase Order: SCH10-Sept2023 Received: 09/14/23 15:34







23-010974/D002.R000 **Report Number:** 

Report Date: 09/21/2023 ORELAP#: OR100028

Purchase Order: SCH10-Sept2023 09/14/23 15:34 Received:

## Explanation of QC Flag Comments:

Code	Explanation
Q	Matrix interferences affecting spike or surrogate recoveries.
Q1	Quality control result biased high. Only non-detect samples reported.
Q2	Quality control outside QC limits. Data considered estimate.
Q3	Sample concentration greater than four times the amount spiked.
Q4	Non-homogenous sample matrix, affecting RPD result and/or % recoveries.
Q5	Spike results above calibration curve.
Q6	Quality control outside QC limits. Data acceptable based on remaining QC.
R	Relative percent difference (RPD) outside control limit.
R1	RPD non-calculable, as sample or duplicate results are less than five times the LOQ.
R2	Sample replicates RPD non-calculable, as only one replicate is within the analytical range.
LOQ1	Quantitation level raised due to low sample volume and/or dilution.
LOQ2	Quantitaion level raised due to matrix interference.
В	Analyte detected in method blank, but not in associated samples.
B1	The sample concentration is greater than 5 times the blank concentration.
B2	The sample concentration is less than 5 times the blank concentration.