

IDT0450

MORPHINE **IDENTIFICATION**

INSTRUCTIONS FOR USE

Read Before Use

As a precaution, gloves and eye protection should be worn when handling drug substances and using test units. Do not puncture any part of unit. Do not remove probe once a capsule is broken. In the unlikely event of a spill, inse affected area thoroughly with water. Keep out of reach of children. All results are presumptive. If necessary, further testing by a competent drug testing lab can provide additional identification.




Sample suspected substance

Tablet

Firmly press Smart-TipTM probe on TipTM probe on the tablet while scrubbing in a circular motion until the tip is completely covered with substance Lock Smart-TipTM probe TipTM probe into test unit. Press firmly.







Gently shake test unit for 2-3 seconds.



Results



The appearance of a purple color within one minute indicates the presence of an opiate, such as Morphine or Codeine




If no purple color develops with one minute, an opiate, such as Morphine or Codeine, is not present.



Proceed to Dilution (steps 6-9), regardless of whether or not a color appeared.





Gently shake test unit for 2-3 seconds.



Wait 5 minutes. The test window will warm up. This is normal.



Gently shake test unit for 2-3 seconds.



The test window will warm up. This is normal.

画

Disposal

After Dilution, the test kit is safe to dispose of in the trash.

IMPORTANT INFORMATION

until

- **IDenta Touch&Know**TM **Discreet Drug Test Kits** are used to determine the presence of certain drugs in substances such as powders, 1. crystals, granules, flakes, pills, tablets, or liquids.
- All results are presumptive. If necessary, further testing by a competent drug testing lab can provide additional verification. 2.
- A color change indicates a positive result for the listed drugs only 3. if it occurs within the first minute.

If there is NO color change within the time indicated, the listed drugs **are NOT present in the tested sample.** After a minute, the sampler and the liquid may begin changing color as a normal part of the testing process and this color change does <u>NOT</u> indicate a positive result.



Manufactured by IDenta Ltd



