

CERTIFICATE OF ANALYSIS

Product Number: HIS-0100

Description: Positive Control Sera: Human antibody against Histone

Lot Number: 5823

Appearance: Lyophilized

Assays:

ELISA:

Test	Results
SSA	188 EU/ml
SSB	181 EU/ml
SCL	16 EU/ml
Sm	3 EU/ml
Sm/RNP	14 EU/ml
Jo-1	7 EU/ml
Other: Whole Histone	312 EU/ml


Interpretation of Results:

<20 EU/ml: Negative

N/A: Not Available

Infectious Disease Testing:

This product has tested non-reactive for HBsAg, HIVAg, α -HIV, α -HBC, and α -HCV.



Quality Director

Date 2/26/2015