

## CERTIFICATE OF ANALYSIS

**Product Number:** HSA-0300  
**Description:** Positive Control Sera: Human antibody against SS-A (Ro)  
**Lot Number:** 4114  
**Appearance:** Lyophilized

**Assays:**


**Western Blot:**

Test date: 11/1/2010

<b>Blot type</b>	<b>Results</b>
Hep-2 blot	Primary band at 60kD and a secondary bands at 220k, 70kD, 50kD and 30kD

**Infectious Disease Testing:**

This product has tested non-reactive for HBsAg, HIVAg,  $\alpha$ -HIV,  $\alpha$ -HBC, and  $\alpha$ -HCV.



\_\_\_\_\_  
Quality Director

Date

\_\_\_\_\_  
11/2/2010