

CERTIFICATE OF ANALYSIS

Product Number: HTG-0100

Description: Positive Control Sera: Human antibody against Thyroglobulin

Lot Number: 2899

Appearance: Lyophilized

Assays:

ELISA:

Test	Results
SSA	3 EU/ml
SSB	3 EU/ml
SCL	2 EU/ml
Sm	4 EU/ml
Sm/RNP	4 EU/ml
Jo-1	5 EU/ml

Interpretation of Results:

<16 EU/ml: Negative

16-20 EU/ml: Equivocal

>20 EU/ml: Positive

N/A: Not Available

ELISA:


Test	Results
Thyroglobulin	2500 IU/ml

Interpretation of Results:

<100EU/ml: Normal

Infectious Disease Testing:

This product has tested non-reactive for HBsAg, HIVAg, α -HIV, α -HBC, and α -HCV.



Quality Director

10/21/2010

Date