


Certificate of Analysis

Product: SSA-3000
Lot Number: 6216
Appearance: Frozen clear liquid
Expiration Date: 01/May/2024
Material of Origin: Bovine Spleen from USA origin



Parameter	Specification	Result
Protein Concentration by BCA	0.048-0.448 mg/ml	0.111 mg/ml
Antigen titer by direct coat ELISA		
Vs Positive antibody 1	75-125%	141%
Vs Positive antibody 2	75-125%	141%
Vs Positive antibody 3	75-125%	159%
Mean	80-120%	147%
Noise/Signal ratio by direct coat ELISA		
Vs Normal serum	<0.139	-0.013
Vs SSB antibody	<1.624	0.282
Vs SCL-70 antibody	<0.201	0.074
Vs Sm antibody	<0.130	-0.002
Vs Sm-RNP Complex antibody	<0.156	0.005
Vs Mitochondrial antibody	<0.157	0.003
Vs Jo-1 antibody	<0.176	0.008

Double Immunodiffusion Results:

	Standard	Lot 6216
Titer vs specific antibody	1:4	1:8
Reaction with anti-host serum proteins	Negative	Negative

Absorbance at 280nm: 0.401 at 1: 1 dilution

[Signature]

Quality Director

1/23/2019
Date