

Medical Device Details

Sponsor:	Mango Medtech Limited 12 Madden Street Auckland New Zealand
WAND Reference:	220420-WAND-6YVTQ7
Sponsors Own Reference:	M-Close Kit
GMDN:	Laparoscopy wound closure system [60390]
Class:	IIa
Intended Purpose:	Laparoscopic port closure system including laparoscopic nerve block kit used in conjunction with or independently to perform abdominal wall nerve blocks.
Manufacturer:	New Wave Endo-Surgery Inc 6601 Lyons Road, Suite D-8 Coconut Creek Florida United States

ARTG ID:

- ▣ > The device is supplied sterile.
 - > Sterilisation method: Gamma irradiation
- > The device is intended to be invasive.
- > The device is intended for single use.
- > The device is not an active device.
- > The device does not contain material or ingredients of microbial origin.
- > The device does not contain material or ingredients of recombinant origin.
- > The device does not contain material or ingredients manufactured or formulated using a genetically modified organism.
- > The device does not contain material or ingredients of human origin.
- > The device does not contain human blood or its components.
- ▣ > The device consists of: Single product
 - ▣ Components:
 - > Component: Nerve-block Sharp Needle [61547], Device includes inbuilt needles designed to administer targeted local anaesthesia.
 - > The device does not contain material or ingredients of animal origin.
 - > The device is not medicated.
 - > The device is not formulated.
 - > The product does not contain a medicine that has consent for marketing in New Zealand.
 - > The product does not contain a medical device which incorporates a medicine as an integral part and that has an action ancillary to the device.

Status: Active