

1. Part Number: 8008P-250S

2. Photograph:



3. Product Information:

Description: SteriWare® Graduated Beaker 250ml - Sterile

Use: Single use, disposable beaker

Material of Construction: Polypropylene (PP) - virgin

- Conforms to FDA CFR 177.1520
- Conforms to EU Regulations 10/2011
- Conforms to EC Regulations 1935/2004
- BSE/TSE free

Method of Construction: Single piece, injection moulded

Nominal Volume: 250ml

Brimful Volume: 320ml

Max. Diameter (inc. spout): 80mm

Top diameter (excl. spout) 74mm

Bottom Diameter: 60mm

Overall Height: 108mm

Graduation Marks: From 50ml to 250ml (every 10ml)


Nominal Weight: 33g

Moulding Environment: Class 100,000 Cleanroom

Primary Packing Environment: Class 10,000 Cleanroom

Individually Bagged? Yes (heat sealed PE/PET bag)

Method of Sterilisation: Gamma Irradiation (25 to 45 kGy)



Sampling Systems Ltd.

4-5 Forge Mills Park, Coleshill, Warwickshire, B46 1JH, UK.
Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com

Quantity per Box: 40
Recommended Storage Conditions: Dry and ambient temperature
Shelf Life: 5 years from the date of manufacture
ATEX: This device has not been ATEX tested

4. Change Control:

Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to change.control@sampling.com

5. Country of Origin:

Made and packed in the UK.

6. Repeated Use:

This is a single-use, disposable device. Repeated use will lead to failure.

7. Approval:

Signed:



Name: Kuldeep Flora

Position: Quality Manager

Date: 6-Mar-23