

## Specification Sheet

Specification Number: 8017H-05S Revision: 07 Date Issued: 10-Mar-20

**1. Part Number:** 8017H-05S

### 2. Photograph:



#### 3. Product Information:

Description: SteriWare® Bowl 500ml - Sterile

Use: Single use bowl

Material of Construction: High Density Polyethylene (HDPE) – virgin

• Conforms to FDA 21 CFR 177.1520

• Conforms to EU Regulations 10/2011

• Conforms to EC Regulations 1935/2004

• BSE/TSE free

Method of Construction: Single piece, injection moulded

Bowl Diameter: 129mm
Bowl Width (inc, spout & handle): 180mm
Bowl Height: 70mm
Nominal Weight: 61g

Moulding & Primary Packing Environment: Class 100,000 Cleanroom

Individually Bagged? Yes (heat sealed PE/PET bag)
Method of Sterilisation: Gamma Irradiation (25-45kGy)

Quantity per Box: 100

Recommended Storage Conditions: Dry and ambient temperature

Shelf Life: 5 years from the date of manufacture

ATEX: Suitable for use in ATEX rated areas – Contact Sampling Systems

for a copy of the report to confirm suitability for your application



### Sampling Systems Ltd.

4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



# **Specification Sheet**

Specification Number: 8017H-05S Revision: 07 Date Issued: 10-Mar-20

4. Change Control: Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to

change.control@sampling.com

**5. Country of Origin:** Made and packed in the UK.

**6. Repeated Use:** This is a single-use, disposable sampler. Repeated use will lead to

failure.

**7. Approval:** Signed:

Name: Kuldeep Flora

Position: Quality Manager

Date: 10-Mar-20