



**1. Part Number:** 8051H-1100

**2. Photograph:**



**3. Product Information:**

|   |   |
|---|---|
| Description:                            | SteriWare® Jumbo Pipette 1100mm   |
| Use:                                    | Single use, disposable liquid sampler   |
| Material of Construction:               | High Density Polyethylene (HDPE) - virgin <ul style="list-style-type: none"> <li>• Conforms to FDA 21 CFR 177.1520</li> <li>• Conforms to EU Regulations 10/2011</li> <li>• Conforms to EC Regulations 1935/2004</li> <li>• BSE/TSE free</li> </ul> <p>A blue dye is added to one of the tips. The blue dye is certified:</p> <ul style="list-style-type: none"> <li>• FDA acceptable for food contact</li> <li>• Free of materials of animal origin</li> </ul> |
| Method of Construction:                 | Ultrasonic welding  |
| Nominal Length                          | 1090mm  |
| Diameter of Outer Tube:                 | 21mm  |
| Diameter of hole in tip:                | Handle End (Blue) = 1.5mm<br>Product End (Natural) = 3mm  |
| Nominal Weight of Sampler:              | 123g  |
| Assembly & Primary Packing Environment: | Class 100,000 Cleanroom   |
| Individually Bagged?                    | Yes (heat sealed PE/PET bag)  |
| Method of Sterilisation:                | N/A   |
| Number of Samplers per Box:             | 37  |
| Recommended Storage Conditions:         | Dry and ambient temperature   |
| Shelf Life:                             | 5 years from the date of manufacture  |

**4. Change Control:**

Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to [change.control@sampling.com](mailto:change.control@sampling.com)

**5. Country of Origin:**

Made and packed in the UK.

**6. Repeated Use:**

This is a single-use, disposable sampler. Repeated use will lead to failure.

**7. Approval:**

Signed:



Name: Kuldeep Flora

Position: Quality Manager

Date: 29-Jun-21