

Specification Sheet Specification Number: 8064H-1000 Revision: 07

Date Issued:

07 29<u>-Jun-21</u>

#### 1. Part Number:

8064H-1000

# 2. Photograph:



## **3. Product Information:**

Description:	SteriWare <sup>®</sup> Powder Probe 1000mm
Use:	Single use powder sampler
Material of Construction:	<ul> <li>High Density Polyethylene (HDPE) – virgin</li> <li>Conforms to FDA 21 CFR 177.1520</li> <li>Conforms to EU Regulations 10/2011</li> <li>Conforms to EC Regulations 1935/2004</li> <li>BSE/TSE free</li> </ul>
Method of Construction:	Ultrasonic welding
Nominal Length:	945mm
Blade Length:	845mm
Diameter of Blade:	22mm
Width of opening in blade:	10mm
Nominal Weight of Sampler:	123g
Moulding & Primary Packing Environment:	Class 100,000 Cleanroom
Individually Bagged?	Yes (heat sealed PE/PET bag)
Method of Sterilisation:	N/A
Quantity per Box:	100
Recommended Storage Conditions:	Dry and ambient temperature
Shelf Life:	5 years from the date of manufacture

### 4. Change Control:

Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to *change.control@sampling.com* 

### 5. Country of Origin:

Made and packed in the UK.



4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



**Specification Sheet Specification Number:** Revision: Date Issued:

8064H-1000 07 29-<u>Jun-21</u>

6. Repeated Use:

This is a single-use, disposable sampler. Repeated use will lead to failure.

7. Approval:

Signed:

Name: Position: Date:

Kuldeep Flora **Quality Manager** 29-Jun-21



4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com