

## **Specification Sheet**

Specification Number: 8069H-1000S Revision: 02

Date Issued: 13-Mar-20

**1. Part Number:** 8069H-1000S

## 2. Photograph:



## 3. Product Information:

Description: SteriWare® Mono Slot Sampler – 1000mm - Sterile

Use: Single use, disposable powder sampler

Material of Construction: HDPE (High Density Polyethylene) - virgin

Conforms to FDA CFR 177.1520
Conforms to EU Regulations 10/2011

Conforms to EO Regulations 10/2011
 Conforms to EC Regulations 1935/2004

BSE/TSE free

Method of Construction: Ultrasonic welding

Nominal Length: 968mm

Diameter of Blade: 22mm

Diameter of Outer Tube: 26mm

Number of Sampling Slots: 1

Slot Positions:



Sampling Slot Width: 8mm
Sampling Slot Length: 60mm
Nominal Weight of Sampler: 240g

Moulding & Primary Packing Environment: Class 100,000 Cleanroom
Individually Bagged? Yes (heat sealed PE/PET bag)
Method of Sterilisation: Gamma Irradiation (25-50kGy)

Quantity per Box: 20

Recommended Storage Conditions: Dry and ambient temperature

Shelf Life: 5 years from the date of manufacture
ATEX: This product has not been ATEX tested



4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



## **Specification Sheet**

Specification Number: 8069H-1000S Revision: 02 Date Issued: 13-Mar-20

4. Change Control: Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the

SteriWare range please email your contact details to

change.control@sampling.com

**5. Country of Origin:** Made and packed in the UK.

**6. Repeated Use:** This is a single-use, disposable sampler. Repeated use will lead to

failure.

7. Approval: Signed: Name: Kuldeep Flora

Position: Quality Manager

Date: 13-Mar-20