

**1. Part Number:** 8073H-1000

**2. Photograph:**




**3. Product Information:**

**Description:** SteriWare Slot Sampler - 1000mm  
**Use:** Single use, disposable powder sampler  
**Material of Construction:** High Density Polyethylene (HDPE) - virgin  

- Conforms to FDA 21 CFR 177.1520
- Conforms to EU Regulations 10/2011
- Conforms to EC Regulations 1935/2004
- BSE/TSE free

**Method of Construction:** Ultrasonic welding  
**Overall Length:** 968mm  
**Diameter of Outer Tube:** 26mm  
**Diameter of Inner Blade:** 22mm  
**Nominal Weight of Sampler:** 234g  
**Slot Positions:**



**Number of Sampling Slots:** 3  
**Sampling Slot Width:** 8mm  
**Assembly & Primary Packing Environment:** Class 100,000 Cleanroom  
**Individually Bagged?** Yes (heat sealed PE/PET bag)  
**Method of Sterilisation:** N/A  
**Number of Samplers per Box:** 20  
**Recommended Storage Conditions:** Dry and ambient temperature  
**Shelf Life:** 5 years from the date of manufacture  
**ATEX:** Suitable for use in ATEX rated areas – Contact Sampling Systems for a copy of the report to confirm suitability for your application



**Sampling Systems Ltd.**

4-5 Forge Mills Park, Coleshill, Warwickshire, B46 1JH, UK.

Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com

**4. Change Control:**

Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to

[change.control@sampling.com](mailto:change.control@sampling.com)

**5. Country of Origin:**

Made and packed in the UK.

**6. Repeated Use:**

This is a single-use, disposable sampler. Repeated use will lead to failure.

**7. Approval:**

Signed:



Name: Kuldeep Flora

Position: Quality Manager

Date: 29-Jun-21