

Certificate of Conformity

Konformitätsbescheinigung
Certificat de Conformité
Certificato di conformità
Certificado de conformidad
Certificado de Conformidade



Page 1 of 1

Part Number	8080A-60 8080A-150 8080A-450 8080A-900 8080A-1200 8080A-5500 8080A-10500	8080L-60 8080L-150 8080L-450 8080L-900 8080L-1200
Description	PureBags	
Material	LDPE (Low Density Polyethylene)	
Temperature Rating	80°C to – 100°C	
Free from:	Pyrogens, R-Nase & D-Nase	
Shelf Life:	5 years from date of manufacture	
Date	07 October 2020	

We confirm that these bags are manufactured from LDPE which conforms to FDA 21 CFR 177.1520.

The manufacturer confirms these bags are manufactured from fully synthetic polymers. No animal sourced materials are used in the production of PureBags.

The bags are made from extruded virgin polyethylene at temperatures up to 200°C. The polyethylene tube manufacturing process is done by extrusion and the internal tubing is never exposed to the external environment. This ensures the bags sterility up until the point at which it is opened by the user.

The sterility of the manufactured bags is controlled by periodically sampling and analysing various production lines. In addition, sterilization by the extrusion process has been scientifically validated. Independent testing has confirmed that the sterile sampling bags we supply remain sterile for at least five years if they are not opened.

The information detailed above has been given to us by our suppliers, it is passed on in good faith and is to the best of our knowledge correct.

For and on behalf of Sampling Systems Ltd.

A handwritten signature in blue ink, appearing to read 'Kuldeep Flora'.

Kuldeep Flora
Quality Manager

This certificate is valid for 3 years from the date given above. Download the most recent certificate at: www.sampling.com/certificates

