


**1. Part Number:** 8095H-150

**2. Photograph:**



**3. Product Information:**

|  |   |
|--|---|
| <b>Description:</b>                                | SteriWare Powder Spatula HDPE - 150mm   |
| <b>Use:</b>  | Single use, disposable powder spatula   |
| <b>Material of Construction:</b>                   | High Density Polyethylene (HDPE) – rigid, white <ul style="list-style-type: none"> <li>• FDA acceptable CFR 177.1520</li> <li>• EU Regulations 10/2011 compliant</li> <li>• EC Regulations 1935/2004 compliant</li> <li>• BSE/TSE free</li> </ul> |
| <b>Method of Construction:</b>                     | Single piece, injection moulded   |
| <b>Blade Shape:</b>                                | U   |
| <b>Blade Style:</b>                                | Open Tip  |
| <b>Nominal Length:</b>                             | 230mm   |
| <b>Blade Length:</b>                               | 145mm   |
| <b>Blade Width:</b>                                | 18mm  |
| <b>Nominal Weight:</b>                             | 17g   |
| <b>Moulding &amp; Primary Packing Environment:</b> | Class 100,000 Cleanroom   |
| <b>Individually Bagged?</b>                        | Yes (heat sealed PE/PET bag)  |
| <b>Method of Sterilisation:</b>                    | N/A   |
| <b>Number of Samplers per Box:</b>                 | 100   |
| <b>Recommended Storage Conditions:</b>             | Dry and ambient temperature   |
| <b>Shelf Life:</b>                                 | 5 years from the date of manufacture  |
| <b>ATEX:</b>                                       | This device has not been ATEX tested  |

**4. Change Control:**

Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to

[change.control@sampling.com](mailto:change.control@sampling.com)

**5. Country of Origin:**

Made and packed in the UK.

**6. Repeated Use:**

This is a single-use, disposable sampler. Repeated use will lead to failure.

**7. Approval:**

Signed:



Name: Kuldeep Flora

Position: Quality Manager

Date: 26-Oct-20