

Specification Number: 8370H-100

Revision: Date Issued:

07 14-Sep-20

1. Part Number:

8370H-100

2. Photograph:



3. Product Information:

| Description: | SteriWare [®] Funnel 100mm |
|---|---|
| Use: | Single use, disposable funnel |
| Material of Construction: | High Density Polyethylene (HDPE) - virgin Conforms to FDA CFR 177.1520 Conforms to EU Regulations 10/2011 Conforms to EC Regulations 1935/2004 BSE/TSE free |
| Method of Construction: | Single piece, injection moulded |
| Max. Diameter: | 117mm |
| ID at Top of Funnel: | 100mm |
| Overall Height: | 103mm |
| OD at Bottom of Spout: | 11mm |
| ID at Bottom of Spout: | 8mm |
| OD at Top of Spout: | 17mm |
| Nominal Weight: | 16g |
| Moulding & Primary Packing Environment: | Class 100,000 Cleanroom |
| Individually Bagged? | Yes (heat sealed PE/PET bag) |
| Method of Sterilisation: | N/A |
| Quantity per Box: | 50 |
| Recommended Storage Conditions: | Dry and ambient temperature |
| Shelf Life: | 5 years from the date of manufacture |
| ATEX: | This device has not been ATEX tested |



4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK.

Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



Specification Sheet Specification Number: 8370H-100 Revision: 14-Se<u>p-20</u> Date Issued:

07

4. Change Control:

Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to change.control@sampling.com

5. Country of Origin:

Made and packed in the UK.

6. Repeated Use:

7. Approval:

This is a single-use, disposable device. Repeated use will lead to failure.

Signed:

Name: Kuldeep Flora Position: **Quality Manager** 14-Sep-20 Date:



Sampling Systems Ltd. 4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com