

## **Specification Sheet**

Specification Number: Revision: Date Issued:

8404H-250S 01 15-Jul-21

1. Part Number:

8404H-250S

2. Photograph:



## 3. Product Information:

Description	Wide Mouth HDPE Bottle 250ml - Sterile
Materials of Construction:	BottleHDPE (conforms to FDA 177.1520 & USP Class VI)CapPP(conforms to FDA 177.1520 & USP Class VI)
Manufacturing Environment:	Controlled environment within factory
Packing Environment:	Class 10,000 Cleanroom
Nominal Volume:	250ml
Brimful Volume:	290ml
Nominal Height:	131mm
Diameter:	62mm
Internal Neck Diameter:	32mm
Weight:	37g
Individually Bagged:	Yes (in a PE/PET laminate film bag)
Quantity per box:	50
Method of Sterilisation:	Gamma Irradiation (25-45 kGy)
BSE/TSE Free:	Yes
Supplied with lid screwed on:	Yes
Recommended Storage Conditions:	Dry and ambient temperature
Shelf Life:	5 years from the date of manufacture
ATEX:	This device has not been ATEX tested



4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



**Specification Sheet Specification Number:** 8404H-250S Revision: Date Issued: 15-Jul-21

01

4. Change Control: Sampling Systems operates a policy of continual improvement.

> To be kept informed of all major changes to any product in the SteriWare range please email your contact details to change.control@sampling.com

- 5. Country of Origin: Packed and sterilised in the UK.
- 6. Repeated Use:

This is a single-use, disposable device. Repeated use will lead to failure.

7. Approval:

Signed:

Name: Kuldeep Flora Position: **Quality Manager** 15-Jul-21 Date:



Sampling Systems Ltd. 4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com