

1. Part Number: 8432P-1000

2. Photograph:



3. Product Information:

Description:	SteriWare® Graduated Jug 1000ml
Use:	Single use, disposable jug
Material of Construction:	Polypropylene (PP) – virgin <ul style="list-style-type: none"> • Conforms to FDA CFR 177.1520 • Conforms to EU Regulations 10/2011 • Conforms to EC Regulations 1935/2004 • BSE/TSE free
Method of Construction:	Single piece, injection moulded
Nominal Volume:	1000ml
Brimful Volume:	1270ml
Top Diameter (excl. spout):	120mm
Max Width (inc. spout and handle)	170mm
Base Diameter:	93mm
Height:	170mm
Nominal Weight:	123g
Graduation Marks:	To 1000ml (every 10ml)
Moulding & Primary Packing Environment:	Class 100,000 Cleanroom
Individually Bagged?	Yes (heat sealed PE/PET bag)
Method of Sterilisation:	N/A
Quantity per Box:	18

Recommended Storage Conditions:

Dry and ambient temperature

Shelf Life:

5 years from the date of manufacture

ATEX:

This device has not been ATEX tested

4. Change Control:

Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to

change.control@sampling.com**5. Country of Origin:**

Made and packed in the UK.

6. Repeated Use:

This is a single-use, disposable sampler. Repeated use will lead to failure.

7. Approval:

Signed:



Name: Kuldeep Flora

Position: Quality Manager

Date: 26-Mar-20