

Specification Sheet

Specification Number: 8444A-1S Revision: 04

Date Issued: 4-Sep-19

1. Part Number: 8444A-1S

2. Photograph:



3. Product Information:

Description: SteriWare® Capsule Counting Tray
Use: Single use, tray for counting capsules

Material of Construction: Polystyrene – virgin, white

Conforms to FDA CFR 177.1640
Conforms to EU Regulations 10/2011
Conforms to EC Regulations 1935/2004

BSE/TSE free

Method of Construction: Single piece, injection moulded

Length: 260mm
Height 10mm
Width 130mm
Nominal Weight: 50g

Moulding & Primary Packing Environment: Class 100,000 Cleanroom

Individually Bagged? Yes (heat sealed PE/PET bag)

Method of Sterilisation: Gamma Irradiation (Minimum 25 kGy)

Number of Samplers per Box: 20

Recommended Storage Conditions: Dry and ambient temperature

Shelf Life: 5 years from the date of manufacture
ATEX: This device has not been ATEX tested



Sampling Systems Ltd.

4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



Specification Sheet

Specification Number:8444A-1SRevision:04Date Issued:4-Sep-19

4. Change Control: Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the

SteriWare range please email your contact details to

change.control@sampling.com

5. Country of Origin: Made and packed in the UK.

6. Repeated Use: This is a single-use, disposable device. Repeated use will lead to

failure.

7. Approval: Signed:

Name: Kuldeep Flora
Position: Quality Manager

Date: 4-Sep-19