

**Specification Sheet** 

Specification Number: Revision: Date Issued: 8785B-75S 02 26-Mar-20

1. Part Number:

8785B-75S

2. Photograph:



## 3. Product Information:

Description:	
Use:	
Material of Construction:	

Method of Construction:

Overall Length: Blade Length:

Max. Blade Width:

Individually Bagged? Method of Sterilisation:

Nominal Weight:

SteriWare<sup>®</sup> Scraper 75mm Blue - Sterile Single use, disposable scraper

Polypropylene – blue - virgin

- Conforms to FDA CFR 177.1520
- Conforms to EU Regulations 10/2011
- Conforms to EC Regulations 1935/2004
- BSE/TSE free

Single piece, injection moulded

19	Umn	٦

75mm

75mm

36g

100

Moulding & Primary Packing Environment: Class 100,000 Cleanroom

Yes (heat sealed PE/PET bag)

Dry and ambient temperature

Gamma Irradiation (Minimum 25 kGy)

5 years from the date of manufacture

This device has not been ATEX tested

Quantity per Box:

Recommended Storage Conditions:

Shelf Life:

ATEX:

QMS<sup>o</sup> ISO 9001 : 2015 R E G I S T E R E D Cert No. 14128455

4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



**Specification Sheet Specification Number:** 8785B-75S Revision: Date Issued: 26-Mar-20

02

4. Change Control:

Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to change.control@sampling.com

5. Country of Origin:

Made and packed in the UK.

6. Repeated Use:

7. Approval:

This is a single-use, disposable sampler. Repeated use will lead to failure.

Signed:

Name: Kuldeep Flora Position: **Quality Manager** Date: 26-Mar-20



Sampling Systems Ltd. 4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com