

	<h1 style="margin: 0;">Sampler Specification</h1>	Specification Number	8921H-1100
		Revision	01

Part No. 8921H-1100

Description SteriWare LiquiDip Duo –1100mm

Material of Construction: High Density Polyethylene (HDPE) - virgin

- Conforms to FDA CFR 177.1520
- Conforms to EU Regulations 10/2011
- Conforms to EC Regulations 1935/2004



Method of Construction: Ultrasonic welding

Overall Length 1125mm

Maximum Diameter: 26mm

Outer Diameter of Tube: 21mm

Diameter of hole in tip: 6mm

Nominal Weight of Sampler: 128g


Assembly & Packing Environment: Class 100,000 Medical Cleanroom

Individually Bagged? Yes (heat sealed PE bag)

Method of Sterilisation: N/A

Number of Samplers per Box: 20

Recommended Storage Conditions: Dry and ambient temperature

Specification Approved By:	Signature: 	Name: Jonathan Fenner Position: Director Date: 27-Apr-17
Page 1 of 1	Sampling Systems Ltd. 4-5 Forge Mills Park, Station Road, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com	